

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031615\
 Data File : BM000713.D
 Acq On : 16 Mar 2015 12:25
 Operator : TP/IZ
 Sample : SSTD0.485
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.485

Quant Time: Mar 17 04:36:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 17 04:25:21 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.64	152	8106	0.40	ng/ul	0.00
2) Naphthalene-d8	10.41	136	28990	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.27	164	15417	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.02	188	30522	0.40	ng/ul	0.00
16) Chrysene-d12	21.23	240	21113	0.40	ng/ul	0.00
20) Perylene-d12	23.42	264	17666	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.00	152	15178	0.40	ng/ul	0.00
14) Fluoranthene-d10	19.07	212	24902	0.37	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.46	128	31455	0.41	ng/ul	100
5) 2-Methylnaphthalene	12.08	142	20943	0.41	ng/ul	100
7) Acenaphthylene	14.00	152	28383	0.42	ng/ul	100
8) Acenaphthene	14.33	153	21870	0.42	ng/ul	100
9) Fluorene	15.34	166	25114	0.42	ng/ul	100
11) Pentachlorophenol	16.68	266	2212	0.65	ng/ul	100
12) Phenanthrene	17.07	178	37311	0.41	ng/ul	100
13) Anthracene	17.16	178	33545	0.41	ng/ul	100
15) Fluoranthene	19.09	202	37684	0.36	ng/ul	100
17) Pyrene	19.45	202	37302	0.45	ng/ul	100
18) Benzo(a)anthracene	21.21	228	26510	0.41	ng/ul	100
19) Chrysene	21.26	228	31361	0.40	ng/ul	100
21) Benzo(b)fluoranthene	22.76	252	28297	0.42	ng/ul	100
22) Benzo(k)fluoranthene	22.80	252	25335	0.40	ng/ul	100
23) Benzo(a)pyrene	23.31	252	24729	0.41	ng/ul	100
24) Indeno(1,2,3-cd)pyrene	25.61	276	27066	0.41	ng/ul	100
25) Dibenzo(a,h)anthracene	25.62	278	21813	0.42	ng/ul	100
26) Benzo(g,h,i)perylene	26.27	276	23691	0.41	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031615\
Data File : BM000713.D
Acq On : 16 Mar 2015 12:25
Operator : TP/IZ
Sample : SST0.485
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.485

Quant Time: Mar 17 04:36:49 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 17 04:25:21 2015
Response via : Initial Calibration

