


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031615\
 Data File : BM000710.D
 Acq On : 16 Mar 2015 10:39
 Operator : TP/IZ
 Sample : DFTPP38
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP38

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031615.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 17 04:53:53 2015



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2507

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.7	85357	PASS
68	69	0.00	2	1.5	1205	PASS
69	198	0.00	100	34.8	78917	PASS
70	69	0.00	2	0.4	317	PASS
127	198	10	80	45.2	102381	PASS
197	198	0.00	2	0.6	1393	PASS
198	198	100	100	100.0	226474	PASS
199	198	5	9	6.8	15297	PASS
275	198	10	60	24.1	54658	PASS
365	198	1	100	2.6	5806	PASS
441	443	0.01	100	75.1	31568	PASS
442	198	50	100	95.6	216448	PASS
443	442	15	24	19.4	42048	PASS

Scan 2507 (17.433 min): BM000710.D
DFTPP38

m/z	Abundance
40.00	3341.0
42.00	310.0
43.10	1113.0
44.10	7123.0
45.00	490.0
60.00	320.0
207.00	468.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
36.90	401.0
38.00	1219.0
39.10	7792.0
40.00	4013.0
41.00	370.0
43.00	668.0
44.00	7336.0
45.00	682.0
49.10	1119.0
50.10	27712.0
51.10	112088.0
52.10	6393.0
53.10	420.0
55.00	698.0
56.00	3228.0
57.00	8391.0
57.90	522.0
60.10	390.0
61.10	1147.0
62.00	1528.0
63.00	4478.0
64.00	830.0
65.10	2003.0
68.00	1624.0
69.00	102032.0
70.00	559.0
73.10	1017.0
74.00	9635.0
75.10	16200.0
76.00	4673.0
77.10	114928.0
78.00	8121.0
79.00	7621.0
80.00	5312.0
81.00	8557.0
82.00	2089.0
83.10	2006.0
84.00	328.0
85.10	1468.0
86.00	2331.0
87.00	1214.0
87.80	361.0
91.10	1858.0
92.00	2539.0
93.00	12773.0
94.00	849.0
96.00	694.0
98.00	9215.0
99.00	7698.0
100.00	686.0
101.00	4278.0
102.90	1327.0
104.00	3077.0
105.00	2646.0
105.90	776.0
107.00	34432.0
108.10	5850.0
109.10	969.0
110.00	65920.0
111.00	10169.0
112.00	1230.0
113.10	422.0
116.00	1936.0
117.00	26304.0
118.00	1856.0
118.90	317.0
120.10	443.0
121.90	1752.0
123.00	3488.0
124.00	1451.0
125.00	1332.0
126.10	713.0
127.00	122640.0
128.00	8720.0
129.00	45000.0
130.00	4216.0
131.00	995.0
134.00	1365.0
135.00	3498.0
136.00	1413.0
137.00	2003.0
138.10	591.0
140.00	620.0
141.00	5458.0
141.90	1918.0
143.00	1342.0
144.00	350.0
145.00	327.0
145.90	987.0
147.10	2887.0
148.00	6598.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

148.90	1587.0
150.00	548.0
151.10	819.0
151.70	419.0
152.20	348.0
153.00	2081.0
153.90	1269.0
155.00	3164.0
156.10	4660.0
157.00	1033.0
158.10	897.0
158.90	1118.0
160.10	2448.0
161.00	3171.0
162.00	844.0
163.00	334.0
164.00	572.0
165.00	2782.0
166.00	1607.0
167.00	10879.0
168.00	7765.0
169.00	1207.0
170.90	523.0
172.00	1009.0
173.00	1564.0
174.00	2214.0
175.10	4135.0
176.00	1751.0
177.10	1870.0
177.90	713.0
179.00	9223.0
180.00	7467.0
181.00	2975.0
182.10	672.0
184.00	643.0
185.00	4434.0
186.00	32664.0
187.10	9890.0
188.00	1403.0
189.00	2056.0
191.10	1005.0
192.00	2235.0
193.00	3608.0
194.00	711.0
194.80	489.0
196.10	7535.0
197.00	1323.0
198.00	254592.0
199.00	18144.0
199.90	1363.0
200.80	468.0
201.70	1194.0
203.00	1737.0
204.00	8528.0
205.10	14035.0
206.10	65024.0
207.10	8100.0
208.00	2459.0
209.00	989.0
210.20	909.0
211.10	2290.0
215.00	904.0
215.90	1181.0
217.00	16143.0
218.00	1892.0
221.10	11228.0
221.80	2331.0
223.00	3061.0
224.00	33000.0
225.00	8784.0
226.00	1565.0
227.00	12749.0
228.00	1763.0
229.00	2821.0
230.00	426.0
231.10	1230.0
232.90	302.0
234.10	807.0
235.00	818.0
236.10	532.0
237.00	2381.0
238.90	616.0
240.00	409.0
241.00	899.0
242.00	1444.0
243.00	1661.0
244.10	26424.0
245.10	3422.0
246.00	4346.0
246.90	854.0
248.90	1261.0
251.20	354.0
252.90	472.0
254.10	1172.0
255.00	130368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

256.00	18152.0
257.00	2398.0
258.00	6522.0
259.00	1227.0
265.00	2691.0
266.10	328.0
271.10	357.0
272.00	389.0
273.00	3707.0
274.00	9746.0
275.00	57760.0
276.00	8151.0
277.00	4508.0
278.00	920.0
283.00	600.0
285.00	590.0
293.00	1141.0
294.10	617.0
294.90	392.0
296.00	13536.0
296.90	1886.0
303.00	1711.0
303.90	456.0
314.00	639.0
315.00	1195.0
316.10	846.0
321.00	566.0
322.00	324.0
323.00	5218.0
324.00	861.0
326.80	752.0
327.90	627.0
331.90	360.0
332.90	364.0
334.00	3060.0
335.00	993.0
341.00	581.0
346.10	936.0
347.00	339.0
352.00	1292.0
353.00	828.0
354.00	1886.0
365.00	5584.0
366.00	683.0
371.10	437.0
372.00	2268.0
373.10	604.0
383.00	706.0
389.90	411.0
401.90	1405.0
403.00	1295.0
404.00	825.0
421.10	994.0
421.90	1570.0
423.00	9530.0
424.00	4189.0
424.80	791.0
441.00	28320.0
442.00	192256.0
443.00	38048.0
444.00	3185.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
37.00	405.0
38.00	1138.0
39.00	7485.0
40.00	3637.0
40.90	395.0
42.00	469.0
43.10	709.0
44.00	7044.0
45.00	758.0
49.00	810.0
50.10	24344.0
51.10	97160.0
52.10	4753.0
55.00	729.0
56.00	3103.0
57.00	7546.0
57.90	528.0
59.90	391.0
61.10	1260.0
62.00	1323.0
63.00	4248.0
64.00	667.0
65.00	1436.0
68.00	1178.0
69.00	90624.0
70.00	394.0
73.00	775.0
74.10	8918.0
75.00	14130.0
76.10	4330.0
77.10	106360.0
78.10	6730.0
79.00	6840.0
80.00	5540.0
81.00	7257.0
82.00	1874.0
83.00	1829.0
83.90	381.0
84.90	1675.0
86.00	1912.0
87.00	1218.0
88.10	423.0
91.00	1653.0
92.00	1713.0
93.00	11817.0
94.10	508.0
96.00	477.0
98.00	9023.0
99.00	8272.0
100.00	878.0
101.00	4909.0
103.00	1223.0
104.00	3048.0
104.90	2626.0
106.00	1063.0
107.00	31640.0
108.10	5002.0
109.10	1041.0
110.00	64144.0
111.00	9415.0
112.00	949.0
112.90	370.0
115.90	1531.0
117.00	24824.0
118.00	1870.0
119.10	301.0
122.00	1824.0
123.00	3356.0
124.00	1292.0
125.00	1236.0
126.10	417.0
127.00	121392.0
128.00	8286.0
129.00	48224.0
130.00	3702.0
130.90	736.0
132.00	425.0
134.10	1252.0
135.00	4235.0
136.10	1701.0
137.00	2252.0
138.00	325.0
139.90	518.0
141.00	6139.0
142.20	2077.0
143.10	1292.0
145.00	436.0
146.00	1044.0
147.00	3372.0
148.00	7038.0
148.90	1672.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

149.90	328.0
151.20	730.0
151.70	522.0
153.00	1773.0
154.00	1494.0
155.00	3327.0
156.00	5717.0
157.10	825.0
158.00	1210.0
159.00	702.0
160.00	1968.0
161.00	2778.0
162.10	647.0
163.80	361.0
165.00	2345.0
166.00	1979.0
167.00	11714.0
168.00	6695.0
169.00	1222.0
170.20	384.0
171.10	488.0
171.90	992.0
172.90	1392.0
174.00	2556.0
175.00	4740.0
175.90	1360.0
177.00	2067.0
178.00	699.0
179.00	9492.0
180.00	7119.0
181.00	2854.0
182.00	559.0
184.00	854.0
185.00	4239.0
186.00	33400.0
187.00	10502.0
188.10	1576.0
189.00	2067.0
190.00	327.0
191.10	975.0
192.00	2999.0
193.00	3712.0
194.10	763.0
194.70	449.0
196.00	7889.0
197.00	1621.0
198.00	272640.0
199.00	17792.0
200.10	1299.0
201.40	1342.0
203.00	1562.0
204.10	9000.0
205.10	15567.0
206.10	65560.0
207.10	8684.0
208.10	2386.0
209.00	595.0
210.00	1102.0
210.90	2219.0
214.90	517.0
215.90	1437.0
217.00	17680.0
218.00	2279.0
221.10	11683.0
223.10	3198.0
224.10	36576.0
225.10	8896.0
226.10	1076.0
227.00	14737.0
228.00	2152.0
229.10	2610.0
230.00	339.0
231.00	1423.0
233.00	312.0
233.90	911.0
235.00	1111.0
236.10	595.0
237.00	2012.0
239.10	535.0
239.90	503.0
240.90	818.0
242.00	1804.0
243.10	1887.0
244.10	29160.0
245.10	3660.0
246.00	5090.0
247.00	1146.0
248.00	396.0
249.10	1127.0
252.90	446.0
254.10	1133.0
255.00	146496.0
256.00	21864.0
256.90	1661.0
258.00	8309.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

259.00	1405.0
260.00	413.0
265.00	3296.0
272.00	350.0
273.00	3641.0
274.00	11758.0
275.00	65032.0
276.10	9201.0
277.00	4647.0
278.00	874.0
283.00	803.0
283.90	301.0
285.10	982.0
291.90	331.0
293.10	1208.0
294.00	395.0
294.80	352.0
296.00	16712.0
296.90	2267.0
302.10	393.0
303.00	1797.0
303.90	472.0
314.10	880.0
315.00	1574.0
316.00	1428.0
321.00	749.0
322.00	506.0
323.10	6109.0
324.10	815.0
327.00	1157.0
328.00	467.0
332.90	593.0
334.10	4040.0
335.10	849.0
341.00	698.0
346.10	1444.0
347.00	448.0
352.00	2256.0
353.00	1394.0
353.90	1727.0
355.00	337.0
365.00	6839.0
366.00	1023.0
371.00	599.0
372.00	3399.0
373.00	589.0
383.00	681.0
390.10	333.0
391.00	355.0
401.90	1196.0
403.10	2189.0
403.90	939.0
420.90	1343.0
422.00	1921.0
423.00	13721.0
424.00	4621.0
425.10	676.0
441.00	38496.0
442.00	259776.0
443.00	50976.0
444.00	4651.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
38.10	652.0
39.10	3853.0
40.00	3151.0
41.00	395.0
43.10	505.0
44.00	7466.0
45.00	771.0
49.10	449.0
50.00	11270.0
51.10	46824.0
52.10	2721.0
56.00	1322.0
57.00	3095.0
61.00	632.0
62.00	754.0
63.10	1646.0
65.10	1095.0
68.00	813.0
69.00	44096.0
73.00	501.0
74.10	4062.0
75.00	6485.0
76.00	2447.0
77.10	50776.0
78.10	3607.0
79.00	3596.0
80.00	2495.0
81.00	3824.0
82.10	844.0
83.00	870.0
84.90	881.0
86.00	1060.0
87.00	544.0
91.00	909.0
92.00	947.0
93.00	5754.0
94.00	426.0
98.00	4771.0
99.00	3690.0
101.00	2145.0
103.00	629.0
104.00	1750.0
105.10	1003.0
106.00	485.0
107.00	16120.0
108.10	2380.0
109.10	369.0
110.00	32648.0
111.00	5105.0
112.00	555.0
115.90	813.0
117.00	11986.0
118.00	1118.0
120.10	345.0
122.00	1064.0
123.00	1832.0
124.00	863.0
125.00	695.0
126.10	411.0
127.00	63112.0
128.00	3983.0
129.00	24392.0
130.00	1954.0
131.10	395.0
133.90	557.0
135.00	2116.0
135.90	712.0
137.00	996.0
141.00	2915.0
142.20	935.0
142.80	816.0
146.00	797.0
147.00	1962.0
148.00	4046.0
149.00	885.0
151.20	618.0
152.90	1066.0
154.20	757.0
154.90	1715.0
156.10	3343.0
156.90	524.0
158.00	746.0
159.00	526.0
160.10	1350.0
161.00	1356.0
162.00	507.0
164.90	1071.0
166.00	682.0
167.00	6548.0
168.00	4227.0
169.00	487.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

171.90	592.0
173.10	832.0
174.00	1400.0
175.00	3202.0
176.00	872.0
177.00	1093.0
179.00	5440.0
180.00	4048.0
181.00	1531.0
182.10	306.0
183.90	459.0
185.00	2554.0
186.00	18088.0
187.10	5215.0
188.10	655.0
189.10	1105.0
191.10	457.0
192.00	1896.0
193.10	1990.0
194.00	426.0
196.10	4139.0
197.10	1237.0
198.00	152192.0
199.00	9955.0
200.00	713.0
201.60	709.0
203.00	1014.0
204.00	5161.0
205.10	8685.0
206.10	38456.0
207.10	5263.0
207.90	1180.0
209.00	463.0
210.00	629.0
211.00	1712.0
215.00	449.0
216.00	743.0
217.00	10403.0
218.10	1015.0
221.00	5534.0
221.80	2036.0
223.00	2621.0
224.00	20320.0
225.00	5603.0
226.10	471.0
227.00	8457.0
228.00	1369.0
229.00	2069.0
231.10	821.0
234.10	491.0
235.10	704.0
236.00	319.0
237.00	604.0
241.00	593.0
242.10	1225.0
243.00	1241.0
244.00	17024.0
245.10	1938.0
246.00	2952.0
247.00	493.0
249.00	492.0
254.10	615.0
255.00	88144.0
256.00	12115.0
257.00	972.0
258.00	4824.0
259.00	977.0
265.00	1897.0
273.00	2826.0
274.00	6226.0
275.00	41184.0
276.10	5289.0
277.00	3352.0
278.00	699.0
283.00	315.0
285.00	517.0
293.10	748.0
296.00	9770.0
297.10	1475.0
303.00	1201.0
304.10	303.0
313.90	505.0
315.10	1111.0
316.10	887.0
321.10	455.0
323.00	3890.0
324.00	683.0
327.00	651.0
328.00	310.0
332.90	452.0
334.00	2727.0
335.00	871.0
341.00	508.0
345.80	785.0
351.90	1201.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

353.10	863.0
354.00	1096.0
355.10	399.0
364.90	4995.0
365.90	691.0
372.00	2152.0
373.00	323.0
383.10	859.0
389.90	312.0
401.90	782.0
402.90	1434.0
404.00	397.0
421.00	1112.0
422.00	1334.0
423.00	9781.0
424.00	2988.0
441.00	27888.0
442.00	197312.0
443.00	37120.0
444.00	2980.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38