

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031623\
 Data File : BM039015.D
 Acq On : 16 Mar 2023 12:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 17 05:49:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 05:25:34 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.586	0.541	7.7	96	0.00
3	Pyridine	1.624	1.649	-1.5	103	0.00
4	n-Nitrosodimethylamine	0.661	0.682	-3.2	106	0.00
5 S	2-Fluorophenol	1.316	1.281	2.7	100	0.00
6	Aniline	2.245	2.390	-6.5	109	0.00
7 S	Phenol-d6	1.710	1.753	-2.5	106	0.00
8	2-Chlorophenol	1.401	1.416	-1.1	104	0.00
9	Benzaldehyde	0.814	0.766	5.9	106	0.00
10 C	Phenol	1.816	1.863	-2.6	106	0.00
11	bis(2-Chloroethyl)ether	1.476	1.536	-4.1	108	0.00
12	1,3-Dichlorobenzene	1.489	1.442	3.2	100	0.00
13 C	1,4-Dichlorobenzene	1.515	1.471	2.9	100	0.00
14	1,2-Dichlorobenzene	1.459	1.432	1.9	101	0.00
15	Benzyl Alcohol	1.226	1.346	-9.8	112	0.00
16	2,2'-oxybis(1-Chloropropane	1.864	2.094	-12.3	118	0.00
17	2-Methylphenol	1.247	1.320	-5.9	108	0.00
18	Hexachloroethane	0.559	0.564	-0.9	103	0.00
19 P	n-Nitroso-di-n-propylamine	1.077	1.275	-18.4	118	0.00
20	3+4-Methylphenols	1.666	1.796	-7.8	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
22	Acetophenone	0.561	0.548	2.3	110	0.00
23 S	Nitrobenzene-d5	0.407	0.412	-1.2	113	0.00
24	Nitrobenzene	0.428	0.432	-0.9	113	0.00
25	Isophorone	0.749	0.801	-6.9	118	0.00
26 C	2-Nitrophenol	0.169	0.182	-7.7	117	0.00
27	2,4-Dimethylphenol	0.326	0.326	0.0	111	0.00
28	bis(2-Chloroethoxy)methane	0.481	0.492	-2.3	116	0.00
29 C	2,4-Dichlorophenol	0.310	0.308	0.6	110	0.00
30	1,2,4-Trichlorobenzene	0.338	0.313	7.4	106	0.00
31	Naphthalene	1.140	1.095	3.9	109	0.00
32	Benzoic acid	0.216	0.233	-7.9	120	0.00
33	4-Chloroaniline	0.486	0.492	-1.2	112	0.00
34 C	Hexachlorobutadiene	0.194	0.177	8.8	104	0.00
35	Caprolactam	0.094	0.099	-5.3	113	0.00
36 C	4-Chloro-3-methylphenol	0.330	0.344	-4.2	115	0.00
37	2-Methylnaphthalene	0.764	0.758	0.8	112	0.00
38	1-Methylnaphthalene	0.716	0.718	-0.3	113	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	0.650	0.594	8.6	108	0.00
41 P	Hexachlorocyclopentadiene	0.357	0.349	2.2	112	0.00
42 S	2,4,6-Tribromophenol	0.231	0.224	3.0	110	0.00
43 C	2,4,6-Trichlorophenol	0.421	0.416	1.2	111	0.00
44	2,4,5-Trichlorophenol	0.493	0.484	1.8	110	0.00
45 S	2-Fluorobiphenyl	1.536	1.473	4.1	112	0.00
46	1,1'-Biphenyl	1.694	1.631	3.7	113	0.00
47	2-Chloronaphthalene	1.278	1.218	4.7	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031623\
 Data File : BM039015.D
 Acq On : 16 Mar 2023 12:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 17 05:49:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 05:25:34 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.367	0.418	-13.9	121	0.00
49	Acenaphthylene	2.035	2.021	0.7	113	0.00
50	Dimethylphthalate	1.539	1.541	-0.1	114	0.00
51	2,6-Dinitrotoluene	0.313	0.328	-4.8	114	0.00
52 C	Acenaphthene	1.251	1.204	3.8	111	0.00
53	3-Nitroaniline	0.363	0.384	-5.8	112	0.00
54 P	2,4-Dinitrophenol	0.169	0.184	-8.9	121	0.00
55	Dibenzofuran	1.957	1.880	3.9	111	0.00
56 P	4-Nitrophenol	0.306	0.306	0.0	110	0.00
57	2,4-Dinitrotoluene	0.401	0.433	-8.0	112	0.00
58	Fluorene	1.533	1.509	1.6	111	0.00
59	2,3,4,6-Tetrachlorophenol	0.380	0.372	2.1	109	0.00
60	Diethylphthalate	1.497	1.562	-4.3	115	0.00
61	4-Chlorophenyl-phenylether	0.753	0.747	0.8	113	0.00
62	4-Nitroaniline	0.362	0.389	-7.5	110	0.00
63	Azobenzene	1.584	1.696	-7.1	118	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.127	-5.8	117	0.00
66 c	n-Nitrosodiphenylamine	0.674	0.664	1.5	112	0.00
67	4-Bromophenyl-phenylether	0.216	0.215	0.5	113	0.00
68	Hexachlorobenzene	0.241	0.228	5.4	109	0.00
69	Atrazine	0.184	0.194	-5.4	112	0.00
70 C	Pentachlorophenol	0.178	0.170	4.5	107	0.00
71	Phenanthrene	1.174	1.125	4.2	109	0.00
72	Anthracene	1.171	1.154	1.5	109	0.00
73	Carbazole	1.068	1.057	1.0	107	0.00
74	Di-n-butylphthalate	1.187	1.343	-13.1	118	0.00
75 C	Fluoranthene	1.239	1.204	2.8	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzidine	0.417	0.451	-8.2	101	0.00
78	Pyrene	1.502	1.513	-0.7	103	0.00
79 S	Terphenyl-d14	1.094	1.133	-3.6	104	0.00
80	Butylbenzylphthalate	0.558	0.677	-21.3	114	0.00
81	Benzo(a)anthracene	1.440	1.438	0.1	100	0.00
82	3,3'-Dichlorobenzidine	0.443	0.451	-1.8	101	0.00
83	Chrysene	1.401	1.361	2.9	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.819	1.025	-25.2#	115	0.00
85 c	Di-n-octyl phthalate	1.414	1.664	-17.7	114	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	1.522	1.480	2.8	92	0.00
88	Benzo(b)fluoranthene	1.278	1.306	-2.2	98	0.00
89	Benzo(k)fluoranthene	1.330	1.285	3.4	94	0.00
90 C	Benzo(a)pyrene	1.230	1.230	0.0	95	0.00
91	Dibenzo(a,h)anthracene	1.289	1.257	2.5	93	0.00
92	Benzo(g,h,i)perylene	1.261	1.214	3.7	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031623\
Data File : BM039015.D
Acq On : 16 Mar 2023 12:36
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 17 05:49:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 05:25:34 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0