


Data Path : Z:\HPCHEM1\BNA_M\Data\BM031716\
 Data File : BM004656.D
 Acq On : 17 Mar 2016 08:47
 Operator : UM/SJ
 Sample : DFTPP37
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP37

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM031516.M
 Title : SVOA CALIBRATION
 Last Update : Thu Mar 17 01:59:21 2016



AutoFind: Scans 2544, 2545, 2546; Background Corrected with Scan 2539

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.6	52650	PASS
68	69	0.00	2	1.6	871	PASS
69	198	0.00	100	33.8	54586	PASS
70	69	0.00	2	0.4	210	PASS
127	198	10	80	45.3	73157	PASS
197	198	0.00	2	0.5	749	PASS
198	198	100	100	100.0	161536	PASS
199	198	5	9	6.8	10923	PASS
275	198	10	60	24.7	39893	PASS
365	198	1	100	3.1	4975	PASS
441	443	0.01	100	79.0	21395	PASS
442	198	50	100	85.5	138189	PASS
443	442	15	24	19.6	27073	PASS

m/z	Abundance
38.00	465.0
39.10	3111.0
40.00	302.0
44.00	705.0
49.10	308.0
50.10	11719.0
51.10	46456.0
52.10	2484.0
55.00	302.0
56.00	1346.0
57.00	3307.0
61.00	668.0
62.00	664.0
63.00	1915.0
64.00	337.0
65.00	921.0
68.00	732.0
69.00	47376.0
70.00	329.0
73.00	440.0
74.00	4848.0
75.00	7979.0
76.10	2831.0
77.10	58688.0
78.10	4360.0
79.00	3449.0
80.00	2799.0
81.00	3676.0
82.00	1018.0
83.00	968.0
85.00	727.0
86.00	1132.0
87.00	542.0
88.00	301.0
91.00	916.0
92.00	900.0
93.00	6381.0
94.00	421.0
96.10	470.0
98.00	4792.0
99.00	4709.0
100.00	435.0
101.00	2181.0
103.00	690.0
104.00	1409.0
105.00	1270.0
106.00	441.0
107.00	18256.0
108.00	2674.0
109.10	467.0
110.00	32448.0
111.00	5196.0
112.00	647.0
116.00	803.0
117.00	14537.0
118.00	1073.0
122.00	1031.0
123.00	1749.0
124.00	809.0
125.00	666.0
127.00	59824.0
128.00	4786.0
129.00	22672.0
130.00	2054.0
131.00	369.0
134.00	631.0
135.00	1787.0
136.00	752.0
137.00	1240.0
141.00	2919.0
142.00	964.0
143.00	585.0
146.00	541.0
147.00	1693.0
148.00	3573.0
149.00	1017.0
153.00	904.0
154.00	744.0
155.00	1862.0
156.00	2747.0
157.10	475.0
158.00	499.0
159.00	375.0
160.00	953.0
161.00	1374.0
162.00	454.0
165.00	1016.0
166.00	825.0
167.00	6464.0
168.00	4605.0
169.00	707.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

171.90	523.0
173.00	636.0
174.00	1246.0
175.10	2487.0
176.00	820.0
177.00	1185.0
178.00	462.0
179.00	4353.0
180.00	3480.0
181.00	1463.0
184.00	385.0
185.00	2171.0
186.10	17088.0
187.00	5117.0
188.00	567.0
189.00	1014.0
190.90	625.0
192.00	1522.0
193.00	1762.0
194.00	434.0
196.00	3371.0
196.90	843.0
198.00	121856.0
199.00	8789.0
200.00	655.0
201.40	462.0
203.00	702.0
204.10	4522.0
205.10	7618.0
206.10	31816.0
207.10	5419.0
208.00	1312.0
209.00	481.0
210.00	462.0
211.00	1176.0
215.00	363.0
216.00	745.0
217.00	8233.0
218.00	1079.0
221.00	4430.0
221.80	1238.0
223.00	1735.0
224.10	17320.0
225.10	4323.0
226.10	487.0
227.00	7264.0
228.00	1136.0
229.00	1495.0
231.00	604.0
234.00	445.0
235.00	533.0
235.90	316.0
237.10	834.0
241.00	415.0
242.00	850.0
243.10	925.0
244.10	13071.0
245.00	1759.0
246.00	2458.0
247.00	522.0
249.00	445.0
254.10	509.0
255.10	63888.0
256.00	9533.0
257.00	947.0
258.00	3700.0
259.00	768.0
265.00	1442.0
273.00	1897.0
274.00	5305.0
275.10	27840.0
276.00	3780.0
277.00	2330.0
278.00	397.0
283.10	331.0
285.10	469.0
293.00	604.0
296.00	7692.0
297.00	1155.0
303.00	940.0
314.00	312.0
315.00	794.0
316.10	589.0
323.10	2062.0
324.10	464.0
326.90	437.0
334.10	1690.0
335.00	394.0
345.90	437.0
352.10	720.0
353.00	495.0
354.10	771.0
365.00	3229.0
366.00	565.0
372.10	1337.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

373.00	334.0
402.00	505.0
403.00	651.0
421.00	512.0
422.10	555.0
423.10	4284.0
424.00	1525.0
441.10	11643.0
442.10	78760.0
443.10	16156.0
444.10	1491.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

m/z	Abundance
38.00	616.0
39.10	4054.0
39.90	372.0
44.00	598.0
49.00	425.0
50.10	14979.0
51.10	60584.0
52.00	3066.0
56.00	1764.0
57.00	4020.0
61.00	901.0
62.00	815.0
63.00	2493.0
64.00	331.0
65.10	1230.0
68.00	1080.0
69.00	63176.0
70.00	303.0
73.10	590.0
74.10	6785.0
75.00	10392.0
76.10	3685.0
77.10	79368.0
78.10	5558.0
79.00	4693.0
80.00	3706.0
81.00	5178.0
82.00	1470.0
83.00	1277.0
85.00	1167.0
86.00	1542.0
87.00	727.0
88.00	325.0
91.00	1204.0
92.00	1400.0
93.00	8466.0
94.00	563.0
96.00	631.0
98.00	6609.0
99.00	5904.0
100.00	531.0
101.00	2916.0
103.00	1045.0
104.00	1826.0
105.00	1772.0
106.00	603.0
107.00	24880.0
108.00	3873.0
109.00	675.0
110.00	45232.0
111.00	6672.0
112.00	961.0
116.00	1230.0
117.00	20168.0
118.00	1560.0
120.00	335.0
122.00	1361.0
123.00	2511.0
124.00	1126.0
125.00	1022.0
126.20	544.0
127.00	84704.0
128.00	6032.0
129.00	31712.0
130.00	2708.0
131.00	607.0
134.00	1027.0
135.00	2750.0
136.00	1084.0
137.10	1708.0
140.00	518.0
141.00	4140.0
142.00	1390.0
143.00	976.0
146.10	786.0
147.00	2243.0
148.00	5263.0
149.00	1230.0
150.00	313.0
151.10	534.0
153.00	1377.0
154.00	1089.0
155.00	2372.0
156.00	3435.0
157.00	706.0
158.00	748.0
159.00	672.0
160.00	1500.0
161.00	1987.0
161.90	555.0
165.00	1544.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

166.00	1204.0
167.00	9438.0
168.00	5879.0
169.00	968.0
170.10	343.0
172.00	844.0
173.00	1016.0
174.00	1930.0
175.10	3782.0
176.10	901.0
177.00	1649.0
178.00	555.0
179.00	6241.0
180.00	4925.0
181.00	2224.0
182.00	381.0
184.00	514.0
185.10	3133.0
186.10	25080.0
187.10	7192.0
188.00	742.0
189.00	1468.0
189.90	310.0
191.00	994.0
192.00	2186.0
193.10	2443.0
194.00	515.0
196.00	5042.0
197.10	1036.0
198.00	185920.0
199.00	12439.0
200.00	925.0
201.50	816.0
203.00	1190.0
204.00	6616.0
205.10	11600.0
206.10	48096.0
207.10	8175.0
208.00	1823.0
209.00	792.0
210.20	697.0
211.10	1731.0
215.00	529.0
216.00	944.0
217.00	12677.0
218.00	1495.0
221.10	7412.0
221.80	1659.0
223.00	2803.0
224.10	27048.0
225.10	6948.0
226.10	884.0
227.00	10849.0
228.00	1797.0
229.00	2517.0
229.90	302.0
231.10	931.0
234.00	610.0
235.10	795.0
236.00	510.0
237.00	1053.0
239.00	491.0
240.00	379.0
241.00	533.0
242.10	1438.0
243.10	1650.0
244.10	21112.0
245.10	2648.0
246.00	3985.0
247.00	721.0
249.00	763.0
253.10	465.0
254.10	778.0
255.10	104504.0
256.10	15384.0
257.00	1273.0
258.10	6038.0
259.00	906.0
265.00	2485.0
266.00	305.0
272.10	359.0
273.00	3290.0
274.10	8225.0
275.10	46216.0
276.10	6358.0
277.00	3944.0
278.00	715.0
283.10	581.0
284.10	335.0
285.10	722.0
293.00	945.0
296.00	12817.0
297.00	1761.0
303.10	1475.0
304.00	394.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

314.00	592.0
315.00	1404.0
316.00	785.0
321.00	398.0
323.10	4280.0
324.10	768.0
327.00	764.0
328.00	386.0
333.00	418.0
334.10	2734.0
335.00	660.0
341.00	461.0
346.00	900.0
352.00	1347.0
353.00	780.0
354.10	1318.0
365.00	5620.0
366.00	831.0
371.00	371.0
372.10	2358.0
373.10	582.0
383.00	587.0
402.00	910.0
403.00	1249.0
404.00	516.0
421.10	1076.0
422.00	1203.0
423.10	7879.0
424.10	2423.0
441.10	23912.0
442.10	152832.0
443.10	30200.0
444.10	2858.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

m/z	Abundance
38.10	453.0
39.10	3534.0
40.00	341.0
44.00	596.0
49.00	302.0
50.10	12277.0
51.10	50912.0
52.10	2475.0
56.00	1513.0
57.00	3436.0
61.00	658.0
62.00	735.0
63.00	2119.0
65.10	1109.0
68.00	801.0
69.00	53208.0
73.00	571.0
74.00	5448.0
75.00	8684.0
76.10	3240.0
77.10	67504.0
78.10	4711.0
79.00	3889.0
80.00	3139.0
81.00	4480.0
82.00	1128.0
83.10	1080.0
85.00	945.0
86.00	1376.0
87.00	577.0
91.00	1033.0
92.00	1158.0
93.00	7239.0
94.00	496.0
96.00	528.0
98.00	5748.0
99.00	4960.0
100.00	477.0
101.00	2603.0
102.90	754.0
104.00	1608.0
105.00	1595.0
106.00	517.0
107.00	21512.0
108.00	3389.0
109.00	572.0
110.00	38880.0
111.00	5746.0
112.00	739.0
116.00	1133.0
117.00	17800.0
118.00	1343.0
122.00	1369.0
123.00	2249.0
123.90	994.0
125.00	854.0
126.20	422.0
127.00	74944.0
128.00	5781.0
129.00	27728.0
130.00	2284.0
131.00	476.0
134.00	729.0
135.00	2194.0
136.10	916.0
137.00	1317.0
140.00	348.0
141.00	3719.0
142.00	1268.0
143.00	765.0
146.00	694.0
147.00	2021.0
148.00	4699.0
149.00	1066.0
151.10	585.0
153.00	1244.0
154.00	925.0
155.10	2106.0
156.10	3352.0
157.10	703.0
158.00	620.0
159.00	539.0
160.00	1128.0
161.00	2117.0
162.00	555.0
165.00	1425.0
166.00	1129.0
167.00	8337.0
168.00	5080.0
169.00	818.0
171.00	314.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

172.00	707.0
173.00	1009.0
174.00	1920.0
175.00	3363.0
176.10	820.0
177.00	1497.0
177.90	523.0
179.00	6216.0
180.00	4468.0
181.10	1977.0
182.10	322.0
184.00	627.0
185.10	2928.0
186.10	23984.0
187.10	6621.0
188.00	780.0
189.00	1418.0
191.00	867.0
192.00	1897.0
193.00	2516.0
194.00	407.0
196.00	4918.0
197.10	1212.0
198.00	176832.0
199.00	11541.0
200.00	831.0
201.40	706.0
203.00	1113.0
204.10	6287.0
205.10	10502.0
206.10	46616.0
207.10	7436.0
208.00	1762.0
209.00	730.0
210.10	717.0
211.00	1576.0
214.90	449.0
216.00	919.0
217.00	11738.0
218.00	1625.0
221.10	6809.0
221.80	1753.0
223.00	2762.0
224.10	25304.0
225.10	6622.0
226.10	802.0
227.10	10298.0
228.00	1444.0
229.00	2383.0
230.00	327.0
231.10	802.0
234.00	565.0
235.00	733.0
236.00	552.0
237.00	870.0
239.00	425.0
239.90	332.0
241.00	562.0
242.00	1434.0
243.10	1468.0
244.10	20088.0
245.10	2758.0
246.00	3889.0
247.00	873.0
249.00	616.0
253.00	438.0
254.20	946.0
255.10	103480.0
256.10	14700.0
257.00	1186.0
258.00	5894.0
259.00	1070.0
265.00	2582.0
266.00	305.0
273.00	3007.0
274.10	8316.0
275.10	45624.0
276.10	6117.0
277.00	3754.0
278.00	718.0
283.10	462.0
284.00	304.0
285.00	686.0
293.00	909.0
295.00	327.0
296.00	13348.0
297.00	1622.0
303.10	1470.0
304.00	397.0
314.00	578.0
315.00	1526.0
316.00	823.0
321.10	351.0
323.10	4564.0
324.10	736.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

327.00	745.0
328.00	366.0
332.00	319.0
333.00	442.0
334.10	2771.0
335.10	844.0
341.00	487.0
346.00	917.0
352.10	1228.0
353.00	961.0
354.10	1474.0
365.00	6076.0
366.00	945.0
371.00	391.0
372.10	2163.0
373.10	545.0
383.10	618.0
390.10	312.0
402.00	1015.0
403.10	1483.0
404.00	512.0
421.00	1469.0
422.10	1260.0
423.10	8657.0
424.10	2572.0
441.10	28632.0
442.10	182976.0
443.10	34864.0
444.00	3069.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37

Scan 2539 (17.621 min): BM004656.D
DFTPP37

m/z	Abundance
44.00	814.0
73.00	535.0
207.10	1546.0

Instrument :
BNA_M
ClientSampleId :
DFTPP37