

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031816\
 Data File : BM004689.D
 Acq On : 18 Mar 2016 19:33
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 14:14:00 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.494	0.528	-6.9	85	0.00
3	n-Nitrosodimethylamine	0.517	0.604	-16.8	97	0.00
4 S	2-Fluorophenol	1.154	1.279	-10.8	92	0.00
5 S	Phenol-d6	1.413	1.490	-5.4	90	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
7 S	Nitrobenzene-d5	0.241	0.253	-5.0	89	0.00
8	Naphthalene	1.375	1.544	-12.3	94	0.00
9	2-Methylnaphthalene	0.861	0.986	-14.5	95	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
11 S	2,4,6-Tribromophenol	0.166	0.160	3.6	82	0.00
12 S	2-Fluorobiphenyl	1.431	1.679	-17.3	95	0.00
13	Acenaphthylene	2.572	2.734	-6.3	91	0.00
14	Acenaphthene	1.702	1.962	-15.3	95	0.00
15	Fluorene	1.908	2.151	-12.7	94	-0.02
16 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
17	Hexachlorobenzene	0.308	0.351	-14.0	90	0.00
18	Pentachlorophenol	0.169	0.163	3.6	86	0.00
19	Phenanthrene	2.004	2.406	-20.1	96	0.00
20	Anthracene	1.429	1.531	-7.1	88	0.00
21	Fluoranthene	1.895	2.116	-11.7	91	-0.02
22 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
23	Pyrene	1.835	1.830	0.3	90	0.00
24 S	Terphenyl-d14	0.632	0.649	-2.7	93	0.00
25	Benzo(a)anthracene	2.000	1.869	6.6	82	0.00
26	Chrysene	1.487	1.683	-13.2	105	0.00
27	Indeno(1,2,3-cd)pyrene	1.674	1.731	-3.4	92	0.00
28 I	Perylene-d12	1.000	1.000	0.0	95	0.00
29	Benzo(b)fluoranthene	1.819	2.066	-13.6	96	0.00
30	Benzo(k)fluoranthene	1.802	1.981	-9.9	91	0.00
31 C	Benzo(a)pyrene	1.706	1.837	-7.7	90	0.00
32	Dibenzo(a,h)anthracene	1.527	1.681	-10.1	92	0.00
33	Benzo(g,h,i)perylene	1.613	1.794	-11.2	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0