

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031816\
 Data File : BM004698.D
 Acq On : 19 Mar 2016 01:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 01:51:22 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.494	0.594	-20.2	93	0.00
3	n-Nitrosodimethylamine	0.517	0.560	-8.3	87	0.00
4 S	2-Fluorophenol	1.154	1.193	-3.4	83	0.00
5 S	Phenol-d6	1.413	1.377	2.5	80	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.241	0.232	3.7	79	0.00
8	Naphthalene	1.375	1.554	-13.0	91	0.00
9	2-Methylnaphthalene	0.861	0.980	-13.8	90	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
11 S	2,4,6-Tribromophenol	0.166	0.161	3.0	76	0.00
12 S	2-Fluorobiphenyl	1.431	1.760	-23.0	92	0.00
13	Acenaphthylene	2.572	2.653	-3.1	81	0.00
14	Acenaphthene	1.702	1.977	-16.2	88	-0.02
15	Fluorene	1.908	2.412	-26.4#	97	-0.02
16 I	Phenanthrene-d10	1.000	1.000	0.0	114	-0.02
17	Hexachlorobenzene	0.308	0.259	15.9	81	-0.02
18	Pentachlorophenol	0.169	0.091	46.2#	58	0.00
19	Phenanthrene	2.004	1.637	18.3	79	0.00
20	Anthracene	1.429	1.352	5.4	94	-0.02
21	Fluoranthene	1.895	1.700	10.3	89	-0.01
22 I	Chrysene-d12	1.000	1.000	0.0	85	-0.01
23	Pyrene	1.835	2.164	-17.9	88	-0.01
24 S	Terphenyl-d14	0.632	0.741	-17.2	87	-0.01
25	Benzo(a)anthracene	2.000	1.964	1.8	71	0.01
26	Chrysene	1.487	2.318	-55.9#	120	0.00
27	Indeno(1,2,3-cd)pyrene	1.674	2.007	-19.9	88	-0.01
28 I	Perylene-d12	1.000	1.000	0.0	94	0.00
29	Benzo(b)fluoranthene	1.819	2.061	-13.3	95	0.00
30	Benzo(k)fluoranthene	1.802	1.895	-5.2	86	-0.01
31 C	Benzo(a)pyrene	1.706	1.744	-2.2	85	-0.01
32	Dibenzo(a,h)anthracene	1.527	1.617	-5.9	88	-0.01
33	Benzo(g,h,i)perylene	1.613	1.761	-9.2	90	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0