

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031816\
 Data File : BM004698.D
 Acq On : 19 Mar 2016 01:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 01:51:22 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	1,4-Dioxane	0.400	0.473	-18.2	93	0.00
3	n-Nitrosodimethylamine	0.400	0.434	-8.5	87	0.00
4 S	2-Fluorophenol	0.400	0.414	-3.5	83	0.00
5 S	Phenol-d6	0.400	0.390	2.5	80	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.400	0.385	3.8	79	0.00
8	Naphthalene	0.400	0.452	-13.0	91	0.00
9	2-Methylnaphthalene	0.400	0.455	-13.7	90	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	89	0.00
11 S	2,4,6-Tribromophenol	0.400	0.388	3.0	76	0.00
12 S	2-Fluorobiphenyl	0.400	0.492	-23.0	92	0.00
13	Acenaphthylene	0.400	0.413	-3.2	81	0.00
14	Acenaphthene	0.400	0.465	-16.3	88	-0.02
15	Fluorene	0.400	0.506	-26.5#	97	-0.02
16 I	Phenanthrene-d10	5.000	5.000	0.0	114	-0.02
17	Hexachlorobenzene	0.400	0.337	15.8	81	-0.02
18	Pentachlorophenol	0.400	0.311	22.3	58	0.00
19	Phenanthrene	0.400	0.327	18.3	79	0.00
20	Anthracene	0.400	0.378	5.5	94	-0.02
21	Fluoranthene	0.400	0.359	10.3	89	-0.01
22 I	Chrysene-d12	5.000	5.000	0.0	85	-0.01
23	Pyrene	0.400	0.472	-18.0	88	-0.01
24 S	Terphenyl-d14	0.400	0.469	-17.2	87	-0.01
25	Benzo(a)anthracene	0.400	0.393	1.8	71	0.01
26	Chrysene	0.400	0.624	-56.0#	120	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.480	-20.0	88	-0.01
28 I	Perylene-d12	5.000	5.000	0.0	94	0.00
29	Benzo(b)fluoranthene	0.400	0.453	-13.2	95	0.00
30	Benzo(k)fluoranthene	0.400	0.421	-5.2	86	-0.01
31 C	Benzo(a)pyrene	0.400	0.409	-2.2	85	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.424	-6.0	88	-0.01
33	Benzo(g,h,i)perylene	0.400	0.437	-9.2	90	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0