

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031915\
 Data File : BM000730.D
 Acq On : 19 Mar 2015 15:12
 Operator : TP/IZ
 Sample : SSTD3.201
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.201

Quant Time: Mar 19 15:53:51 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:05:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	2966	0.40	ng/ul	0.00
2) Naphthalene-d8	10.41	136	10436	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.27	164	5347	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.02	188	10310	0.40	ng/ul	0.00
16) Chrysene-d12	21.22	240	8919	0.40	ng/ul	0.00
20) Perylene-d12	23.41	264	7149	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.00	152	40982	3.42	ng/ul	0.00
14) Fluoranthene-d10	19.06	212	71820	3.52	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.46	128	84301	3.39	ng/ul	98
5) 2-Methylnaphthalene	12.08	142	55895	3.53	ng/ul	99
7) Acenaphthylene	14.00	152	78328	4.22	ng/ul	98
8) Acenaphthene	14.33	153	57257	3.82	ng/ul	98
9) Fluorene	15.34	166	65486	3.88	ng/ul	97
11) Pentachlorophenol	16.68	266	5022	4.11	ng/ul	97
12) Phenanthrene	17.05	178	98594	3.58	ng/ul	98
13) Anthracene	17.15	178	91534	3.78	ng/ul	97
15) Fluoranthene	19.09	202	110388	3.62	ng/ul	98
17) Pyrene	19.45	202	109599	3.44	ng/ul	99
18) Benzo(a)anthracene	21.20	228	89500	3.61	ng/ul	99
19) Chrysene	21.25	228	95521	3.26	ng/ul	99
21) Benzo(b)fluoranthene	22.75	252	89328	3.67	ng/ul	95
22) Benzo(k)fluoranthene	22.79	252	86399	3.44	ng/ul	94
23) Benzo(a)pyrene	23.31	252	80640	3.63	ng/ul	94
24) Indeno(1,2,3-cd)pyrene	25.60	276	92094	3.58	ng/ul	100
25) Dibenzo(a,h)anthracene	25.61	278	74923	3.60	ng/ul	94
26) Benzo(g,h,i)perylene	26.26	276	79443	3.47	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031915\
Data File : BM000730.D
Acq On : 19 Mar 2015 15:12
Operator : TP/IZ
Sample : SSTD3.201
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
Client Sample ID :
SSTD3.201

Quant Time: Mar 19 15:53:51 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:05:41 2015
Response via : Initial Calibration

