

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031915\
 Data File : BM000732.D
 Acq On : 19 Mar 2015 17:34
 Operator : TP/IZ
 Sample : G1452-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 X1K80

Quant Time: Mar 20 11:03:43 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:39:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	2568	0.40	ng/ul	0.00
2) Naphthalene-d8	10.41	136	9053	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.27	164	4967	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.02	188	8876	0.40	ng/ul	0.00
16) Chrysene-d12	21.22	240	7314	0.40	ng/ul	0.00
20) Perylene-d12	23.41	264	6681	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.00	152	3391	0.33	ng/ul	0.00
14) Fluoranthene-d10	19.07	212	6660	0.38	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.46	128	12568	0.58	ng/ul	100
5) 2-Methylnaphthalene	12.08	142	9209	0.67	ng/ul	100
7) Acenaphthylene	14.00	152	11379	0.66	ng/ul	99
8) Acenaphthene	14.33	153	10940	0.78	ng/ul	99
11) Pentachlorophenol	16.68	266	564	0.51	ng/ul	96
13) Anthracene	17.15	178	12765	0.61	ng/ul	99
17) Pyrene	19.45	202	24488	0.94	ng/ul	100
19) Chrysene	21.26	228	16641	0.69	ng/ul#	88
21) Benzo(b)fluoranthene	22.75	252	18892	0.83	ng/ul	97
22) Benzo(k)fluoranthene	22.79	252	29378	1.25	ng/ul	97
25) Dibenzo(a,h)anthracene	25.62	278	16698	0.86	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031915\
Data File : BM000732.D
Acq On : 19 Mar 2015 17:34
Operator : TP/IZ
Sample : G1452-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
X1K80

Quant Time: Mar 20 11:03:43 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:39:49 2015
Response via : Initial Calibration

