

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031924\
 Data File : BM044673.D
 Acq On : 19 Mar 2024 21:12
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020077

Quant Time: Mar 19 23:32:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 15 11:12:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	192532	20.000	ng/u1	0.00
20) Naphthalene-d8	10.475	136	786957	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.339	164	623035	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.098	188	1313677	20.000	ng/u1	0.00
79) Chrysene-d12	21.291	240	1193866	20.000	ng/u1	0.00
88) Perylene-d12	23.538	264	1400056	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.257	96	43761	7.709	ng/uL	0.00
4) Pyridine-d5	3.651	84	326221	19.525	ng/u1	0.00
7) Phenol-d5	6.869	99	373240	18.856	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.045	67	237203	19.400	ng/u1	0.00
11) 2-Chlorophenol-d4	7.228	132	287418	19.759	ng/u1	0.00
15) 4-Methylphenol-d8	8.398	113	299204	19.247	ng/u1	0.00
21) Nitrobenzene-d5	8.851	128	138017	20.827	ng/u1	0.00
24) 2-Nitrophenol-d4	9.569	143	144636	20.247	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.098	165	282173	23.209	ng/u1	0.00
31) 4-Chloroaniline-d4	10.622	131	438407	21.711	ng/u1	0.00
46) Dimethylphthalate-d6	13.763	166	1027318	21.020	ng/u1	0.00
49) Acenaphthylene-d8	14.033	160	1068305	19.113	ng/u1	0.00
54) 4-Nitrophenol-d4	14.557	143	202631	19.993	ng/u1	0.00
60) Fluorene-d10	15.339	176	852162	20.749	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.468	200	171152	20.071	ng/u1	0.00
73) Anthracene-d10	17.198	188	1301004	20.719	ng/u1	0.00
81) Pyrene-d10	19.497	212	1463565	19.988	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.397	264	1487869	20.219	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.287	88	46059	7.837	ng/uL	98
5) Pyridine	3.669	79	340611	19.655	ng/u1	96
6) Benzaldehyde	6.851	77	210476	24.412	ng/u1	95
8) Phenol	6.892	94	383841	19.000	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.134	93	306759	18.492	ng/u1	97
12) 2-Chlorophenol	7.257	128	293110	19.536	ng/u1	98
13) 2-Methylphenol	8.134	108	284198	18.538	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.216	45	452189	20.641	ng/u1	98
16) Acetophenone	8.522	105	458412	19.013	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.504	70	273509	21.957	ng/u1	94
18) 4-Methylphenol	8.463	108	318047	19.058	ng/u1	96
19) Hexachloroethane	8.751	117	125189	20.370	ng/u1	97
22) Nitrobenzene	8.892	77	337401	21.470	ng/u1	95
23) Isophorone	9.416	82	816106	25.454	ng/u1	98
25) 2-Nitrophenol	9.598	139	160752	20.641	ng/u1	97
26) 2,4-Dimethylphenol	9.657	107	327467	23.108	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.904	93	430046	21.327	ng/u1	99
29) 2,4-Dichlorophenol	10.122	162	281057	23.155	ng/u1	99
30) Naphthalene	10.522	128	912728	20.478	ng/u1	99
32) 4-Chloroaniline	10.645	127	429021	22.039	ng/u1	100
33) Hexachlorobutadiene	10.792	225	151521	22.272	ng/u1	96
34) Caprolactam	11.433	113	67915	14.368	ng/u1	96
35) 4-Chloro-3-methylphenol	11.769	107	382491	27.173	ng/u1	99
36) 2-Methylnaphthalene	12.145	142	632563	21.342	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031924\
 Data File : BM044673.D
 Acq On : 19 Mar 2024 21:12
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020077


Quant Time: Mar 19 23:32:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 15 11:12:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.363	142	629169	21.518	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.510	216	311298	17.563	ng/ul	99
40) Hexachlorocyclopentadiene	12.480	237	181201	15.400	ng/ul	98
41) 2,4,6-Trichlorophenol	12.763	196	252718	20.338	ng/ul	99
42) 2,4,5-Trichlorophenol	12.833	196	283522	21.291	ng/ul	98
43) 1,1'-Biphenyl	13.169	154	925650	18.710	ng/ul	99
44) 2-Chloronaphthalene	13.210	162	694009	17.876	ng/ul	99
45) 2-Nitroaniline	13.427	65	266147	21.836	ng/ul	99
47) Dimethylphthalate	13.810	163	1042065	20.954	ng/ul	99
48) 2,6-Dinitrotoluene	13.933	165	206547	20.793	ng/ul	97
50) Acenaphthylene	14.063	152	1282150	19.681	ng/ul	99
51) 3-Nitroaniline	14.263	138	220380	20.539	ng/ul	96
52) Acenaphthene	14.404	153	858863	20.054	ng/ul	99
53) 2,4-Dinitrophenol	14.468	184	117293	18.459	ng/ul	92
55) 4-Nitrophenol	14.568	109	154777	21.795	ng/ul	93
56) Dibenzofuran	14.745	168	1185975	20.819	ng/ul	99
57) 2,4-Dinitrotoluene	14.721	165	299212	20.867	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	14.968	232	237612	21.562	ng/ul	99
59) Diethylphthalate	15.180	149	1073470	20.799	ng/ul	99
61) Fluorene	15.398	166	1006430	21.601	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.398	204	483214	22.346	ng/ul	98
63) 4-Nitroaniline	15.433	138	208657	20.347	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.486	198	188732	20.711	ng/ul	100
67) N-Nitrosodiphenylamine	15.615	169	834788	21.099	ng/ul	99
68) 4-Bromophenyl-phenylether	16.292	248	288615	21.965	ng/ul	98
69) Hexachlorobenzene	16.392	284	332082	22.189	ng/ul	94
70) Atrazine	16.574	200	182997	14.154	ng/ul	98
71) Pentachlorophenol	16.739	266	200325	20.271	ng/ul	98
72) Phenanthrene	17.139	178	1551232	20.816	ng/ul	100
74) Anthracene	17.227	178	1578746	20.948	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.127	216	323626	18.444	ng/uL	99
76) Pentachlorobenzene	14.651	250	372081	21.934	ng/uL	98
77) Carbazole	17.509	167	1460910	20.628	ng/ul	99
78) Di-n-butylphthalate	18.080	149	1887300	19.812	ng/ul	99
80) Fluoranthene	19.162	202	1734592	19.892	ng/ul	98
82) Pyrene	19.527	202	1832965	20.190	ng/ul	96
83) Butylbenzylphthalate	20.445	149	821422	18.401	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.221	252	513692	19.891	ng/ul	99
85) Benzo(a)anthracene	21.280	228	1818417	20.283	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.227	149	1277906	19.634	ng/ul	99
87) Chrysene	21.333	228	1674217	20.170	ng/ul	100
89) Di-n-octyl phthalate	22.115	149	2216023	17.686	ng/ul	100
90) Benzo(b)fluoranthene	22.868	252	1792325	19.132	ng/ul	99
91) Benzo(k)fluoranthene	22.909	252	1848352	20.145	ng/ul	99
93) Benzo(a)pyrene	23.444	252	1738466	19.838	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.815	276	2197622	22.358	ng/ul	100
95) Dibenzo(a,h)anthracene	25.826	278	1850694	22.581	ng/ul	99
96) Benzo(g,h,i)perylene	26.497	276	1754964	22.374	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 19 23:32:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 15 11:12:49 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.837 ng/uL
RT: 3.287 min Scan# 34
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion: 88 Resp: 46059
Ion Ratio Lower Upper
88 100
43 28.8 23.8 35.6
58 75.7 61.6 92.4



#5
Pyridine
Concen: 19.655 ng/uL
RT: 3.669 min Scan# 99
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion: 79 Resp: 340611
Ion Ratio Lower Upper
79 100
52 75.0 57.4 86.0
51 35.7 26.6 39.8









#6
Benzaldehyde
Concen: 24.412 ng/ul
RT: 6.851 min Scan# 641
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion: 77 Resp: 210476
Ion Ratio Lower Upper
77 100
105 97.0 81.4 122.0
106 95.5 80.2 120.4



#8
Phenol
Concen: 19.000 ng/ul
RT: 6.892 min Scan# 647
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion: 94 Resp: 383841
Ion Ratio Lower Upper
94 100
65 28.8 21.3 31.9
66 35.6 25.4 38.2









#10
Bis(2-Chloroethyl)ether
Concen: 18.492 ng/ul
RT: 7.134 min Scan# 689
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion: 93 Resp: 306759
Ion Ratio Lower Upper
93 100
63 78.6 60.2 90.4
95 32.3 26.8 40.2



#12
2-Chlorophenol
Concen: 19.536 ng/ul
RT: 7.257 min Scan# 709
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion: 128 Resp: 293110
Ion Ratio Lower Upper
128 100
64 49.1 37.4 56.2
130 32.3 25.4 38.0





#13

2-Methylphenol

Concen: 18.538 ng/ul

RT: 8.134 min Scan# 858

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :





SSTD020077

Tgt Ion:108 Resp: 284198

Ion Ratio Lower Upper

108 100

107 85.7 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.641 ng/ul

RT: 8.216 min Scan# 872

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion: 45 Resp: 452189


Ion Ratio Lower Upper

45 100

77 14.2 12.4 18.6

79 11.0 9.0 13.6









#16
Acetophenone
Concen: 19.013 ng/ul
RT: 8.522 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:105 Resp: 458412
Ion Ratio Lower Upper
105 100
77 77.8 59.5 89.3
51 28.0 20.2 30.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.957 ng/ul
RT: 8.504 min Scan# 921
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion: 70 Resp: 273509
Ion Ratio Lower Upper
70 100
42 59.5 42.7 64.1
101 9.6 7.1 10.7
130 20.2 17.3 25.9









#18
4-Methylphenol
Concen: 19.058 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:108 Resp: 318047
Ion Ratio Lower Upper
108 100
107 110.2 91.8 137.8



#19
Hexachloroethane
Concen: 20.370 ng/ul
RT: 8.751 min Scan# 963
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:117 Resp: 125189
Ion Ratio Lower Upper
117 100
201 91.3 69.0 103.6
199 55.5 44.4 66.6









#22
Nitrobenzene
Concen: 21.470 ng/ul
RT: 8.892 min Scan# 987
Delta R.T. -0.012 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion: 77 Resp: 337401
Ion Ratio Lower Upper
77 100
123 43.6 38.2 57.2
65 13.7 10.2 15.2



#23
Isophorone
Concen: 25.454 ng/ul
RT: 9.416 min Scan# 1076
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion: 82 Resp: 816106
Ion Ratio Lower Upper
82 100
95 6.3 4.7 7.1
138 16.4 14.1 21.1









#25
2-Nitrophenol
Concen: 20.641 ng/ul
RT: 9.598 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:139 Resp: 160752
Ion Ratio Lower Upper
139 100
65 52.3 40.2 60.2
109 28.0 21.0 31.4



#26
2,4-Dimethylphenol
Concen: 23.108 ng/ul
RT: 9.657 min Scan# 1117
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:107 Resp: 327467
Ion Ratio Lower Upper
107 100
121 52.5 41.0 61.6
122 87.5 71.4 107.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.327 ng/ul
 RT: 9.904 min Scan# 1159
 Delta R.T. -0.006 min
 Lab File: BM044673.D
 Acq: 19 Mar 2024 21:12




Instrument :
 BNA_M
 ClientSampleId :
 SST020077


Tgt Ion: 93 Resp: 430046
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.3 37.9
 123 11.0 9.3 13.9



#29
 2,4-Dichlorophenol
 Concen: 23.155 ng/ul
 RT: 10.122 min Scan# 1196
 Delta R.T. -0.006 min
 Lab File: BM044673.D
 Acq: 19 Mar 2024 21:12

Tgt Ion: 162 Resp: 281057
 Ion Ratio Lower Upper
 162 100
 164 65.0 52.7 79.1
 98 37.1 30.1 45.1





#30

Naphthalene

Concen: 20.478 ng/ul

RT: 10.522 min Scan# 1265

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:128 Resp: 912728

Ion Ratio Lower Upper

128 100

129 11.3 8.7 13.1

127 12.8 10.5 15.7



#32

4-Chloroaniline

Concen: 22.039 ng/ul

RT: 10.645 min Scan# 1285

Delta R.T. -0.006 min

Lab File: BM044673.D




Acq: 19 Mar 2024 21:12


Tgt Ion:127 Resp: 429021

Ion Ratio Lower Upper

127 100

129 31.4 25.1 37.7





#33

Hexachlorobutadiene

Concen: 22.272 ng/ul

RT: 10.792 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:225 Resp: 151521

Ion Ratio Lower Upper

225 100

223 61.1 50.1 75.1

227 65.7 49.2 73.8



#34

Caprolactam

Concen: 14.368 ng/ul

RT: 11.433 min Scan# 1419

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:113 Resp: 67915


Ion Ratio Lower Upper

113 100

55 176.2 134.7 202.1

56 129.4 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 27.173 ng/ul
RT: 11.769 min Scan# 1476
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:107 Resp: 382491
Ion Ratio Lower Upper
107 100
144 25.9 21.7 32.5
142 80.5 65.1 97.7



#36
2-Methylnaphthalene
Concen: 21.342 ng/ul
RT: 12.145 min Scan# 1540
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:142 Resp: 632563
Ion Ratio Lower Upper
142 100
141 87.9 71.1 106.7









#37
1-Methylnaphthalene
Concen: 21.518 ng/ul
RT: 12.363 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:142 Resp: 629169
Ion Ratio Lower Upper
142 100
141 92.9 73.0 109.6
116 3.8 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.563 ng/ul
RT: 12.510 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:216 Resp: 311298
Ion Ratio Lower Upper
216 100
214 78.6 63.4 95.2
179 21.6 17.5 26.3
108 19.5 15.2 22.8









#40
Hexachlorocyclopentadiene
Concen: 15.400 ng/ul
RT: 12.480 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:237 Resp: 181201
Ion Ratio Lower Upper
237 100
235 63.6 50.1 75.1
272 14.6 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 20.338 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:196 Resp: 252718
Ion Ratio Lower Upper
196 100
198 96.5 76.7 115.1
200 31.6 24.1 36.1









#42
2,4,5-Trichlorophenol
Concen: 21.291 ng/ul
RT: 12.833 min Scan# 1657
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:196 Resp: 283522
Ion Ratio Lower Upper
196 100
198 95.8 78.1 117.1
200 30.2 24.7 37.1



#43
1,1'-Biphenyl
Concen: 18.710 ng/ul
RT: 13.169 min Scan# 1714
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:154 Resp: 925650
Ion Ratio Lower Upper
154 100
153 40.2 31.8 47.8
76 14.4 11.0 16.4





#44

2-Chloronaphthalene

Concen: 17.876 ng/ul

RT: 13.210 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:162 Resp: 694009

Ion Ratio Lower Upper

162 100

164 31.8 25.6 38.4

127 38.4 31.4 47.0



#45

2-Nitroaniline

Concen: 21.836 ng/ul

RT: 13.427 min Scan# 1758

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion: 65 Resp: 266147


Ion Ratio Lower Upper

65 100

92 68.5 55.4 83.0

138 104.6 84.0 126.0





#47

Dimethylphthalate

Concen: 20.954 ng/ul

RT: 13.810 min Scan# 1823

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:163 Resp: 1042065

Ion Ratio Lower Upper

163 100

194 4.0 3.0 4.6

164 10.3 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 20.793 ng/ul

RT: 13.933 min Scan# 1844

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:165 Resp: 206547


Ion Ratio Lower Upper

165 100

89 63.9 48.4 72.6

121 22.2 17.7 26.5









#50
Acenaphthylene
Concen: 19.681 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:152 Resp: 1282150
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2
153 13.0 10.6 15.8



#51
3-Nitroaniline
Concen: 20.539 ng/ul
RT: 14.263 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:138 Resp: 220380
Ion Ratio Lower Upper
138 100
108 9.7 7.3 10.9
92 116.5 89.3 133.9









#52
Acenaphthene
Concen: 20.054 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:153 Resp: 858863
Ion Ratio Lower Upper
153 100
152 48.5 38.2 57.2
154 88.9 70.3 105.5



#53
2,4-Dinitrophenol
Concen: 18.459 ng/ul
RT: 14.468 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:184 Resp: 117293
Ion Ratio Lower Upper
184 100
63 79.2 57.2 85.8
154 64.0 47.9 71.9









#55
4-Nitrophenol
Concen: 21.795 ng/ul
RT: 14.568 min Scan# 1952
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:109 Resp: 154777
Ion Ratio Lower Upper
109 100
139 142.0 126.2 189.4
65 144.6 117.0 175.4



#56
Dibenzofuran
Concen: 20.819 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:168 Resp: 1185975
Ion Ratio Lower Upper
168 100
139 38.0 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 20.867 ng/ul
RT: 14.721 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:165 Resp: 299212
Ion Ratio Lower Upper
165 100
63 53.0 37.4 56.0
182 2.6 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.562 ng/ul
RT: 14.968 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:232 Resp: 237612
Ion Ratio Lower Upper
232 100
131 49.9 40.6 61.0
130 2.6 2.2 3.4
166 32.8 27.0 40.4





#59

Diethylphthalate

Concen: 20.799 ng/ul

RT: 15.180 min Scan# 2057

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:149 Resp: 1073470

Ion Ratio Lower Upper

149 100

177 21.3 16.6 24.8

150 12.0 9.9 14.9



#61

Fluorene

Concen: 21.601 ng/ul

RT: 15.398 min Scan# 2093

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:166 Resp: 1006430


Ion Ratio Lower Upper

166 100

165 98.3 77.3 115.9

167 13.4 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 22.346 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:204 Resp: 483214
Ion Ratio Lower Upper
204 100
206 33.0 26.1 39.1
141 60.6 50.3 75.5



#63
4-Nitroaniline
Concen: 20.347 ng/ul
RT: 15.433 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:138 Resp: 208657
Ion Ratio Lower Upper
138 100
92 52.4 39.8 59.8
108 77.3 54.6 82.0









#66
4,6-Dinitro-2-methylphenol
Concen: 20.711 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:198 Resp: 188732
Ion Ratio Lower Upper
198 100
182 3.8 3.3 4.9
77 26.4 21.0 31.4



#67
N-Nitrosodiphenylamine
Concen: 21.099 ng/ul
RT: 15.615 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:169 Resp: 834788
Ion Ratio Lower Upper
169 100
168 68.7 54.3 81.5
167 37.5 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 21.965 ng/ul
RT: 16.292 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:248 Resp: 288615
Ion Ratio Lower Upper
248 100
250 98.0 78.7 118.1
141 77.8 65.4 98.2



#69
Hexachlorobenzene
Concen: 22.189 ng/ul
RT: 16.392 min Scan# 2262
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:284 Resp: 332082
Ion Ratio Lower Upper
284 100
142 36.5 33.7 50.5
249 29.6 24.8 37.2





#70

Atrazine

Concen: 14.154 ng/ul

RT: 16.574 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:200 Resp: 182997

Ion Ratio Lower Upper

200 100

173 26.9 22.4 33.6

215 47.6 38.9 58.3



#71

Pentachlorophenol

Concen: 20.271 ng/ul

RT: 16.739 min Scan# 2321

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:266 Resp: 200325


Ion Ratio Lower Upper

266 100

264 64.3 51.0 76.4

268 63.3 52.6 78.8





#72

Phenanthrene

Concen: 20.816 ng/ul

RT: 17.139 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:178 Resp: 1551232

Ion Ratio Lower Upper

178 100

179 15.1 12.3 18.5

176 19.4 15.4 23.2



#74

Anthracene

Concen: 20.948 ng/ul

RT: 17.227 min Scan# 2404

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:178 Resp: 1578746


Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6

176 18.9 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.444 ng/uL

RT: 13.127 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:216 Resp: 323626

Ion Ratio Lower Upper

216 100

214 77.0 62.9 94.3

218 47.8 38.6 57.8



#76

Pentachlorobenzene

Concen: 21.934 ng/uL

RT: 14.651 min Scan# 1966

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:250 Resp: 372081


Ion Ratio Lower Upper

250 100

248 61.7 50.1 75.1

252 61.4 50.6 75.8





#77

Carbazole

Concen: 20.628 ng/ul

RT: 17.509 min Scan# 2452

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:167 Resp: 1460910

Ion Ratio Lower Upper

167 100

166 21.7 17.2 25.8

139 13.5 10.4 15.6



#78

Di-n-butylphthalate

Concen: 19.812 ng/ul

RT: 18.080 min Scan# 2549

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:149 Resp: 1887300


Ion Ratio Lower Upper

149 100

150 8.9 7.3 10.9

104 5.0 3.8 5.6





#80

Fluoranthene

Concen: 19.892 ng/ul

RT: 19.162 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:202 Resp: 1734592

Ion Ratio Lower Upper

202 100

101 12.4 10.7 16.1

100 9.3 8.2 12.2



#82

Pyrene

Concen: 20.190 ng/ul

RT: 19.527 min Scan# 2795

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:202 Resp: 1832965


Ion Ratio Lower Upper

202 100

101 13.7 12.5 18.7

100 11.6 10.0 15.0









#83
Butylbenzylphthalate
Concen: 18.401 ng/ul
RT: 20.445 min Scan# 2951
Delta R.T. -0.006 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:149 Resp: 821422
Ion Ratio Lower Upper
149 100
91 75.4 55.9 83.9
206 18.8 15.4 23.0



#84
3,3'-Dichlorobenzidine
Concen: 19.891 ng/ul
RT: 21.221 min Scan# 3083
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:252 Resp: 513692
Ion Ratio Lower Upper
252 100
254 64.9 51.6 77.4
126 12.1 11.1 16.7









#85
Benzo(a)anthracene
Concen: 20.283 ng/ul
RT: 21.280 min Scan# 3093
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12




Instrument :
BNA_M
ClientSampleId :
SSTD020077


Tgt Ion:228 Resp: 1818417
Ion Ratio Lower Upper
228 100
229 19.7 15.8 23.8
226 27.0 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.634 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. -0.000 min
Lab File: BM044673.D
Acq: 19 Mar 2024 21:12

Tgt Ion:149 Resp: 1277906
Ion Ratio Lower Upper
149 100
167 27.1 22.1 33.1
279 3.7 2.6 4.0





#87

Chrysene

Concen: 20.170 ng/ul

RT: 21.333 min Scan# 3102

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:228 Resp: 1674217

Ion Ratio Lower Upper

228 100

226 30.1 23.9 35.9

229 19.6 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 17.686 ng/ul




RT: 22.115 min Scan# 3235


Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Tgt Ion:149 Resp: 2216023





#90

Benzo(b)fluoranthene

Concen: 19.132 ng/ul

RT: 22.868 min Scan# 3363

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:252 Resp: 1792325

Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 11.1 9.3 13.9



#91

Benzo(k)fluoranthene

Concen: 20.145 ng/ul

RT: 22.909 min Scan# 3370

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:252 Resp: 1848352


Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 11.9 10.0 15.0





#93

Benzo(a)pyrene

Concen: 19.838 ng/ul

RT: 23.444 min Scan# 3461

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:252 Resp: 1738466

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 13.3 10.9 16.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.358 ng/ul

RT: 25.815 min Scan# 3864

Delta R.T. 0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12




Tgt Ion:276 Resp: 2197622


Ion Ratio Lower Upper

276 100

138 29.0 23.1 34.7

277 25.0 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 22.581 ng/ul

RT: 25.826 min Scan# 3866

Delta R.T. -0.000 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Instrument :

BNA_M

ClientSampleId :

SSTD020077





Tgt Ion:278 Resp: 1850694

Ion Ratio Lower Upper

278 100

139 18.5 15.8 23.6

279 23.8 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 22.374 ng/ul

RT: 26.497 min Scan# 3980

Delta R.T. -0.006 min

Lab File: BM044673.D

Acq: 19 Mar 2024 21:12

Tgt Ion:276 Resp: 1754964

Ion Ratio Lower Upper

276 100

138 24.3 20.8 31.2

277 24.2 19.0 28.4

