

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032015\
 Data File : BM000738.D
 Acq On : 20 Mar 2015 12:08
 Operator : TP/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.403

Quant Time: Mar 21 04:07:37 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:39:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	0.952	1.087	-14.2	102	0.00
4 SURR	2-Methylnaphthalene-d10	0.460	0.507	-10.2	99	0.00
5	2-Methylnaphthalene	0.608	0.682	-12.2	101	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
7	Acenaphthylene	1.389	1.527	-9.9	100	0.00
8 C	Acenaphthene	1.122	1.260	-12.3	100	0.00
9	Fluorene	1.263	1.401	-10.9	101	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
11	Pentachlorophenol	0.050	0.043	14.0	93	0.00
12	Phenanthrene	1.069	1.215	-13.7	104	0.00
13	Anthracene	0.939	1.045	-11.3	101	0.00
14 SURR	Fluoranthene-d10	0.793	0.898	-13.2	103	-0.01
15 C	Fluoranthene	1.184	1.340	-13.2	104	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
17	Pyrene	1.429	1.570	-9.9	103	0.00
18	Benzo(a)anthracene	1.112	1.206	-8.5	103	0.01
19	Chrysene	1.313	1.513	-15.2	107	0.01
20 I	Perylene-d12	1.000	1.000	0.0	104	0.00
21	Benzo(b)fluoranthene	1.361	1.515	-11.3	103	0.00
22	Benzo(k)fluoranthene	1.404	1.697	-20.9#	107	0.00
23 C	Benzo(a)pyrene	1.244	1.441	-15.8	105	0.00
24	Indeno(1,2,3-cd)pyrene	1.440	1.656	-15.0	104	0.00
25	Dibenzo(a,h)anthracene	1.164	1.332	-14.4	103	0.00
26	Benzo(g,h,i)perylene	1.281	1.475	-15.1	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0