

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032015\
 Data File : BM000743.D
 Acq On : 20 Mar 2015 16:41
 Operator : TP/IZ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.404

Quant Time: Mar 21 04:14:08 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM031915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 21 04:02:13 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
3	Naphthalene	0.952	1.104	-16.0	117	0.00
4 SURR	2-Methylnaphthalene-d10	0.460	0.514	-11.7	114	0.00
5	2-Methylnaphthalene	0.608	0.698	-14.8	117	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
7	Acenaphthylene	1.389	1.639	-18.0	119	0.00
8 C	Acenaphthene	1.122	1.310	-16.8	116	0.00
9	Fluorene	1.263	1.456	-15.3	117	-0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
11	Pentachlorophenol	0.050	0.042	16.0	106	0.00
12	Phenanthrene	1.069	1.220	-14.1	121	0.00
13	Anthracene	0.939	1.057	-12.6	119	0.00
14 SURR	Fluoranthene-d10	0.793	0.881	-11.1	117	0.00
15 C	Fluoranthene	1.184	1.318	-11.3	118	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	119	0.01
17	Pyrene	1.429	1.602	-12.1	118	0.00
18	Benzo(a)anthracene	1.112	1.268	-14.0	121	0.01
19	Chrysene	1.313	1.522	-15.9	120	0.01
20 I	Perylene-d12	1.000	1.000	0.0	117	0.00
21	Benzo(b)fluoranthene	1.361	1.505	-10.6	115	0.00
22	Benzo(k)fluoranthene	1.404	1.687	-20.2#	119	0.00
23 C	Benzo(a)pyrene	1.244	1.442	-15.9	118	0.00
24	Indeno(1,2,3-cd)pyrene	1.440	1.623	-12.7	114	0.00
25	Dibenzo(a,h)anthracene	1.164	1.304	-12.0	113	0.00
26	Benzo(g,h,i)perylene	1.281	1.439	-12.3	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0