

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032016\
 Data File : BM004705.D
 Acq On : 20 Mar 2016 14:28
 Operator : UM/SJ
 Sample : MDL-05-W
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-05-W

Quant Time: Mar 21 02:35:16 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	27833	5.00	ng	0.00
6) Naphthalene-d8	10.61	136	107437	5.00	ng	-0.02
10) Acenaphthene-d10	14.46	164	55531	5.00	ng	0.00
16) Phenanthrene-d10	17.19	188	108402	5.00	ng	-0.03
22) Chrysene-d12	21.39	240	119930	5.00	ng	0.00
28) Perylene-d12	23.69	264	96895	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.42	112	40411	6.29	ng	0.00
5) Phenol-d6	6.99	99	47499	6.04	ng	0.00
7) Nitrobenzene-d5	8.99	82	16005	3.09	ng	0.00
11) 2,4,6-Tribromophenol	15.95	330	9387	5.10	ng	0.00
12) 2-Fluorobiphenyl	13.09	172	69007	4.34	ng	0.00
24) Terphenyl-d14	19.84	244	59037	3.89	ng	0.00


Target Compounds						Qvalue
2) 1,4-Dioxane	3.37	88	255	0.02	ng	# 77
3) n-Nitrosodimethylamine	3.68	42	96	0.03	ng	# 93
8) Naphthalene	10.66	128	1431	0.05	ng	# 89
9) 2-Methylnaphthalene	12.29	142	853	0.05	ng	# 90
13) Acenaphthylene	14.17	152	1004	0.04	ng	99
14) Acenaphthene	14.53	154	1017	0.05	ng	87
15) Fluorene	15.53	166	908	0.04	ng	# 100
17) Hexachlorobenzene	16.53	284	293	0.04	ng	# 100
18) Pentachlorophenol	16.86	266	121	0.16	ng	# 75
19) Phenanthrene	17.24	178	2090	0.05	ng	# 95
20) Anthracene	17.34	178	1135	0.04	ng	99
21) Fluoranthene	19.26	202	1674	0.04	ng	98
23) Pyrene	19.64	202	1796	0.04	ng	97
25) Benzo(a)anthracene	21.37	228	3833	0.08	ng	# 96
26) Chrysene	21.37	228	3865	0.11	ng	96
27) Indeno(1,2,3-cd)pyrene	26.02	276	1520	0.04	ng	98
29) Benzo(b)fluoranthene	22.99	252	1664	0.05	ng	# 71
30) Benzo(k)fluoranthene	23.04	252	1619	0.05	ng	# 69
31) Benzo(a)pyrene	23.59	252	1361	0.04	ng	# 58
32) Dibenzo(a,h)anthracene	26.04	278	1155	0.04	ng	# 68
33) Benzo(g,h,i)perylene	26.73	276	1421	0.05	ng	# 84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032016\
Data File : BM004705.D
Acq On : 20 Mar 2016 14:28
Operator : UM/SJ
Sample : MDL-05-W
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-05-W

Quant Time: Mar 21 02:35:16 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 19 07:50:46 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.84 min Scan# 323

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampled :

MDL-05-W




Tgt Ion: 152 Resp: 27833

Ion Ratio Lower Upper

152 100

150 140.4 115.9 173.9

115 44.7 38.4 57.6



#2

1,4-Dioxane

Concen: 0.02 ng

RT: 3.37 min Scan# 45

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28




Tgt Ion: 88 Resp: 255


Ion Ratio Lower Upper

88 100

58 36.1 43.9 65.9#

43 17.3 19.9 29.9#





#3

n-Nitrosodimethylamine

Concen: 0.03 ng

RT: 3.68 min Scan# 65

Delta R.T. 0.02 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion: 42 Resp: 96

Ion Ratio Lower Upper

42 100

74 143.8 120.9 181.3




44 0.0 6.0 9.0#



Abundance Ion 42.00 (41.70 to 42.70): BM004681.D (-59) (-)

Ion 74.00 (73.70 to 74.70): BM004681.D (-59) (-)

Ion 44.00 (43.70 to 44.70): BM004681.D (-59) (-)



#4

2-Fluorophenol

Concen: 6.29 ng

RT: 5.42 min Scan# 178

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28


Tgt Ion: 112 Resp: 40411

Ion Ratio Lower Upper

112 100

64 46.5 37.0 55.6



63 23.8 19.3 28.9




Abundance Ion 112.00 (111.70 to 112.70): BM004681.D (-172) (-)

Ion 64.00 (63.70 to 64.70): BM004681.D (-172) (-)

Ion 63.00 (62.70 to 63.70): BM004681.D (-172) (-)





#5

Phenol-d6

Concen: 6.04 ng

RT: 6.99 min Scan# 273

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion: 99 Resp: 47499

Ion Ratio Lower Upper

99 100

42 14.2 12.7 19.1




71 27.9 22.9 34.3



Abundance Ion 99.00 (98.70 to 99.70): BM004705.D (-549) (-)

Ion 42.00 (41.70 to 42.70): BM004705.D (-549) (-)

Ion 71.00 (70.70 to 71.70): BM004705.D (-549) (-)



#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.61 min Scan# 560

Delta R.T. -0.02 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Tgt Ion: 136 Resp: 107437


Ion Ratio Lower Upper

136 100

137 10.2 9.4 14.2

54 6.9 2.2 3.4#

68 5.9 2.2 3.4#






Abundance Ion 136.00 (135.70 to 136.70): BM004705.D (-549) (-)

Ion 137.00 (136.70 to 137.70): BM004705.D (-549) (-)

Ion 54.00 (53.70 to 54.70): BM004705.D (-549) (-)

Ion 68.00 (67.70 to 68.70): BM004705.D (-549) (-)





#7

Nitrobenzene-d5

Concen: 3.09 ng

RT: 8.99 min Scan# 427

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion: 82 Resp: 16005

Ion Ratio Lower Upper

82 100

128 51.6 41.8 62.8




54 35.0 32.9 49.3



Abundance Ion 82.00 (81.70 to 82.70): BM004705.D (-408) (-)

Ion 128.00 (127.70 to 128.70): BM004705.D (-408) (-)

Ion 54.00 (53.70 to 54.70): BM004705.D (-408) (-)



#8

Naphthalene

Concen: 0.05 ng

RT: 10.66 min Scan# 563

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28


Tgt Ion: 128 Resp: 1431

Ion Ratio Lower Upper

128 100

129 16.1 8.9 13.3#



127 18.1 11.3 16.9#




Abundance Ion 128.00 (127.70 to 128.70): BM004705.D (-551) (-)

Ion 129.00 (128.70 to 129.70): BM004705.D (-551) (-)

Ion 127.00 (126.70 to 127.70): BM004705.D (-551) (-)





#9

2-Methylnaphthalene

Concen: 0.05 ng

RT: 12.29 min Scan# 706

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W





Tgt Ion:142 Resp: 853

Ion Ratio Lower Upper

142 100

141 88.5 66.2 99.2

115 41.2 23.8 35.8#



#10

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.46 min Scan# 866

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28




Tgt Ion:164 Resp: 55531


Ion Ratio Lower Upper

164 100

162 100.5 85.4 128.0

160 45.2 40.2 60.2









#11
 2,4,6-Tribromophenol
 Concen: 5.10 ng
 RT: 15.95 min Scan# 933
 Delta R.T. 0.00 min
 Lab File: BM004705.D
 Acq: 20 Mar 2016 14:28

Instrument :
 BNA_M
 ClientSampleId :
 MDL-05-W

Tgt Ion:330 Resp: 9387
 Ion Ratio Lower Upper
 330 100
 332 94.2 93.9 140.9
 141 0.0 42.2 63.2#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#12
 2-Fluorobiphenyl
 Concen: 4.34 ng
 RT: 13.09 min Scan# 783
 Delta R.T. 0.00 min
 Lab File: BM004705.D
 Acq: 20 Mar 2016 14:28

Tgt Ion:172 Resp: 69007
 Ion Ratio Lower Upper
 172 100
 171 32.7 28.2 42.2
 170 20.9 19.4 29.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#13

Acenaphthylene

Concen: 0.04 ng

RT: 14.17 min Scan# 853

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion:152 Resp: 1004

Ion Ratio Lower Upper

152 100

151 18.8 15.8 23.6




153 12.5 10.2 15.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#14

Acenaphthene

Concen: 0.05 ng

RT: 14.53 min Scan# 869

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28


Tgt Ion:154 Resp: 1017

Ion Ratio Lower Upper

154 100

153 89.6 83.8 125.6



152 44.5 40.3 60.5




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#15
Fluorene
 Concen: 0.04 ng
 RT: 15.53 min Scan# 914
 Delta R.T. 0.00 min
 Lab File: BM004705.D
 Acq: 20 Mar 2016 14:28




Instrument :
 BNA_M
 ClientSampleId :
 MDL-05-W


Tgt Ion:166 Resp: 908
 Ion Ratio Lower Upper
 166 100
 165 0.0 0.0 0.0
 167 13.2 10.6 15.8



#16
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.19 min Scan# 982
 Delta R.T. -0.03 min
 Lab File: BM004705.D
 Acq: 20 Mar 2016 14:28

Tgt Ion:188 Resp: 108402
 Ion Ratio Lower Upper
 188 100
 94 17.2 1.4 2.0#
 80 17.6 1.0 1.6#









#17
Hexachlorobenzene
Concen: 0.04 ng
RT: 16.53 min Scan# 956
Delta R.T. 0.00 min
Lab File: BM004705.D
Acq: 20 Mar 2016 14:28

Instrument :
BNA_M
ClientSampleId :
MDL-05-W

Tgt Ion: 284 Resp: 293
Ion Ratio Lower Upper
284 100
142 0.0 0.0 0.0
249 0.0 0.0 0.0




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#18
Pentachlorophenol
Concen: 0.16 ng
RT: 16.86 min Scan# 969
Delta R.T. 0.00 min
Lab File: BM004705.D
Acq: 20 Mar 2016 14:28

Tgt Ion: 266 Resp: 121
Ion Ratio Lower Upper
266 100
268 82.4 49.1 73.7#
264 84.7 52.9 79.3#



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E









#19
Phenanthrene
Concen: 0.05 ng
RT: 17.24 min Scan# 984
Delta R.T. 0.00 min
Lab File: BM004705.D
Acq: 20 Mar 2016 14:28




Instrument :
BNA_M
ClientSampleId :
MDL-05-W


Tgt Ion:178 Resp: 2090
Ion Ratio Lower Upper
178 100
176 18.6 15.5 23.3
179 19.1 12.2 18.4#



#20
Anthracene
Concen: 0.04 ng
RT: 17.34 min Scan# 988
Delta R.T. 0.00 min
Lab File: BM004705.D
Acq: 20 Mar 2016 14:28

Tgt Ion:178 Resp: 1135
Ion Ratio Lower Upper
178 100
176 18.8 14.5 21.7
179 15.8 12.6 18.8





#21

Fluoranthene

Concen: 0.04 ng

RT: 19.26 min Scan# 1087

Delta R.T. -0.02 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion:202 Resp: 1674

Ion Ratio Lower Upper

202 100

101 10.6 8.7 13.1




203 18.8 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#22

Chrysene-d12

Concen: 5.00 ng

RT: 21.39 min Scan# 1203

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28


Tgt Ion:240 Resp: 119930

Ion Ratio Lower Upper

240 100

120 5.1 7.4 11.2#



236 27.4 24.5 36.7




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#23

Pyrene

Concen: 0.04 ng

RT: 19.64 min Scan# 1109

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion:202 Resp: 1796

Ion Ratio Lower Upper

202 100

200 21.2 16.8 25.2




203 20.3 13.9 20.9



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Terphenyl-d14

Concen: 3.89 ng

RT: 19.84 min Scan# 1121

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28


Tgt Ion:244 Resp: 59037

Ion Ratio Lower Upper

244 100

212 8.5 7.7 11.5



122 1.9 3.4 5.0#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#25
Benzo(a)anthracene
Concen: 0.08 ng
RT: 21.37 min Scan# 1202
Delta R.T. 0.00 min
Lab File: BM004705.D
Acq: 20 Mar 2016 14:28




Instrument :
BNA_M
ClientSampleId :
MDL-05-W


Tgt Ion:228 Resp: 3833
Ion Ratio Lower Upper
228 100
226 30.1 24.2 36.2
229 21.8 14.2 21.2#



#26
Chrysene
Concen: 0.11 ng
RT: 21.37 min Scan# 1202
Delta R.T. -0.06 min
Lab File: BM004705.D
Acq: 20 Mar 2016 14:28

Tgt Ion:228 Resp: 3865
Ion Ratio Lower Upper
228 100
226 30.1 20.9 31.3
229 21.8 17.4 26.2









#27
 Indeno(1,2,3-cd)pyrene
 Concen: 0.04 ng
 RT: 26.02 min Scan# 1584
 Delta R.T. 0.00 min
 Lab File: BM004705.D
 Acq: 20 Mar 2016 14:28




Instrument :
 BNA_M
 ClientSampleId :
 MDL-05-W


Tgt Ion: 276 Resp: 1520
 Ion Ratio Lower Upper
 276 100
 138 33.7 25.9 38.9
 277 25.3 19.8 29.6



#28
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.69 min Scan# 1361
 Delta R.T. 0.00 min
 Lab File: BM004705.D
 Acq: 20 Mar 2016 14:28

Tgt Ion: 264 Resp: 96895
 Ion Ratio Lower Upper
 264 100
 260 25.0 21.0 31.4
 265 22.8 17.5 26.3





#29

Benzo(b)fluoranthene

Concen: 0.05 ng

RT: 22.99 min Scan# 1294

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W





Tgt Ion:252 Resp: 1664

Ion Ratio Lower Upper

252 100

253 33.9 17.9 26.9#

125 29.4 11.4 17.0#



#30

Benzo(k)fluoranthene

Concen: 0.05 ng

RT: 23.04 min Scan# 1298

Delta R.T. -0.01 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28




Tgt Ion:252 Resp: 1619


Ion Ratio Lower Upper

252 100

253 35.4 19.3 28.9#

125 32.1 10.6 15.8#





#31

Benzo(a)pyrene

Concen: 0.04 ng

RT: 23.59 min Scan# 1351

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W


Tgt Ion: 252 Resp: 1361

Ion Ratio Lower Upper

252 100

253 40.6 18.6 28.0#




125 37.3 12.3 18.5#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#32

Dibenzo(a,h)anthracene

Concen: 0.04 ng

RT: 26.04 min Scan# 1586

Delta R.T. 0.00 min

Lab File: BM004705.D

Acq: 20 Mar 2016 14:28


Tgt Ion: 278 Resp: 1155

Ion Ratio Lower Upper

278 100

139 35.7 17.6 26.4#



279 41.9 19.9 29.9#




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#33

Benzo(g,h,i)perylene

Concen: 0.05 ng

RT: 26.73 min Scan# 1652

Delta R.T. 0.00 min

Lab File: BM004705.D


Acq: 20 Mar 2016 14:28

Instrument :

BNA_M

ClientSampleId :

MDL-05-W



Tgt Ion: 276 Resp: 1421

Ion Ratio Lower Upper

276 100

277 32.7 19.1 28.7#

138 38.3 24.5 36.7#

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

