

Data Path : Z:\HPCHEM1\BNA_M\Data\BM032016\
 Data File : BM004700.D
 Acq On : 20 Mar 2016 11:30
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 02:10:48 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	0.00
2	1,4-Dioxane	0.400	0.460	-15.0	106	0.00
3	n-Nitrosodimethylamine	0.400	0.416	-4.0	98	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	93	0.00
5 S	Phenol-d6	0.400	0.369	7.8	89	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	108	-0.02
7 S	Nitrobenzene-d5	0.400	0.358	10.5	85	0.00
8	Naphthalene	0.400	0.452	-13.0	105	0.00
9	2-Methylnaphthalene	0.400	0.444	-11.0	102	-0.01
10 I	Acenaphthene-d10	5.000	5.000	0.0	103	0.00
11 S	2,4,6-Tribromophenol	0.400	0.327	18.3	75	0.00
12 S	2-Fluorobiphenyl	0.400	0.477	-19.2	104	0.00
13	Acenaphthylene	0.400	0.403	-0.8	93	0.00
14	Acenaphthene	0.400	0.438	-9.5	97	0.00
15	Fluorene	0.400	0.453	-13.2	101	-0.02
16 I	Phenanthrene-d10	5.000	5.000	0.0	105	-0.03
17	Hexachlorobenzene	0.400	0.405	-1.3	89	0.00
18	Pentachlorophenol	0.400	0.357	10.8	66	0.00
19	Phenanthrene	0.400	0.459	-14.8	102	0.00
20	Anthracene	0.400	0.395	1.3	90	0.00
21	Fluoranthene	0.400	0.418	-4.5	94	-0.02
22 I	Chrysene-d12	5.000	5.000	0.0	109	0.00
23	Pyrene	0.400	0.386	3.5	92	-0.02
24 S	Terphenyl-d14	0.400	0.375	6.3	90	0.00
25	Benzo(a)anthracene	0.400	0.360	10.0	84	0.00
26	Chrysene	0.400	0.464	-16.0	115	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.403	-0.8	95	0.00
28 I	Perylene-d12	5.000	5.000	0.0	102	0.00
29	Benzo(b)fluoranthene	0.400	0.459	-14.8	104	0.00
30	Benzo(k)fluoranthene	0.400	0.420	-5.0	93	0.00
31 C	Benzo(a)pyrene	0.400	0.409	-2.2	92	0.00
32	Dibenzo(a,h)anthracene	0.400	0.420	-5.0	95	0.00
33	Benzo(g,h,i)perylene	0.400	0.436	-9.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0