

Data Path : Z:\HPCHEM1\BNA_M\Data\BM032016\
 Data File : BM004700.D
 Acq On : 20 Mar 2016 11:30
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 02:10:48 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.494	0.580	-17.4	106	0.00
3	n-Nitrosodimethylamine	0.517	0.537	-3.9	98	0.00
4 S	2-Fluorophenol	1.154	1.145	0.8	93	0.00
5 S	Phenol-d6	1.413	1.305	7.6	89	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	108	-0.02
7 S	Nitrobenzene-d5	0.241	0.216	10.4	85	0.00
8	Naphthalene	1.375	1.555	-13.1	105	0.00
9	2-Methylnaphthalene	0.861	0.957	-11.1	102	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
11 S	2,4,6-Tribromophenol	0.166	0.136	18.1	75	0.00
12 S	2-Fluorobiphenyl	1.431	1.705	-19.1	104	0.00
13	Acenaphthylene	2.572	2.590	-0.7	93	0.00
14	Acenaphthene	1.702	1.865	-9.6	97	0.00
15	Fluorene	1.908	2.159	-13.2	101	-0.02
16 I	Phenanthrene-d10	1.000	1.000	0.0	105	-0.03
17	Hexachlorobenzene	0.308	0.312	-1.3	89	0.00
18	Pentachlorophenol	0.169	0.114	32.5#	66	0.00
19	Phenanthrene	2.004	2.301	-14.8	102	0.00
20	Anthracene	1.429	1.413	1.1	90	0.00
21	Fluoranthene	1.895	1.978	-4.4	94	-0.02
22 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
23	Pyrene	1.835	1.771	3.5	92	-0.02
24 S	Terphenyl-d14	0.632	0.592	6.3	90	0.00
25	Benzo(a)anthracene	2.000	1.798	10.1	84	0.00
26	Chrysene	1.487	1.726	-16.1	115	0.00
27	Indeno(1,2,3-cd)pyrene	1.674	1.687	-0.8	95	0.00
28 I	Perylene-d12	1.000	1.000	0.0	102	0.00
29	Benzo(b)fluoranthene	1.819	2.088	-14.8	104	0.00
30	Benzo(k)fluoranthene	1.802	1.890	-4.9	93	0.00
31 C	Benzo(a)pyrene	1.706	1.743	-2.2	92	0.00
32	Dibenzo(a,h)anthracene	1.527	1.603	-5.0	95	0.00
33	Benzo(g,h,i)perylene	1.613	1.758	-9.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0