

Data Path : Z:\HPCHEM1\BNA_M\Data\BM032016\
 Data File : BM004709.D
 Acq On : 20 Mar 2016 16:52
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 04:26:26 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 07:50:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.494	0.575	-16.4	103	0.00
3	n-Nitrosodimethylamine	0.517	0.521	-0.8	92	0.00
4 S	2-Fluorophenol	1.154	1.070	7.3	85	0.00
5 S	Phenol-d6	1.413	1.208	14.5	80	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	106	-0.02
7 S	Nitrobenzene-d5	0.241	0.201	16.6	77	0.00
8	Naphthalene	1.375	1.540	-12.0	102	0.00
9	2-Methylnaphthalene	0.861	0.946	-9.9	99	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
11 S	2,4,6-Tribromophenol	0.166	0.121	27.1#	66	0.00
12 S	2-Fluorobiphenyl	1.431	1.686	-17.8	101	0.00
13	Acenaphthylene	2.572	2.480	3.6	87	0.00
14	Acenaphthene	1.702	1.846	-8.5	94	0.00
15	Fluorene	1.908	2.070	-8.5	95	-0.02
16 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.03
17	Hexachlorobenzene	0.308	0.314	-1.9	88	0.00
18	Pentachlorophenol	0.169	0.095	43.8#	54	0.00
19	Phenanthrene	2.004	2.319	-15.7	100	0.00
20	Anthracene	1.429	1.393	2.5	87	0.00
21	Fluoranthene	1.895	1.948	-2.8	91	-0.02
22 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
23	Pyrene	1.835	1.858	-1.3	92	0.00
24 S	Terphenyl-d14	0.632	0.629	0.5	91	0.00
25	Benzo(a)anthracene	2.000	1.665	16.8	74	0.00
26	Chrysene	1.487	2.031	-36.6#	128	0.00
27	Indeno(1,2,3-cd)pyrene	1.674	1.638	2.2	88	0.00
28 I	Perylene-d12	1.000	1.000	0.0	99	0.00
29	Benzo(b)fluoranthene	1.819	1.919	-5.5	93	0.00
30	Benzo(k)fluoranthene	1.802	1.976	-9.7	94	0.00
31 C	Benzo(a)pyrene	1.706	1.641	3.8	84	0.00
32	Dibenzo(a,h)anthracene	1.527	1.529	-0.1	87	0.00
33	Benzo(g,h,i)perylene	1.613	1.686	-4.5	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0