


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032017\
 Data File : BM009287.D
 Acq On : 20 Mar 2017 10:07
 Operator : SJ/MA
 Sample : DFTPP51
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP51

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031317MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 21 16:14:58 2017



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.0	67482	PASS
68	69	0.00	2	1.9	1346	PASS
69	198	0.00	100	51.4	69403	PASS
70	69	0.00	2	1.0	685	PASS
127	198	10	80	47.4	63984	PASS
197	198	0.00	2	0.6	778	PASS
198	198	100	100	100.0	134917	PASS
199	198	5	9	7.2	9696	PASS
275	198	10	60	25.1	33882	PASS
365	198	1	100	3.7	5007	PASS
441	443	0.01	100	72.9	12788	PASS
442	198	50	100	64.9	87597	PASS
443	442	15	24	20.0	17546	PASS

m/z	Abundance
36.00	422.0
38.00	657.0
39.10	4304.0
40.00	1491.0
41.00	1433.0
42.10	307.0
43.10	1583.0
44.00	970.0
48.90	356.0
50.00	13093.0
51.00	60712.0
52.00	2807.0
54.00	454.0
55.00	1286.0
55.90	1647.0
57.00	4424.0
58.10	421.0
61.10	1191.0
62.00	745.0
63.00	2557.0
65.10	1848.0
67.00	425.0
68.00	1093.0
69.00	60968.0
70.10	470.0
70.90	548.0
73.20	1912.0
74.10	4964.0
75.00	9121.0
76.00	3134.0
77.10	61032.0
78.00	4238.0
79.00	4869.0
79.90	3193.0
81.00	3728.0
82.00	1294.0
83.10	1158.0
84.00	340.0
84.90	1119.0
86.00	1434.0
87.00	759.0
88.10	438.0
91.00	1266.0
91.90	1661.0
93.00	6900.0
94.10	540.0
95.00	730.0
96.20	674.0
97.20	805.0
98.00	6983.0
99.00	5280.0
100.00	458.0
101.00	2103.0
103.00	462.0
104.00	1751.0
105.10	2551.0
106.10	386.0
107.10	16672.0
108.00	2870.0
108.90	957.0
110.00	26312.0
111.00	4167.0
111.90	853.0
115.90	983.0
117.00	17688.0
117.90	1507.0
121.20	333.0
121.90	1617.0
122.90	1558.0
123.90	794.0
125.10	784.0
126.10	412.0
127.00	47216.0
128.00	3978.0
129.00	18120.0
130.00	2050.0
130.90	435.0
132.90	528.0
134.00	456.0
135.00	1706.0
136.10	693.0
137.10	967.0
137.60	570.0
138.10	419.0
139.80	345.0
141.00	3020.0
142.20	1302.0
143.10	927.0
145.70	398.0
146.90	1699.0
148.00	4163.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

148.80	800.0
149.20	843.0
151.00	815.0
152.10	616.0
152.90	1045.0
154.20	453.0
155.00	1519.0
156.10	955.0
157.10	418.0
157.80	467.0
158.90	831.0
159.90	1216.0
161.10	1585.0
162.00	422.0
163.70	312.0
164.90	1620.0
166.10	667.0
167.00	4512.0
168.00	4231.0
168.90	669.0
170.90	327.0
172.90	677.0
174.20	1225.0
175.00	1541.0
176.20	555.0
176.70	535.0
177.20	566.0
178.10	327.0
178.90	4023.0
180.00	3582.0
181.00	1646.0
183.90	729.0
185.00	2809.0
186.00	13630.0
187.00	5290.0
187.80	438.0
188.20	343.0
188.90	1183.0
190.70	464.0
191.20	588.0
192.00	884.0
192.90	2273.0
194.10	765.0
194.90	420.0
196.10	1881.0
198.00	91536.0
199.00	6759.0
199.80	593.0
201.50	999.0
203.10	917.0
203.90	3358.0
205.00	7068.0
206.00	30752.0
207.00	6717.0
208.00	2116.0
209.20	447.0
209.90	619.0
211.00	1621.0
214.90	364.0
216.00	300.0
217.00	7425.0
217.80	899.0
219.30	380.0
221.00	3630.0
222.10	524.0
223.00	1824.0
224.10	15138.0
225.00	3967.0
226.90	6570.0
227.90	1376.0
229.00	1013.0
231.00	450.0
232.80	381.0
233.90	715.0
235.00	1028.0
235.80	321.0
237.00	606.0
238.80	368.0
239.80	650.0
240.80	426.0
241.90	973.0
243.00	782.0
244.00	10674.0
244.90	1354.0
245.90	2081.0
246.90	379.0
249.00	1010.0
253.00	446.0
254.10	455.0
255.00	56440.0
256.00	6661.0
256.80	593.0
258.00	3737.0
259.00	527.0
265.00	2149.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

266.80	582.0
272.90	1572.0
274.00	4340.0
275.00	21632.0
275.90	3523.0
276.90	1791.0
277.70	509.0
281.00	897.0
284.90	636.0
293.00	556.0
296.00	6678.0
297.00	822.0
303.00	1803.0
315.00	539.0
323.00	2496.0
326.90	568.0
333.70	1177.0
334.90	303.0
340.90	459.0
352.10	640.0
354.10	842.0
365.00	3356.0
372.20	926.0
403.10	319.0
421.00	348.0
421.80	445.0
423.10	2478.0
424.00	905.0
441.10	6081.0
442.10	44984.0
443.10	8462.0
444.00	887.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
37.00	400.0
38.20	643.0
39.10	7637.0
40.10	1079.0
41.20	1366.0
42.10	351.0
43.00	1185.0
44.00	1553.0
49.00	818.0
50.00	22336.0
51.10	79040.0
52.00	4836.0
52.90	345.0
55.10	1676.0
55.90	2711.0
57.00	6199.0
57.90	808.0
59.10	595.0
60.90	733.0
62.00	1184.0
63.00	4023.0
64.10	974.0
65.00	2435.0
65.80	467.0
67.10	414.0
68.00	1445.0
69.00	80568.0
70.00	1034.0
71.10	777.0
73.00	2214.0
74.00	9410.0
75.10	14517.0
76.00	4697.0
77.10	98208.0
78.10	7563.0
79.10	7492.0
79.90	5107.0
81.00	7287.0
82.20	1338.0
83.10	2261.0
83.90	655.0
84.90	1530.0
86.10	2161.0
87.10	1132.0
90.90	1877.0
92.10	2609.0
92.90	11619.0
93.90	715.0
95.10	388.0
97.00	807.0
98.00	9420.0
99.00	6713.0
100.00	960.0
101.00	4568.0
102.20	372.0
102.90	899.0
104.00	2645.0
105.00	2776.0
106.00	992.0
107.00	26504.0
107.90	4550.0
109.10	1195.0
110.00	41504.0
111.10	8199.0
111.90	756.0
113.10	862.0
115.00	739.0
116.00	1305.0
117.00	25112.0
118.00	2620.0
119.10	669.0
120.00	337.0
120.90	388.0
122.10	2378.0
123.00	2352.0
124.00	1276.0
125.10	1434.0
126.10	435.0
127.00	76728.0
127.90	5568.0
129.00	31032.0
130.00	2002.0
130.80	579.0
132.00	522.0
132.90	460.0
133.90	837.0
135.00	3704.0
136.10	1414.0
137.10	1372.0
139.20	516.0
141.00	4771.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

141.90	1461.0
142.90	938.0
143.80	447.0
146.10	848.0
147.00	3965.0
148.00	7179.0
149.10	925.0
150.00	372.0
151.20	689.0
152.00	662.0
153.00	1216.0
154.10	1684.0
155.00	3320.0
156.00	3600.0
156.80	772.0
157.20	547.0
157.90	1203.0
159.00	701.0
160.00	1634.0
161.10	1820.0
162.10	410.0
163.20	384.0
164.00	392.0
165.20	1403.0
166.00	1480.0
167.00	9502.0
168.00	5708.0
169.00	1271.0
172.00	1266.0
173.10	1679.0
174.10	1602.0
174.90	3492.0
176.00	579.0
176.90	1228.0
177.90	800.0
178.90	7369.0
180.00	5262.0
181.10	2158.0
182.10	639.0
183.80	484.0
184.20	472.0
185.10	3641.0
186.10	21208.0
187.00	7117.0
187.90	958.0
188.90	2495.0
189.90	306.0
191.10	879.0
191.90	1904.0
193.00	4070.0
194.20	901.0
195.90	4334.0
197.10	985.0
198.00	165568.0
199.00	11831.0
199.90	917.0
201.00	541.0
201.60	969.0
202.10	465.0
202.90	1951.0
204.00	6385.0
205.00	10076.0
206.00	43840.0
207.00	7968.0
208.10	2054.0
209.00	1347.0
210.20	1436.0
210.90	1578.0
213.10	482.0
214.90	476.0
215.90	1117.0
216.90	14317.0
217.90	1906.0
220.20	424.0
221.00	6011.0
222.00	1691.0
223.00	2320.0
224.00	26928.0
225.10	6958.0
225.80	564.0
227.00	11990.0
228.00	2452.0
229.00	3001.0
230.90	845.0
233.20	337.0
233.90	393.0
234.90	924.0
235.80	548.0
236.90	1055.0
240.20	411.0
241.00	906.0
242.00	1415.0
243.10	1939.0
244.10	22448.0
244.90	2011.0

Instrument :

BNA_M

ClientSampleId :

DFTPP51

245.90	3477.0
247.00	1050.0
249.00	870.0
249.90	325.0
253.10	651.0
254.00	1427.0
255.00	105088.0
256.00	14761.0
256.90	650.0
258.00	6172.0
259.10	670.0
265.00	2837.0
266.20	625.0
270.90	582.0
272.00	402.0
273.00	4049.0
274.00	7947.0
275.00	37272.0
276.00	4729.0
276.90	4829.0
277.80	529.0
278.10	572.0
281.00	913.0
283.00	1049.0
283.80	416.0
284.20	431.0
285.00	761.0
289.00	788.0
292.80	1201.0
296.00	13582.0
296.90	1886.0
302.00	848.0
303.00	1423.0
304.20	372.0
312.70	308.0
314.10	498.0
315.00	1223.0
315.90	588.0
323.00	3060.0
323.90	1171.0
326.90	634.0
328.00	412.0
334.10	2731.0
334.90	741.0
341.00	597.0
346.00	479.0
352.00	1088.0
353.10	1017.0
354.10	714.0
355.10	840.0
364.90	5114.0
365.90	618.0
370.80	325.0
372.10	2133.0
373.00	462.0
382.90	830.0
390.10	313.0
391.00	335.0
402.10	676.0
403.00	1212.0
403.70	444.0
405.10	324.0
421.00	768.0
421.70	447.0
422.10	567.0
423.10	4456.0
424.00	1623.0
441.10	15539.0
442.10	98488.0
443.10	21640.0
444.00	1376.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	1125.0
39.10	5353.0
40.10	1194.0
41.00	1283.0
42.20	347.0
43.00	1752.0
43.90	979.0
44.90	411.0
49.00	363.0
50.00	20488.0
51.10	63832.0
52.00	3188.0
52.90	368.0
55.00	1124.0
56.00	2276.0
57.00	5258.0
57.90	582.0
59.10	681.0
60.90	962.0
61.90	929.0
63.00	3337.0
64.00	578.0
64.90	867.0
66.10	393.0
66.90	475.0
68.00	1502.0
69.00	67880.0
70.20	551.0
71.20	845.0
73.00	3172.0
74.00	8107.0
75.00	11134.0
75.90	3875.0
77.10	79504.0
78.00	5785.0
79.00	4985.0
80.00	4290.0
81.00	5949.0
82.00	1274.0
83.20	1944.0
84.10	496.0
85.00	2067.0
86.00	1742.0
86.90	1043.0
91.00	1779.0
92.00	2389.0
93.00	11159.0
94.00	734.0
94.90	564.0
96.10	547.0
96.90	557.0
97.30	475.0
98.00	7177.0
99.00	6485.0
99.80	324.0
101.00	2591.0
102.80	1050.0
104.00	1879.0
105.00	2858.0
105.80	369.0
107.00	22272.0
108.00	3360.0
108.90	1299.0
110.00	36600.0
111.00	6327.0
111.90	435.0
112.90	556.0
115.10	743.0
116.10	1131.0
117.00	28688.0
117.90	1876.0
119.10	615.0
120.00	518.0
121.10	509.0
122.20	1330.0
122.90	1794.0
124.00	1512.0
124.80	1044.0
126.10	398.0
127.00	68008.0
128.00	5081.0
129.00	27696.0
129.90	2101.0
130.90	506.0
132.00	403.0
132.90	532.0
134.00	1906.0
134.90	2857.0
135.80	1095.0
137.10	1935.0
137.70	346.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

139.10	469.0
140.90	3939.0
142.00	1676.0
142.90	993.0
144.00	504.0
146.00	1558.0
146.90	2787.0
148.00	5556.0
149.10	437.0
150.90	338.0
151.90	407.0
153.00	1959.0
153.90	1657.0
155.10	1571.0
156.00	3945.0
156.70	721.0
157.20	851.0
158.00	1469.0
159.00	580.0
160.00	1669.0
161.00	3026.0
161.80	622.0
162.20	423.0
163.00	497.0
164.10	404.0
165.10	1864.0
165.80	729.0
166.90	8838.0
168.10	5263.0
169.00	686.0
170.90	665.0
171.90	647.0
173.10	525.0
174.00	2157.0
175.00	2744.0
176.20	1203.0
177.10	2457.0
177.90	585.0
178.90	6702.0
180.00	3688.0
181.00	2092.0
184.00	1027.0
185.00	5043.0
186.00	24144.0
187.00	6409.0
187.80	585.0
189.00	1476.0
190.90	861.0
192.00	2486.0
193.10	2324.0
194.10	1170.0
195.00	690.0
195.80	2738.0
197.00	1351.0
198.00	147648.0
199.00	10499.0
200.10	1222.0
200.80	358.0
201.40	402.0
202.10	504.0
202.90	1990.0
204.10	7250.0
205.10	9787.0
206.00	41544.0
207.00	7787.0
208.00	2420.0
209.00	1467.0
210.20	456.0
211.10	1496.0
211.80	510.0
215.10	343.0
215.90	1009.0
217.00	11383.0
218.00	1419.0
220.30	643.0
221.00	6481.0
221.80	1487.0
223.00	3347.0
224.00	20016.0
225.10	5922.0
225.90	320.0
227.00	11540.0
227.90	1566.0
229.00	2444.0
230.10	562.0
231.10	979.0
233.80	515.0
234.10	519.0
234.80	529.0
235.80	537.0
236.90	1491.0
239.00	442.0
239.80	401.0
241.00	348.0
242.00	2149.0

Instrument :

BNA_M

ClientSampleId :

DFTPP51

242.90	2569.0
244.00	20072.0
245.00	3697.0
246.00	4929.0
246.90	480.0
248.00	456.0
249.00	715.0
250.70	426.0
252.10	458.0
253.00	432.0
254.10	874.0
255.00	102736.0
256.00	14377.0
256.90	1269.0
258.00	8287.0
259.00	1528.0
264.00	312.0
264.90	4081.0
266.20	330.0
267.00	946.0
272.90	3308.0
274.00	7276.0
275.00	42744.0
276.00	6056.0
276.90	3879.0
278.00	647.0
281.00	1194.0
283.00	366.0
284.10	575.0
285.00	605.0
288.70	324.0
292.90	646.0
296.00	12106.0
297.00	1889.0
302.90	1933.0
303.80	399.0
307.90	371.0
313.80	533.0
315.00	1916.0
316.00	1054.0
321.10	529.0
322.20	345.0
323.10	3834.0
323.80	429.0
324.20	590.0
327.00	904.0
327.90	486.0
332.90	472.0
334.00	3944.0
334.90	801.0
340.90	537.0
346.00	1417.0
352.00	921.0
352.90	803.0
354.00	822.0
359.00	320.0
365.00	6551.0
366.00	912.0
372.20	1988.0
373.10	456.0
390.00	393.0
402.00	1001.0
403.00	1244.0
415.10	414.0
421.10	823.0
422.00	893.0
423.10	6089.0
424.00	1563.0
441.10	16744.0
442.10	119320.0
443.10	22536.0
444.10	2932.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
39.10	955.0
40.00	587.0
41.00	1611.0
43.20	824.0
44.10	635.0
51.10	379.0
54.00	407.0
54.90	1306.0
57.10	845.0
67.10	842.0
69.20	402.0
71.20	1106.0
73.10	1739.0
74.00	340.0
75.90	484.0
77.00	393.0
81.20	493.0
82.20	314.0
83.20	328.0
84.00	404.0
85.10	829.0
91.10	428.0
94.90	537.0
95.90	366.0
96.90	308.0
103.00	499.0
105.00	417.0
109.00	711.0
111.20	319.0
112.80	477.0
114.90	446.0
126.20	422.0
133.00	606.0
134.90	388.0
147.00	650.0
167.00	402.0
185.00	577.0
186.10	410.0
191.00	564.0
196.10	309.0
207.00	2447.0
207.90	899.0
208.90	792.0
265.00	304.0
281.00	1080.0

Instrument :

BNA_M

ClientSampleId :

DFTPP51