

Data Path : Z:\HPCHEM1\BNA_M\Data\BM032216\
 Data File : BM004740.D
 Acq On : 22 Mar 2016 23:32
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.464

Quant Time: Mar 23 06:07:46 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM032216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 23 05:10:55 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2 SURR	1,4-Dioxane-d8	2.424	2.466	-1.7	86	0.00
3	1,4-Dioxane	2.954	2.885	2.3	85	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
5	Naphthalene	11.781	11.503	2.4	84	0.00
6 SURR	2-Methylnaphthalene-d10	5.695	5.584	1.9	84	0.00
7	2-Methylnaphthalene	7.494	7.314	2.4	84	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
9	Acenaphthylene	18.396	18.297	0.5	89	0.00
10 C	Acenaphthene	16.502	16.137	2.2	86	0.00
11	Fluorene	17.439	17.307	0.8	88	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
13	Pentachlorophenol	0.670	0.688	-2.7	111	0.00
14	Phenanthrene	13.789	13.296	3.6	90	0.00
15	Anthracene	11.226	10.701	4.7	93	0.00
16 SURR	Fluoranthene-d10	10.091	10.204	-1.1	97	0.00
17 C	Fluoranthene	14.171	14.260	-0.6	98	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
19	Pyrene	17.437	16.604	4.8	99	0.00
20	Benzo(a)anthracene	11.843	11.538	2.6	101	0.00
21	Chrysene	16.264	16.134	0.8	108	0.00
22 I	Perylene-d12	1.000	1.000	0.0	109	0.00
23	Benzo(b)fluoranthene	17.883	17.387	2.8	106	0.00
24	Benzo(k)fluoranthene	18.118	17.227	4.9	104	0.00
25 C	Benzo(a)pyrene	15.046	14.486	3.7	105	0.00
26	Indeno(1,2,3-cd)pyrene	16.523	15.775	4.5	103	0.00
27	Dibenzo(a,h)anthracene	12.860	12.249	4.8	103	0.00
28	Benzo(g,h,i)perylene	15.728	15.012	4.6	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0