


Data Path : Z:\HPCHEM1\BNA_M\Data\BM032216\
 Data File : BM004721.D
 Acq On : 22 Mar 2016 10:04
 Operator : UM/SJ
 Sample : DFTPP38
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP38

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM032216.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 23 06:08:08 2016



AutoFind: Scans 2543, 2544, 2545; Background Corrected with Scan 2536

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.2	55597	PASS
68	69	0.00	2	1.7	994	PASS
69	198	0.00	100	32.6	58082	PASS
70	69	0.00	2	0.4	229	PASS
127	198	10	80	44.1	78498	PASS
197	198	0.00	1	0.6	1131	PASS
198	198	100	100	100.0	178112	PASS
199	198	5	9	6.9	12207	PASS
275	198	10	60	25.2	44906	PASS
365	198	1	100	3.1	5587	PASS
441	443	0.01	100	80.7	25616	PASS
442	198	50	100	91.9	163645	PASS
443	442	15	24	19.4	31736	PASS

m/z	Abundance
38.00	637.0
39.10	3691.0
40.00	457.0
44.00	406.0
49.00	412.0
50.10	14903.0
51.10	58064.0
52.00	2854.0
55.00	341.0
56.00	1663.0
57.00	3945.0
61.00	828.0
62.00	853.0
63.00	2297.0
64.00	303.0
65.10	1194.0
68.10	936.0
69.00	59744.0
70.00	303.0
72.90	446.0
74.10	6144.0
75.00	9814.0
76.10	3448.0
77.10	76024.0
78.10	5526.0
79.00	4593.0
80.00	3541.0
81.00	5272.0
82.00	1335.0
83.00	1281.0
85.00	1067.0
86.00	1557.0
87.00	761.0
91.00	1135.0
92.00	1438.0
93.00	7984.0
94.00	446.0
96.00	447.0
98.00	6695.0
99.00	6043.0
100.00	534.0
101.00	3090.0
103.00	936.0
104.00	1762.0
105.00	1748.0
106.00	534.0
107.00	23416.0
108.00	3458.0
109.00	631.0
110.00	41360.0
111.00	6464.0
112.00	745.0
116.00	1249.0
117.00	19216.0
118.00	1590.0
122.00	1397.0
123.00	2195.0
124.00	1039.0
125.00	976.0
126.20	410.0
127.00	77072.0
128.00	5890.0
129.00	29104.0
130.00	2740.0
131.00	581.0
134.00	840.0
135.00	2394.0
136.00	960.0
137.10	1672.0
137.90	345.0
141.00	3771.0
142.00	1268.0
143.00	890.0
146.00	671.0
147.00	1968.0
148.00	4658.0
149.00	1089.0
149.90	377.0
150.90	354.0
153.00	1117.0
154.00	973.0
155.00	2257.0
156.00	3324.0
157.00	520.0
157.90	763.0
158.90	584.0
160.00	1241.0
161.00	1798.0
161.90	489.0
165.00	1256.0
166.00	1047.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

167.00	8341.0
168.00	6243.0
169.00	890.0
171.00	350.0
171.90	662.0
173.00	1071.0
174.00	1654.0
175.00	3137.0
176.00	842.0
177.00	1371.0
178.00	507.0
179.00	5843.0
180.00	4608.0
181.00	1930.0
182.00	329.0
184.00	447.0
185.00	2828.0
186.10	22432.0
187.10	6391.0
188.00	763.0
189.00	1249.0
191.00	760.0
192.00	1972.0
193.10	2222.0
194.00	556.0
196.00	4353.0
197.10	1035.0
198.00	163200.0
199.00	11376.0
200.00	912.0
201.60	666.0
203.00	1150.0
204.10	6148.0
205.10	10037.0
206.10	42368.0
207.10	6286.0
208.00	1518.0
209.00	480.0
210.10	672.0
211.00	1564.0
215.00	560.0
216.00	986.0
217.00	11173.0
218.00	1542.0
221.10	6019.0
223.00	2448.0
224.10	23248.0
225.10	5997.0
226.10	752.0
227.00	9869.0
228.00	1425.0
229.00	2037.0
231.00	823.0
234.00	652.0
235.00	727.0
236.00	473.0
237.00	1180.0
239.00	351.0
240.10	321.0
241.00	526.0
242.10	1242.0
243.10	1372.0
244.10	17648.0
245.10	2416.0
246.00	3201.0
247.00	712.0
249.00	663.0
253.00	425.0
254.10	728.0
255.10	88168.0
256.10	13186.0
257.00	1253.0
258.00	4934.0
259.00	775.0
265.10	1887.0
273.00	2720.0
274.00	7167.0
275.10	38336.0
276.10	5006.0
277.10	3510.0
278.00	731.0
283.10	328.0
285.10	587.0
293.00	780.0
296.00	10334.0
297.00	1616.0
303.00	1180.0
314.00	547.0
315.00	1198.0
316.00	666.0
321.00	370.0
323.10	3338.0
324.00	595.0
327.00	633.0
328.00	327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

333.00	317.0
334.10	2123.0
335.00	551.0
341.10	395.0
346.00	665.0
352.00	932.0
353.00	671.0
354.10	1100.0
365.00	4221.0
366.00	741.0
372.10	1656.0
373.00	523.0
383.00	468.0
402.00	758.0
403.00	1032.0
404.10	487.0
421.00	916.0
422.00	961.0
423.10	6281.0
424.10	2616.0
425.00	305.0
441.10	17976.0
442.10	115640.0
443.10	22544.0
444.10	2291.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
38.00	578.0
39.10	4277.0
40.00	371.0
44.00	387.0
49.00	451.0
50.10	15859.0
51.10	63464.0
52.00	3192.0
56.00	1860.0
57.00	4254.0
61.00	855.0
62.00	995.0
63.00	2695.0
64.00	341.0
65.00	1454.0
68.10	1198.0
69.00	66216.0
70.00	384.0
73.00	487.0
74.00	7043.0
75.00	11445.0
76.10	3842.0
77.10	84256.0
78.10	5961.0
79.00	5108.0
80.00	3651.0
81.00	5646.0
82.00	1503.0
83.00	1372.0
85.00	1322.0
86.00	1623.0
87.00	747.0
88.00	390.0
91.00	1157.0
92.00	1590.0
93.00	9014.0
94.00	654.0
96.10	362.0
98.00	7180.0
99.00	6969.0
100.00	521.0
101.00	3292.0
103.00	1094.0
104.00	2055.0
105.00	1945.0
106.00	680.0
107.00	26880.0
108.10	4021.0
109.00	824.0
110.00	49280.0
111.00	7187.0
112.00	1000.0
113.00	332.0
116.00	1347.0
117.00	22328.0
118.00	1517.0
122.00	1557.0
123.00	2574.0
124.00	1269.0
125.00	1157.0
126.20	561.0
127.00	90312.0
128.10	6895.0
129.00	34104.0
130.00	2830.0
131.00	490.0
134.00	1013.0
135.00	2911.0
136.00	1210.0
137.00	1620.0
138.00	389.0
140.00	453.0
141.00	4488.0
142.00	1475.0
143.00	1046.0
144.00	331.0
146.00	860.0
147.00	2395.0
148.00	5939.0
149.00	1301.0
150.00	435.0
151.00	706.0
151.80	436.0
153.00	1332.0
154.00	1219.0
155.00	2760.0
156.10	3950.0
157.00	699.0
158.00	759.0
159.00	646.0
160.00	1548.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

161.00	2426.0
162.00	686.0
165.00	1751.0
166.10	1622.0
167.00	10380.0
168.00	6688.0
169.10	889.0
170.10	345.0
171.00	408.0
172.00	824.0
173.10	1208.0
174.00	2167.0
175.10	4022.0
176.00	1214.0
177.00	1759.0
178.00	552.0
179.00	7152.0
180.00	5753.0
181.00	2282.0
182.00	393.0
184.00	463.0
185.00	3787.0
186.10	28120.0
187.00	8073.0
188.00	1025.0
189.00	1664.0
189.90	313.0
191.00	790.0
192.00	2409.0
193.00	2897.0
194.00	564.0
195.00	325.0
196.00	5573.0
197.10	1315.0
198.00	207424.0
199.00	14599.0
200.00	1069.0
201.50	1064.0
203.10	1397.0
204.00	7482.0
205.10	12529.0
206.10	54504.0
207.10	7968.0
208.00	1758.0
209.00	569.0
210.00	981.0
211.10	2123.0
215.00	521.0
216.00	1069.0
217.00	13995.0
218.00	1805.0
221.10	8314.0
222.00	1684.0
223.00	3232.0
224.10	30080.0
225.10	7736.0
226.00	822.0
227.00	12438.0
228.10	1763.0
229.10	2735.0
230.00	305.0
231.00	1054.0
234.00	775.0
235.00	836.0
236.00	600.0
237.00	1193.0
239.00	403.0
240.00	441.0
241.00	818.0
242.10	1731.0
243.00	1719.0
244.10	23632.0
245.10	3187.0
246.00	4818.0
247.00	997.0
249.00	865.0
253.00	591.0
254.10	982.0
255.10	117632.0
256.10	17336.0
257.00	1412.0
258.10	6635.0
259.00	1046.0
265.10	2713.0
266.00	343.0
272.00	379.0
273.10	3613.0
274.10	9288.0
275.10	52048.0
276.10	7238.0
277.00	4719.0
278.00	813.0
283.00	530.0
284.10	411.0
285.10	970.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

293.00	989.0
296.00	14909.0
297.00	2034.0
303.00	1545.0
304.00	419.0
314.00	766.0
315.00	1594.0
316.00	840.0
320.90	427.0
323.10	5064.0
324.00	744.0
327.00	898.0
328.00	394.0
332.10	349.0
333.00	369.0
334.10	2930.0
335.00	826.0
341.00	467.0
346.00	988.0
352.10	1430.0
353.00	1118.0
354.10	1454.0
365.00	6314.0
366.00	888.0
371.00	425.0
372.10	2604.0
373.10	648.0
383.00	714.0
390.00	405.0
402.00	1085.0
403.10	1600.0
404.00	501.0
421.00	1539.0
422.00	1376.0
423.10	9422.0
424.10	3151.0
425.00	433.0
441.10	29240.0
442.10	189824.0
443.10	37232.0
444.10	3515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
38.10	481.0
39.10	3072.0
40.00	341.0
44.00	391.0
49.00	357.0
50.10	11298.0
51.10	45264.0
52.00	2284.0
56.00	1252.0
57.00	3053.0
61.00	527.0
62.00	700.0
63.00	1913.0
64.00	335.0
65.00	900.0
68.00	850.0
69.00	48288.0
73.00	525.0
74.00	5118.0
75.00	8172.0
76.10	2987.0
77.10	60384.0
78.10	4089.0
79.00	3732.0
80.00	2600.0
81.00	4089.0
82.00	999.0
83.00	877.0
85.00	973.0
86.00	1175.0
87.00	487.0
91.00	783.0
92.00	1085.0
93.00	6597.0
94.00	362.0
96.00	376.0
98.00	5361.0
99.00	4482.0
100.00	388.0
101.00	2540.0
103.00	778.0
104.00	1520.0
105.00	1350.0
106.00	491.0
107.00	19616.0
108.00	2874.0
109.10	430.0
110.00	34936.0
111.00	5264.0
112.00	640.0
116.00	983.0
117.00	16440.0
118.00	1252.0
122.00	1321.0
123.00	1960.0
124.00	974.0
125.00	806.0
126.20	379.0
127.00	68112.0
128.00	5105.0
129.00	25952.0
130.00	2384.0
131.00	497.0
134.00	616.0
135.00	2244.0
136.00	820.0
137.00	1093.0
140.00	314.0
141.00	3133.0
142.00	1209.0
143.00	783.0
146.00	583.0
147.00	1853.0
148.00	4414.0
149.00	873.0
151.10	436.0
153.00	1170.0
154.00	793.0
155.10	2024.0
156.10	3078.0
157.10	493.0
158.00	716.0
158.90	552.0
160.00	1077.0
161.00	1689.0
162.00	487.0
165.00	1250.0
166.00	1123.0
167.00	7999.0
168.00	4853.0
169.00	782.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

172.00	629.0
173.10	765.0
174.00	1603.0
175.00	3023.0
176.00	763.0
177.00	1374.0
178.10	456.0
179.00	5473.0
180.00	4115.0
181.00	1842.0
182.00	337.0
184.00	500.0
185.00	2938.0
186.10	21992.0
187.10	6288.0
188.00	732.0
189.00	1306.0
191.00	726.0
192.00	1864.0
193.00	2095.0
194.00	378.0
196.00	4892.0
197.10	1043.0
198.00	163712.0
199.00	10647.0
200.00	795.0
201.50	623.0
203.00	1204.0
204.10	5928.0
205.10	9979.0
206.10	42992.0
207.10	5951.0
208.00	1494.0
209.00	637.0
210.10	683.0
211.00	1671.0
211.70	354.0
215.00	481.0
216.10	902.0
217.00	11075.0
218.00	1334.0
221.10	6883.0
221.90	1594.0
223.10	2635.0
224.10	24408.0
225.10	6442.0
226.00	634.0
227.00	9992.0
228.00	1315.0
229.00	2069.0
231.00	853.0
234.00	667.0
235.00	602.0
236.00	442.0
237.00	1087.0
238.90	329.0
240.00	326.0
241.00	485.0
242.00	1257.0
243.00	1419.0
244.10	19064.0
245.10	2437.0
246.00	3543.0
247.00	857.0
249.00	690.0
253.00	351.0
254.10	777.0
255.10	99096.0
256.10	14161.0
257.00	1277.0
258.00	5616.0
259.00	919.0
265.00	2486.0
273.00	3099.0
274.10	8012.0
275.10	44336.0
276.10	5417.0
277.00	3661.0
278.00	639.0
283.10	474.0
285.10	566.0
293.00	895.0
296.00	12515.0
297.00	1858.0
303.00	1585.0
304.10	457.0
314.00	560.0
315.00	1398.0
316.00	804.0
321.00	358.0
323.10	4098.0
324.00	771.0
327.00	833.0
328.00	399.0
333.00	336.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

334.10	2907.0
335.10	768.0
341.00	411.0
346.00	851.0
352.00	1285.0
353.00	985.0
354.10	1471.0
365.00	6228.0
366.00	819.0
372.10	2293.0
373.00	636.0
383.00	721.0
390.00	307.0
402.10	926.0
403.10	1255.0
404.00	487.0
421.00	1192.0
422.00	1269.0
423.10	9388.0
424.10	2669.0
425.10	317.0
441.10	29632.0
442.10	185472.0
443.10	35432.0
444.10	3119.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

Scan 2536 (17.603 min): BM004721.D
DFTPP38

m/z	Abundance
44.00	477.0
207.00	632.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38