

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032224\
 Data File : BM044725.D
 Acq On : 22 Mar 2024 07:21
 Operator : MA/JU
 Sample : P1813-08MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E0LX7MS

Quant Time: Mar 22 07:51:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM031124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 10:51:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.694	152	4086	0.400	ng/ul	0.00
4) Naphthalene-d8	10.468	136	13263	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.331	164	7771	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.086	188	12961	0.400	ng/ul	#-0.02
17) Chrysene-d12	21.286	240	9764	0.400	ng/ul	#-0.02
23) Perylene-d12	23.525	264	11042	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.260	96	1270	0.237	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.063	152	7438	0.390	ng/ul	-0.02
18) Fluoranthene-d10	19.123	212	13772	0.402	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	4388	0.802	ng/ul#	57
5) Naphthalene	10.518	128	13648	0.366	ng/ul	97
7) 2-Methylnaphthalene	12.140	142	8388	0.355	ng/ul	87
8) 1-Methylnaphthalene	12.360	142	8730	0.366	ng/ul	96
10) Acenaphthylene	14.053	152	14195	0.360	ng/ul#	95
11) Acenaphthene	14.396	153	9736	0.341	ng/ul	98
12) Fluorene	15.390	166	10627	0.326	ng/ul	99
14) Pentachlorophenol	16.736	266	4922	1.574	ng/ul	98
15) Phenanthrene	17.128	178	16583	0.426	ng/ul	94
16) Anthracene	17.221	178	17496	0.479	ng/ul	95
19) Fluoranthene	19.151	202	18294	0.383	ng/ul#	90
20) Pyrene	19.518	202	19025	0.388	ng/ul	95
21) Benzo(a)anthracene	21.271	228	17172	0.462	ng/ul	96
22) Chrysene	21.321	228	16929	0.406	ng/ul	96
24) Benzo(b)fluoranthene	22.852	252	17145	0.407	ng/ul	87
25) Benzo(k)fluoranthene	22.899	252	18274	0.400	ng/ul#	89
26) Benzo(a)pyrene	23.428	252	15880	0.415	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	25.782	276	20960	0.476	ng/ul#	97
28) Dibenzo(a,h)anthracene	25.802	278	16779	0.494	ng/ul	97
29) Benzo(g,h,i)perylene	26.470	276	17337	0.449	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032224\
Data File : BM044725.D
Acq On : 22 Mar 2024 07:21
Operator : MA/JU
Sample : P1813-08MS
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E0LX7MS

Quant Time: Mar 22 07:51:01 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM031124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 10:51:03 2024
Response via : Initial Calibration

