


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032317\
Data File : BM009336.D
Acq On : 23 Mar 2017 09:19
Operator : SJ/MA
Sample : DFTPP52
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP52

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121416.M
Title : SVOA CALIBRATION
Last Update : Mon Dec 19 15:00:56 2016



AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.7	83221	PASS
68	69	0.00	2	1.7	1360	PASS
69	198	0.00	100	47.6	81346	PASS
70	69	0.00	2	0.3	259	PASS
127	198	10	80	45.8	78157	PASS
197	198	0.00	2	0.6	1015	PASS
198	198	100	100	100.0	170762	PASS
199	198	5	9	7.0	12011	PASS
275	198	10	60	28.7	48970	PASS
365	198	1	100	3.8	6491	PASS
441	443	0.01	100	81.5	19139	PASS
442	198	50	100	74.6	127405	PASS
443	442	15	24	18.4	23488	PASS

m/z	Abundance
37.00	556.0
38.00	1707.0
39.10	9224.0
40.00	849.0
41.00	367.0
43.90	513.0
48.70	457.0
49.10	750.0
50.10	31000.0
51.10	109752.0
52.00	5896.0
53.20	374.0
55.00	559.0
56.00	2781.0
57.00	6984.0
58.00	638.0
61.00	1043.0
62.10	1753.0
63.00	3399.0
64.00	385.0
65.10	2228.0
65.80	367.0
68.00	1652.0
69.00	104728.0
70.10	450.0
73.10	1788.0
74.00	11490.0
75.00	17176.0
76.00	7047.0
77.10	125544.0
78.10	6709.0
79.10	7812.0
80.00	5585.0
81.00	10381.0
82.00	2448.0
83.00	1752.0
84.90	1865.0
86.10	3024.0
86.90	1873.0
88.00	866.0
91.00	2154.0
92.00	2567.0
92.90	15006.0
93.90	695.0
96.10	400.0
98.00	11873.0
99.00	9876.0
100.10	682.0
101.10	4607.0
101.90	535.0
103.00	1674.0
103.90	3242.0
104.90	3352.0
106.00	950.0
107.00	36336.0
108.00	5157.0
109.10	1310.0
110.00	53040.0
111.00	8208.0
112.00	1517.0
112.80	517.0
115.90	1300.0
117.00	39360.0
117.90	2851.0
118.80	326.0
119.90	453.0
122.00	2289.0
123.00	3045.0
123.90	2163.0
124.90	2118.0
126.10	625.0
127.00	101592.0
128.00	7555.0
129.00	42104.0
130.00	3471.0
131.10	986.0
131.90	672.0
133.90	969.0
135.00	3397.0
136.00	1246.0
137.00	2159.0
137.90	656.0
139.90	467.0
141.00	6063.0
141.90	2380.0
142.80	1109.0
143.90	550.0
145.70	707.0
146.20	741.0
147.10	3270.0
148.00	8594.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

149.00	1918.0
150.90	1038.0
152.90	1868.0
154.00	1629.0
154.90	3277.0
156.00	5410.0
157.10	710.0
157.80	714.0
158.80	832.0
160.00	1990.0
161.00	2190.0
162.10	712.0
164.90	2225.0
166.00	1968.0
167.00	12995.0
168.00	9135.0
168.80	787.0
169.00	798.0
169.90	637.0
170.90	1080.0
171.90	1803.0
173.00	1372.0
174.00	2217.0
175.00	4172.0
175.90	1217.0
176.90	2509.0
178.10	866.0
178.90	8905.0
180.00	6731.0
181.10	2734.0
182.30	497.0
182.90	317.0
184.00	1111.0
185.00	4459.0
186.10	28816.0
186.90	8576.0
188.00	918.0
188.90	2393.0
189.80	494.0
191.00	1146.0
192.00	2998.0
193.10	2939.0
193.90	678.0
195.00	482.0
195.90	5699.0
196.80	1430.0
197.00	1400.0
198.00	199360.0
198.90	14620.0
199.80	1247.0
200.90	479.0
201.20	452.0
202.00	455.0
202.90	2057.0
204.00	9485.0
205.00	14203.0
206.00	57424.0
207.10	9763.0
208.10	2471.0
208.80	478.0
209.00	483.0
209.90	1497.0
210.80	2054.0
211.20	1647.0
214.70	459.0
215.10	529.0
216.10	1618.0
217.00	18936.0
217.90	1850.0
219.10	319.0
220.00	785.0
221.10	8759.0
221.90	2765.0
223.00	4665.0
224.00	33072.0
225.00	7479.0
226.10	1693.0
227.00	14242.0
227.80	1308.0
229.00	2406.0
230.20	330.0
231.10	951.0
231.90	350.0
234.10	994.0
235.10	810.0
235.80	546.0
237.00	2437.0
237.80	593.0
238.90	1127.0
239.80	489.0
240.90	447.0
242.00	1495.0
243.10	2363.0
244.00	20864.0
245.00	4576.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

246.00	6222.0
246.80	712.0
249.00	1114.0
251.70	401.0
253.10	891.0
254.00	618.0
255.00	127496.0
256.00	16029.0
256.90	1938.0
258.00	8499.0
258.80	956.0
265.00	4022.0
266.00	407.0
270.80	337.0
273.00	3459.0
274.00	8559.0
275.00	51944.0
276.10	8089.0
277.00	5777.0
278.00	1010.0
281.00	499.0
282.90	496.0
284.00	405.0
285.10	1305.0
290.00	304.0
290.90	320.0
291.90	396.0
293.00	1182.0
295.10	422.0
296.00	16432.0
296.90	2018.0
300.70	319.0
303.00	1529.0
304.00	501.0
313.10	340.0
314.00	810.0
315.00	1698.0
316.10	581.0
321.00	627.0
323.00	5232.0
324.10	881.0
326.80	688.0
327.70	525.0
333.10	656.0
334.00	2488.0
335.00	634.0
341.00	1105.0
341.90	447.0
346.00	1958.0
347.10	402.0
352.10	1295.0
353.00	764.0
354.20	2091.0
365.00	5827.0
366.00	1379.0
370.90	560.0
371.90	2297.0
373.00	644.0
383.10	648.0
402.00	993.0
403.00	1431.0
403.70	305.0
404.10	367.0
421.10	985.0
422.00	1399.0
423.10	5258.0
424.10	2468.0
425.00	350.0
441.10	15667.0
442.10	116232.0
443.10	22648.0
444.10	2764.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

m/z	Abundance
37.00	683.0
37.90	1929.0
39.10	7071.0
40.00	1482.0
44.10	949.0
49.10	963.0
50.10	22120.0
51.10	95000.0
52.00	4974.0
54.90	433.0
56.00	2484.0
57.10	4892.0
58.90	410.0
61.00	1440.0
61.90	1430.0
63.00	4800.0
64.00	311.0
65.00	2085.0
68.10	1366.0
69.00	95400.0
73.00	1936.0
74.00	9243.0
75.00	17480.0
76.00	5454.0
77.10	112064.0
78.10	7444.0
79.10	6570.0
79.90	5980.0
81.00	9216.0
81.90	1757.0
82.90	1668.0
84.00	342.0
84.90	960.0
86.10	2379.0
87.10	955.0
88.00	815.0
89.10	519.0
91.00	1359.0
92.00	1720.0
93.00	14466.0
94.00	1039.0
94.80	307.0
95.90	1012.0
98.00	9682.0
99.10	9088.0
99.90	497.0
101.00	4524.0
102.00	382.0
103.00	1453.0
104.00	3419.0
105.00	3383.0
106.10	571.0
107.00	30808.0
108.00	5670.0
109.00	1252.0
110.00	51360.0
110.90	8567.0
111.90	1218.0
113.00	550.0
115.90	1098.0
117.00	31664.0
118.00	2745.0
118.90	344.0
120.00	658.0
121.90	1409.0
123.00	2847.0
123.90	1394.0
125.00	1124.0
127.00	89704.0
128.00	5754.0
129.00	36272.0
130.00	3472.0
130.90	1373.0
133.00	604.0
134.00	1666.0
135.00	3372.0
135.90	1019.0
137.00	1538.0
137.60	452.0
139.80	488.0
141.00	6681.0
142.10	1786.0
143.00	1271.0
144.90	338.0
146.10	1425.0
147.00	3659.0
148.00	7772.0
148.90	1354.0
150.90	430.0
151.40	507.0
152.10	503.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

152.90	2896.0
153.90	1209.0
154.90	2091.0
156.00	4140.0
156.90	1072.0
157.70	907.0
158.10	1363.0
158.90	1124.0
159.90	1313.0
161.10	1604.0
161.90	313.0
162.90	364.0
164.10	559.0
164.90	2469.0
166.00	1532.0
167.00	14646.0
167.90	7686.0
169.00	1178.0
170.10	836.0
171.10	664.0
171.80	906.0
172.90	1116.0
173.90	2538.0
175.00	5619.0
176.10	2426.0
177.00	3076.0
178.00	788.0
179.00	9282.0
180.00	7503.0
181.10	2215.0
182.20	459.0
184.20	852.0
185.00	3884.0
186.00	29576.0
187.00	9582.0
187.80	531.0
189.00	2708.0
189.80	621.0
190.90	1100.0
192.00	3281.0
193.00	3417.0
194.20	718.0
195.00	627.0
195.90	5371.0
197.00	1647.0
198.00	202112.0
199.00	12899.0
200.10	1402.0
201.40	588.0
203.10	1985.0
204.00	9499.0
205.10	15830.0
206.00	62416.0
207.00	10060.0
208.00	2420.0
209.00	1433.0
210.00	1269.0
210.80	1943.0
212.00	443.0
214.90	820.0
215.90	1116.0
217.00	17984.0
217.90	2800.0
221.00	8008.0
221.90	2448.0
223.00	5143.0
224.00	33120.0
225.00	9071.0
225.80	547.0
226.20	539.0
227.00	16872.0
228.00	1860.0
228.90	3160.0
230.20	342.0
230.90	1309.0
234.00	1331.0
235.00	1221.0
235.80	1188.0
237.00	1572.0
239.00	987.0
239.70	667.0
241.00	1237.0
241.90	2110.0
243.00	1309.0
244.10	25288.0
245.00	4246.0
246.00	6198.0
247.10	723.0
248.00	328.0
249.10	974.0
250.70	445.0
252.80	517.0
253.30	449.0
254.00	1668.0
255.00	133632.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

256.00	21456.0
257.00	2602.0
258.00	10631.0
258.80	820.0
259.00	817.0
261.10	342.0
265.00	3999.0
266.00	353.0
267.00	351.0
271.00	447.0
272.90	3534.0
274.00	10115.0
275.00	60992.0
276.00	7705.0
277.00	5689.0
278.00	885.0
281.10	427.0
282.10	606.0
282.90	532.0
283.90	1018.0
285.10	1282.0
289.10	348.0
293.00	981.0
294.90	463.0
296.00	21552.0
297.00	2703.0
301.10	517.0
301.70	399.0
303.00	2168.0
304.00	928.0
307.90	595.0
314.10	592.0
315.00	1741.0
315.70	841.0
322.00	402.0
323.10	5440.0
324.10	1046.0
327.00	860.0
328.00	878.0
333.10	853.0
334.00	2680.0
334.90	735.0
340.80	480.0
345.80	860.0
352.10	1470.0
353.00	826.0
354.00	2416.0
365.00	8212.0
365.90	1320.0
370.90	741.0
371.90	2691.0
372.90	671.0
383.00	690.0
390.90	394.0
402.00	1139.0
403.00	2536.0
403.90	415.0
421.10	1456.0
422.00	625.0
423.00	8247.0
424.00	3013.0
441.10	25160.0
442.10	148224.0
443.10	28904.0
443.90	2176.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

m/z	Abundance
38.10	828.0
39.00	3340.0
40.00	767.0
43.80	435.0
44.10	406.0
45.20	454.0
49.20	408.0
50.10	11371.0
51.10	44912.0
52.00	2345.0
55.90	1442.0
57.10	2980.0
61.00	1051.0
61.70	367.0
62.10	328.0
63.10	1398.0
63.90	464.0
65.10	900.0
68.00	1064.0
69.00	43912.0
70.00	328.0
73.10	1684.0
74.00	5561.0
74.90	7026.0
76.10	2396.0
77.10	53576.0
78.10	2409.0
79.00	4990.0
80.00	2938.0
81.00	3503.0
82.00	1374.0
83.00	1208.0
83.90	345.0
85.00	649.0
86.00	965.0
87.00	564.0
90.80	1041.0
92.00	1419.0
93.00	6212.0
93.90	823.0
95.80	577.0
98.00	4649.0
99.00	3676.0
100.10	456.0
101.00	2104.0
103.00	517.0
104.00	1576.0
104.90	1579.0
106.10	360.0
107.00	16504.0
107.90	2151.0
109.00	426.0
110.00	24920.0
111.00	3469.0
112.10	724.0
113.00	405.0
115.90	830.0
116.90	14810.0
118.00	1036.0
120.00	337.0
121.80	697.0
122.20	828.0
123.00	1636.0
123.90	406.0
124.10	408.0
125.00	952.0
126.00	372.0
127.00	43176.0
128.00	3086.0
129.00	20176.0
130.00	1246.0
130.70	504.0
133.90	1029.0
134.90	1521.0
136.00	917.0
136.90	815.0
140.00	431.0
141.00	2611.0
141.90	608.0
143.10	666.0
146.00	482.0
147.00	1659.0
147.90	4215.0
149.10	1121.0
151.10	357.0
152.90	783.0
153.80	387.0
154.90	2148.0
156.00	2137.0
156.80	509.0
157.30	383.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

157.90	706.0
159.10	360.0
159.80	440.0
161.00	2070.0
165.00	1403.0
166.10	1022.0
167.00	7330.0
168.00	3471.0
168.90	632.0
170.90	345.0
171.80	434.0
173.10	390.0
173.80	1721.0
175.00	2488.0
175.80	795.0
176.90	1210.0
177.80	394.0
178.90	5768.0
179.90	3629.0
181.10	1522.0
182.30	303.0
182.90	455.0
183.90	320.0
184.80	1946.0
186.00	18888.0
187.10	4571.0
188.00	533.0
188.90	1554.0
191.10	986.0
192.00	2053.0
193.00	2547.0
194.10	814.0
196.00	3158.0
196.80	452.0
197.90	110816.0
199.00	8514.0
199.70	505.0
201.70	573.0
203.00	973.0
204.00	5647.0
205.00	8582.0
206.00	32136.0
207.00	5209.0
208.10	1625.0
209.00	614.0
210.20	975.0
210.90	1216.0
215.00	1264.0
215.90	804.0
217.00	11631.0
217.90	1202.0
221.00	3484.0
222.10	1452.0
223.00	2197.0
224.00	21176.0
225.10	5584.0
226.10	1057.0
227.00	8989.0
228.10	1306.0
229.00	2271.0
231.00	998.0
234.00	801.0
235.10	434.0
236.00	390.0
237.00	1146.0
238.00	310.0
241.80	372.0
243.00	1147.0
244.00	13786.0
245.10	2263.0
246.00	3823.0
247.00	566.0
249.10	334.0
250.90	496.0
252.90	451.0
254.00	973.0
255.00	85328.0
256.00	12128.0
257.00	1393.0
258.00	6561.0
259.00	534.0
264.90	1709.0
265.90	462.0
271.80	834.0
272.90	3130.0
274.00	6538.0
275.00	33976.0
275.90	3543.0
276.90	3379.0
277.90	573.0
281.00	431.0
282.80	924.0
284.00	490.0
285.00	496.0
292.80	549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

296.00	10738.0
297.00	1866.0
300.90	358.0
303.10	1418.0
303.90	382.0
308.00	321.0
313.90	562.0
315.00	883.0
316.00	526.0
323.10	3680.0
324.10	831.0
326.90	1049.0
333.00	336.0
334.00	3007.0
334.90	730.0
341.10	469.0
345.90	634.0
352.10	708.0
352.90	615.0
353.80	965.0
354.20	919.0
365.00	5436.0
365.80	338.0
370.80	332.0
372.10	1924.0
373.00	432.0
382.90	379.0
384.00	359.0
390.90	401.0
402.10	1223.0
403.00	476.0
421.10	795.0
422.00	1249.0
423.10	5165.0
424.00	1503.0
441.10	16592.0
442.10	117760.0
443.10	18912.0
444.00	1782.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52

m/z	Abundance
38.90	357.0
40.00	942.0
44.10	808.0
72.80	930.0
85.60	360.0
133.00	708.0
134.90	314.0
139.80	368.0
167.10	917.0
207.00	2245.0
209.10	312.0
255.00	709.0
281.20	1095.0

Instrument :
BNA_M
ClientSampleId :
DFTPP52