

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032324\
 Data File : BM044735.D
 Acq On : 22 Mar 2024 15:24
 Operator : MA/JU
 Sample : PB159652BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB159652BL

Quant Time: Mar 22 15:56:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 02:28:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.687	152	205622	20.000	ng	-0.02
21) Naphthalene-d8	10.463	136	756517	20.000	ng	-0.02
39) Acenaphthene-d10	14.327	164	379409	20.000	ng	-0.02
64) Phenanthrene-d10	17.086	188	709514	20.000	ng	-0.02
76) Chrysene-d12	21.286	240	580424	20.000	ng	-0.02
86) Perylene-d12	23.527	264	657131	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.310	112	1221578	85.298	ng	0.00
7) Phenol-d6	6.869	99	1494427	81.111	ng	-0.01
23) Nitrobenzene-d5	8.851	82	1371456	87.335	ng	-0.01
42) 2,4,6-Tribromophenol	15.827	330	338805	80.962	ng	-0.02
45) 2-Fluorobiphenyl	12.945	172	2513228	91.228	ng	-0.02
79) Terphenyl-d14	19.733	244	3137271	94.555	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032324\
Data File : BM044735.D
Acq On : 22 Mar 2024 15:24
Operator : MA/JU
Sample : PB159652BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB159652BL

Quant Time: Mar 22 15:56:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 02:28:24 2024
Response via : Initial Calibration

