

Data Path : Z:\HPCHEM1\BNA M\DATA\BM032416\
Data File : BM004752.D
Acq On : 24 Mar 2016 12:15
Operator : UM/SJ
Sample : PB89112BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK12

Manual Integrations
APPROVED

Sohil
4/6/2016 1:28:00 PM

Quant Time: Apr 05 17:53:39 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM032216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 05 16:23:17 2016
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	1237	0.40	ng/ul	0.00
4) Naphthalene-d8	10.62	136	5017	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.47	164	2653	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.21	188	5078m	0.40	ng/ul	0.00
18) Chrysene-d12	21.40	240	3712	0.40	ng/ul	0.00
22) Perylene-d12	23.69	264	2373m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.32	96	8560	14.28	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.22	152	2246	0.39	ng/ul	0.00
16) Fluoranthene-d10	19.25	212	4283	0.42	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM032416\
Data File : BM004752.D
Acq On : 24 Mar 2016 12:15
Operator : UM/SJ
Sample : PB89112BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK12

Manual Integrations
APPROVED

Sohil
4/6/2016 1:28:00 PM

Quant Time: Apr 05 17:53:39 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM032216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 05 16:23:17 2016
Response via : Initial Calibration

