


Data Path : Z:\HPCHEM1\BNA_M\Data\BM032417\
 Data File : BM009377.D
 Acq On : 24 Mar 2017 13:23
 Operator : SJ/MA
 Sample : DFTPP53
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP53

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032317MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Mar 27 04:18:12 2017



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.0	67722	PASS
68	69	0.00	2	1.8	1157	PASS
69	198	0.00	100	49.0	65064	PASS
70	69	0.00	2	0.4	230	PASS
127	198	10	80	48.8	64760	PASS
197	198	0.00	2	0.5	681	PASS
198	198	100	100	100.0	132680	PASS
199	198	5	9	6.8	8996	PASS
275	198	10	60	28.5	37778	PASS
365	198	1	100	3.9	5118	PASS
441	443	0.01	100	88.6	16041	PASS
442	198	50	100	73.3	97205	PASS
443	442	15	24	18.6	18096	PASS

m/z	Abundance
39.90	802.0
44.10	669.0
73.00	1127.0
84.00	334.0
146.90	337.0
165.10	329.0
193.00	322.0
207.00	1742.0
208.00	471.0
208.90	410.0
220.90	307.0
281.00	444.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
37.10	569.0
38.10	1018.0
39.00	7127.0
39.90	1235.0
43.90	683.0
45.00	350.0
49.10	1096.0
50.00	23400.0
51.10	76752.0
52.10	3625.0
55.00	543.0
56.00	3509.0
57.00	5884.0
57.80	420.0
61.00	1110.0
62.00	1406.0
63.00	3401.0
64.00	576.0
65.10	2065.0
66.00	394.0
68.00	1577.0
69.00	76488.0
73.10	1282.0
74.00	8467.0
75.00	14905.0
76.00	4060.0
77.10	88400.0
78.10	6824.0
79.00	6100.0
79.90	3884.0
81.00	5857.0
82.00	1639.0
82.90	1259.0
83.90	321.0
84.80	645.0
86.00	1593.0
87.10	1369.0
87.90	351.0
89.00	329.0
91.00	1327.0
91.90	1821.0
93.00	10282.0
94.10	775.0
95.80	402.0
98.00	8121.0
99.00	7883.0
100.10	699.0
101.00	3208.0
102.90	936.0
104.10	2268.0
105.00	2043.0
106.00	785.0
107.00	26176.0
108.00	4031.0
109.00	721.0
110.00	39512.0
111.00	6992.0
111.90	577.0
112.80	335.0
115.90	1062.0
117.00	28304.0
118.00	1512.0
122.00	1031.0
123.00	1603.0
123.90	1364.0
124.90	913.0
126.00	481.0
127.00	71920.0
128.00	6623.0
129.00	31200.0
130.00	2237.0
131.00	489.0
134.00	542.0
135.00	2652.0
135.90	838.0
137.00	1665.0
137.90	519.0
139.80	780.0
140.90	4038.0
141.90	1519.0
143.00	1185.0
145.80	915.0
147.10	2301.0
147.90	5940.0
149.00	1192.0
149.90	332.0
151.00	862.0
152.90	1364.0
154.00	972.0
155.00	2247.0
156.00	3199.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

157.10	709.0
158.00	1373.0
159.00	782.0
160.00	843.0
161.00	2012.0
161.80	988.0
164.80	2893.0
166.00	782.0
167.00	9419.0
167.90	5267.0
169.10	659.0
172.10	773.0
173.00	733.0
174.10	1325.0
175.00	3391.0
175.90	520.0
176.80	2150.0
177.90	563.0
178.90	5858.0
179.90	5555.0
181.10	2233.0
183.80	484.0
185.00	3463.0
186.00	19488.0
187.00	7135.0
187.90	793.0
188.90	1193.0
191.00	788.0
191.90	2223.0
193.10	2290.0
194.00	857.0
194.90	337.0
196.00	5569.0
196.60	1333.0
197.10	663.0
197.90	136000.0
198.90	9340.0
199.90	1342.0
201.30	459.0
203.00	1657.0
204.00	6867.0
205.00	10886.0
206.00	43576.0
207.00	7001.0
208.10	1740.0
208.90	570.0
209.80	944.0
211.00	1637.0
212.10	469.0
214.90	745.0
216.20	615.0
217.00	10539.0
218.00	1764.0
219.90	557.0
221.10	4737.0
221.60	1498.0
222.10	1089.0
223.00	3290.0
224.00	22392.0
225.00	6347.0
226.10	1287.0
227.00	11671.0
228.00	1345.0
229.00	2254.0
229.80	300.0
230.90	599.0
233.90	898.0
234.90	1025.0
236.00	730.0
236.90	1588.0
238.80	335.0
240.00	400.0
240.70	566.0
242.00	1849.0
243.10	606.0
244.00	16298.0
245.00	2779.0
245.90	4097.0
246.90	1331.0
249.10	460.0
252.90	559.0
253.70	579.0
255.00	90784.0
256.00	13002.0
257.00	1318.0
258.00	5490.0
258.90	1220.0
264.00	304.0
264.90	2247.0
266.10	310.0
273.00	3066.0
274.00	6278.0
275.00	35224.0
276.00	5274.0
277.00	3910.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

278.00	529.0
281.00	393.0
282.90	701.0
284.00	686.0
285.10	964.0
292.90	367.0
293.80	526.0
296.00	14784.0
296.90	1614.0
302.00	529.0
302.90	1307.0
314.10	481.0
314.90	1112.0
316.00	584.0
321.10	528.0
323.10	3585.0
324.10	1039.0
325.10	342.0
326.90	747.0
327.90	309.0
332.00	462.0
332.80	390.0
334.10	1931.0
335.00	683.0
341.00	877.0
346.00	1365.0
350.00	317.0
352.10	1001.0
352.90	563.0
354.00	1147.0
365.10	5102.0
366.00	1604.0
372.00	1629.0
373.00	316.0
382.90	362.0
401.90	771.0
403.10	599.0
403.90	322.0
405.10	399.0
421.10	341.0
422.10	587.0
423.00	5154.0
424.10	2185.0
424.90	521.0
441.00	12885.0
442.10	75888.0
443.10	14753.0
444.10	735.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
37.20	557.0
38.00	1703.0
39.00	5876.0
40.00	813.0
41.10	522.0
44.00	939.0
49.00	1223.0
50.10	20824.0
51.00	84368.0
52.10	4107.0
54.90	370.0
56.00	2014.0
57.00	4612.0
57.80	326.0
61.00	816.0
62.00	833.0
63.00	3331.0
64.00	370.0
65.00	1119.0
68.10	1194.0
69.00	78536.0
70.00	691.0
73.10	2041.0
74.00	9880.0
75.00	15277.0
76.00	5111.0
77.10	87104.0
78.10	6288.0
79.00	7264.0
80.00	4985.0
81.00	7305.0
82.00	1544.0
83.00	1350.0
85.00	1059.0
86.10	2524.0
87.00	653.0
88.90	344.0
90.90	1577.0
92.00	1208.0
93.00	11310.0
93.90	901.0
95.90	491.0
98.00	9941.0
99.10	5520.0
100.00	434.0
100.90	2564.0
102.90	887.0
104.00	2442.0
105.10	1363.0
105.90	320.0
107.00	27904.0
108.00	4061.0
109.00	934.0
110.00	38632.0
111.00	5505.0
112.00	901.0
112.80	456.0
116.10	999.0
117.00	26720.0
118.00	2018.0
118.90	597.0
122.10	2120.0
122.90	2323.0
123.90	1566.0
124.80	756.0
127.00	75152.0
128.00	6485.0
129.00	28792.0
130.00	3662.0
131.10	801.0
133.80	1063.0
134.90	2256.0
135.80	1024.0
136.90	1436.0
137.60	627.0
138.20	405.0
139.10	335.0
139.90	492.0
141.00	6071.0
141.90	1426.0
143.10	511.0
146.00	873.0
147.00	3096.0
148.00	6759.0
149.10	1271.0
150.00	407.0
151.10	726.0
152.00	452.0
153.10	1279.0
154.00	648.0
154.90	2524.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

156.00	3227.0
156.90	772.0
157.70	546.0
159.10	788.0
160.00	1154.0
161.00	2251.0
162.10	587.0
165.00	1682.0
166.00	1958.0
167.00	10798.0
168.00	7014.0
169.10	798.0
170.80	442.0
172.00	1049.0
172.90	1060.0
174.10	1354.0
175.00	5449.0
176.00	467.0
176.90	1414.0
178.00	790.0
178.90	7411.0
180.00	5373.0
180.90	1881.0
183.90	941.0
185.00	3804.0
186.00	23264.0
187.00	7822.0
188.00	1416.0
188.90	1676.0
190.90	718.0
192.10	2839.0
193.10	3149.0
193.90	516.0
195.00	430.0
196.00	3975.0
197.00	884.0
198.00	162304.0
199.00	11770.0
200.10	633.0
201.40	673.0
203.00	2211.0
204.00	7020.0
205.10	10307.0
206.00	47152.0
207.00	7748.0
208.10	1624.0
209.10	1174.0
210.10	623.0
211.00	1481.0
212.00	568.0
212.90	356.0
215.00	491.0
216.00	1471.0
216.90	15610.0
218.00	1842.0
221.00	5522.0
222.00	2141.0
222.90	2011.0
224.00	25768.0
225.10	7549.0
226.10	933.0
227.00	12208.0
227.90	2351.0
229.00	2866.0
231.00	1175.0
232.90	405.0
233.10	395.0
234.00	1286.0
234.90	1160.0
236.00	1040.0
237.00	1701.0
238.00	369.0
239.80	532.0
240.80	651.0
241.80	719.0
243.00	1495.0
244.00	22344.0
245.00	3342.0
246.00	4406.0
247.00	1066.0
249.10	747.0
251.80	493.0
253.10	301.0
253.80	542.0
254.10	495.0
255.00	109456.0
256.00	14761.0
257.10	919.0
258.00	8787.0
259.00	955.0
259.90	350.0
264.00	355.0
264.90	3045.0
270.10	341.0
272.00	494.0

Instrument :

BNA_M

ClientSampleId :

DFTPP53

273.00	2639.0
274.00	7679.0
275.00	46936.0
276.00	7059.0
277.00	3893.0
278.00	1131.0
281.10	805.0
282.00	321.0
283.00	430.0
283.90	441.0
285.00	927.0
292.90	604.0
293.80	543.0
295.00	734.0
296.00	16408.0
297.00	2586.0
301.90	449.0
303.20	1220.0
304.00	568.0
307.90	347.0
314.10	623.0
315.00	1894.0
316.00	885.0
323.10	4653.0
324.00	1133.0
326.80	728.0
328.00	356.0
333.00	815.0
334.10	2883.0
335.00	910.0
340.90	416.0
346.00	853.0
352.00	1308.0
353.00	724.0
354.10	1375.0
364.90	5641.0
366.00	1040.0
370.80	324.0
372.00	1202.0
372.90	581.0
382.90	395.0
390.10	400.0
402.00	1604.0
403.10	1311.0
403.80	327.0
420.90	987.0
422.10	426.0
423.10	5390.0
424.00	2370.0
425.00	430.0
441.10	18288.0
442.10	116704.0
443.10	21104.0
444.10	2319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
36.90	324.0
38.10	791.0
39.10	2930.0
40.10	592.0
44.00	1002.0
49.10	430.0
50.00	10210.0
51.00	42048.0
52.00	1552.0
56.10	1745.0
57.00	2491.0
60.90	485.0
62.10	573.0
63.00	2852.0
64.90	1037.0
67.90	702.0
69.00	40168.0
73.10	867.0
74.10	4723.0
75.10	8174.0
76.10	2776.0
77.10	50352.0
78.10	2467.0
79.10	3500.0
80.00	3033.0
81.00	4441.0
81.90	1079.0
82.90	937.0
85.20	519.0
86.00	989.0
87.00	599.0
87.90	358.0
91.10	998.0
92.00	1275.0
92.90	5920.0
96.00	526.0
98.00	4754.0
99.00	5033.0
100.10	424.0
100.90	1506.0
102.80	597.0
103.90	1307.0
105.00	1361.0
105.80	677.0
107.00	15249.0
108.00	1802.0
109.10	712.0
110.00	25280.0
111.00	3341.0
114.70	363.0
115.90	848.0
117.00	13834.0
117.90	1079.0
119.00	346.0
122.00	866.0
122.90	1237.0
124.00	1269.0
124.90	401.0
127.00	47208.0
128.00	3045.0
129.00	18752.0
130.10	1777.0
131.00	302.0
132.00	305.0
132.90	302.0
133.90	426.0
134.90	1098.0
136.00	410.0
137.00	992.0
138.10	414.0
140.90	2425.0
141.90	1266.0
143.00	841.0
145.80	404.0
147.00	1695.0
148.00	4052.0
149.10	504.0
151.10	417.0
153.10	827.0
154.00	576.0
155.00	1531.0
156.10	1672.0
156.90	673.0
158.00	576.0
159.80	743.0
160.80	664.0
161.80	621.0
164.90	1527.0
165.80	465.0
166.90	6269.0
168.00	2959.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

169.00	569.0
173.00	833.0
173.90	1194.0
175.00	1501.0
176.00	870.0
177.00	646.0
177.80	646.0
179.00	3263.0
180.00	3150.0
180.80	891.0
181.10	953.0
182.10	315.0
183.90	457.0
184.90	2474.0
186.00	15349.0
187.00	3726.0
188.00	1096.0
188.90	1315.0
189.80	326.0
191.00	497.0
191.90	1475.0
193.00	1877.0
194.00	921.0
194.90	411.0
196.00	3211.0
196.70	335.0
197.10	496.0
198.00	99736.0
199.00	5878.0
200.00	1067.0
202.80	735.0
204.00	4232.0
205.00	8301.0
206.00	28248.0
207.00	6504.0
207.80	1145.0
208.80	303.0
210.00	324.0
211.10	1686.0
216.10	1205.0
217.00	8269.0
218.00	1077.0
221.10	2820.0
221.90	924.0
223.00	2252.0
224.00	17008.0
225.00	4956.0
226.00	839.0
227.00	8158.0
228.00	1300.0
229.00	1518.0
230.90	1041.0
234.00	714.0
235.00	361.0
236.00	396.0
236.80	816.0
239.20	382.0
239.90	380.0
241.00	962.0
242.00	1744.0
242.90	1040.0
244.00	12636.0
245.10	1406.0
246.00	2462.0
247.10	306.0
249.10	535.0
252.10	336.0
254.10	1128.0
255.00	68728.0
256.00	10641.0
257.00	1346.0
258.00	5650.0
258.90	1170.0
265.00	1727.0
271.90	324.0
273.00	2660.0
273.90	5634.0
275.00	31176.0
276.00	5045.0
277.00	4648.0
278.10	788.0
281.10	551.0
283.10	384.0
283.90	534.0
284.90	739.0
293.00	700.0
296.00	9563.0
296.90	1025.0
303.10	2149.0
303.80	407.0
312.80	348.0
314.00	606.0
314.90	1315.0
315.90	557.0
323.00	3227.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

323.80	484.0
327.10	824.0
334.00	1479.0
335.10	549.0
345.70	426.0
346.10	518.0
352.10	1307.0
354.00	1133.0
364.90	4611.0
365.80	561.0
366.00	536.0
370.80	309.0
371.90	1255.0
373.10	429.0
382.90	482.0
402.10	602.0
402.90	922.0
404.20	447.0
421.00	904.0
422.00	1197.0
423.10	5521.0
423.90	1371.0
441.10	16952.0
442.10	99024.0
443.10	18432.0
444.10	2091.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53