


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032422\
 Data File : BM034359.D
 Acq On : 24 Mar 2022 11:54
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Mar 24 02:40:59 2022




AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	120359	PASS
68	69	0.00	2	1.9	2540	PASS
69	198	0.00	100	43.1	132757	PASS
70	69	0.00	2	0.6	839	PASS
127	198	10	80	49.4	152277	PASS
197	198	0.00	2	0.6	1861	PASS
198	198	100	100	100.0	308267	PASS
199	198	5	9	6.9	21325	PASS
275	198	10	60	28.5	87808	PASS
365	198	1	100	4.0	12479	PASS
441	198	0.01	100	12.1	37373	PASS
442	442	50	100	100.0	250837	PASS
443	442	15	24	19.0	47608	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032422\
Data File : BM034359.D
Acq On : 24 Mar 2022 11:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Mar 25 02:06:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 02:40:59 2022
Response via : Initial Calibration



(70) Pentachlorophenol (C)

16.977min (+ 0.006) 590591.13 ng




response 377725

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.30	63.81
264.00	62.90	62.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032422\
Data File : BM034359.D
Acq On : 24 Mar 2022 11:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Mar 25 02:06:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 02:40:59 2022
Response via : Initial Calibration



TIC: BM034359.D\data.ms

(77) Benzidine

19.559min (+ 0.006) 114027.88 ng

response 1854816

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	13.71
183.00	12.20	11.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/24/2022	BNA_M	<u>BM034359.D</u>
Compound Name	Response	Retention Time
DDT	1139210	20.795
DDD	17825	20.353
DDE	2022	19.848
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
19847	1159057	1.71