

Data Path : Z:\HPCHEM1\BNA M\DATA\BM032516\
 Data File : BM004778.D
 Acq On : 25 Mar 2016 15:26
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.442

Quant Time: Apr 05 18:44:58 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 05 18:29:16 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2 SURR	1,4-Dioxane-d8	0.180	0.185	-2.8	98	0.00
3	1,4-Dioxane	0.220	0.222	-0.9	96	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
5	Naphthalene	0.966	0.974	-0.8	97	0.00
6 SURR	2-Methylnaphthalene-d10	0.558	0.568	-1.8	98	0.00
7	2-Methylnaphthalene	0.692	0.709	-2.5	97	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
9	Acenaphthylene	1.840	1.871	-1.7	95	0.00
10 C	Acenaphthene	1.302	1.338	-2.8	96	0.00
11	Fluorene	1.554	1.574	-1.3	95	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
13	Pentachlorophenol	0.100	0.105	-5.0	99	0.00
14	Phenanthrene	1.124	1.161	-3.3	97	0.00
15	Anthracene	1.062	1.071	-0.8	95	0.00
16 SURR	Fluoranthene-d10	0.946	0.958	-1.3	94	0.00
17 C	Fluoranthene	1.279	1.299	-1.6	95	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
19	Pyrene	1.370	1.356	1.0	95	0.00
20	Benzo(a)anthracene	1.230	1.239	-0.7	95	0.00
21	Chrysene	1.188	1.223	-2.9	96	0.00
22 I	Perylene-d12	1.000	1.000	0.0	93	0.00
23	Benzo(b)fluoranthene	1.348	1.449	-7.5	102	0.00
24	Benzo(k)fluoranthene	1.249	1.234	1.2	90	0.00
25 C	Benzo(a)pyrene	1.232	1.264	-2.6	94	0.00
26	Indeno(1,2,3-cd)pyrene	1.140	1.158	-1.6	94	0.00
27	Dibenzo(a,h)anthracene	0.898	0.906	-0.9	94	0.00
28	Benzo(q,h,i)perylene	0.944	0.971	-2.9	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0