

Data Path : Z:\HPCHEM1\BNA M\DATA\BM032516\
 Data File : BM004792.D
 Acq On : 25 Mar 2016 23:57
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.443

Quant Time: Apr 05 19:43:14 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 05 18:29:16 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2 SURR	1,4-Dioxane-d8	0.180	0.181	-0.6	77	0.00
3	1,4-Dioxane	0.220	0.221	-0.5	77	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
5	Naphthalene	0.966	0.963	0.3	78	0.00
6 SURR	2-Methylnaphthalene-d10	0.558	0.544	2.5	76	0.00
7	2-Methylnaphthalene	0.692	0.696	-0.6	77	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
9	Acenaphthylene	1.840	1.825	0.8	73	0.00
10 C	Acenaphthene	1.302	1.349	-3.6	76	0.00
11	Fluorene	1.554	1.569	-1.0	75	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
13	Pentachlorophenol	0.100	0.110	-10.0	83	0.00
14	Phenanthrene	1.124	1.140	-1.4	76	0.00
15	Anthracene	1.062	1.064	-0.2	75	0.00
16 SURR	Fluoranthene-d10	0.946	0.986	-4.2	77	-0.01
17 C	Fluoranthene	1.279	1.360	-6.3	79	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	87	-0.01
19	Pyrene	1.370	1.255	8.4	80	0.00
20	Benzo(a)anthracene	1.230	1.226	0.3	86	-0.01
21	Chrysene	1.188	1.215	-2.3	87	-0.01
22 I	Perylene-d12	1.000	1.000	0.0	96	-0.01
23	Benzo(b)fluoranthene	1.348	1.306	3.1	96	-0.01
24	Benzo(k)fluoranthene	1.249	1.258	-0.7	95	-0.01
25 C	Benzo(a)pyrene	1.232	1.260	-2.3	98	-0.01
26	Indeno(1,2,3-cd)pyrene	1.140	1.314	-15.3	111	-0.01
27	Dibenzo(a,h)anthracene	0.898	1.049	-16.8	113	-0.01
28	Benzo(q,h,i)perylene	0.944	1.098	-16.3	112	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0