

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032524\
 Data File : BM044756.D
 Acq On : 25 Mar 2024 15:22
 Operator : MA/JU
 Sample : P1845-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BELL-CON

Quant Time: Mar 25 15:56:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 02:28:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	183446	20.000	ng	-0.01
21) Naphthalene-d8	10.469	136	713794	20.000	ng	-0.02
39) Acenaphthene-d10	14.333	164	395780	20.000	ng	-0.02
64) Phenanthrene-d10	17.092	188	763141	20.000	ng	-0.01
76) Chrysene-d12	21.297	240	547859	20.000	ng	# 0.00
86) Perylene-d12	23.556	264	583793	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.304	112	1041584	81.522	ng	0.00
7) Phenol-d6	6.875	99	1309076	79.640	ng	0.00
23) Nitrobenzene-d5	8.851	82	808668	54.579	ng	-0.01
42) 2,4,6-Tribromophenol	15.833	330	389416	89.207	ng	-0.01
45) 2-Fluorobiphenyl	12.951	172	1583811	55.113	ng	-0.02
79) Terphenyl-d14	19.739	244	1869382	59.691	ng	0.00
Target Compounds						
						Qvalue
22) Acetophenone	8.516	105	90527	4.827	ng	# 94
32) Benzoic acid	9.781	122	86318	12.617	ng	94
74) Di-n-butylphthalate	18.074	149	129112	2.515	ng	# 94
75) Fluoranthene	19.156	202	146264	3.103	ng	96
78) Pyrene	19.521	202	100360	2.499	ng	# 90
84) Bis(2-ethylhexyl)phtha...	21.227	149	960766	35.240	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032524\
Data File : BM044756.D
Acq On : 25 Mar 2024 15:22
Operator : MA/JU
Sample : P1845-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BELL-CON

Quant Time: Mar 25 15:56:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 02:28:24 2024
Response via : Initial Calibration

