




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032621\
 Data File : BF123475.D
 Acq On : 26 Mar 2021 9:16
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 26 12:47:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 17:06:59 2021
 Response via : Initial Calibration



TIC: BF123475.D\data.ms

(77) Benzidine

12.774min (+ 0.006) 0.00 ng

response 1664557

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.72
183.00	12.10	11.82
0.00	0.00	0.00