

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032621\
 Data File : BM029239.D
 Acq On : 26 Mar 2021 18:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 26 19:11:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 12:33:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.527	0.527	0.0	103	0.00
3	Pyridine	1.271	1.406	-10.6	103	0.00
4	n-Nitrosodimethylamine	0.571	0.657	-15.1	105	0.00
5 S	2-Fluorophenol	1.160	1.198	-3.3	103	0.00
6	Aniline	1.804	1.891	-4.8	103	0.00
7 S	Phenol-d6	1.665	1.736	-4.3	102	0.00
8	2-Chlorophenol	1.309	1.349	-3.1	104	0.00
9	Benzaldehyde	1.046	1.044	0.2	101	0.00
10 C	Phenol	1.645	1.715	-4.3	103	0.00
11	bis(2-Chloroethyl)ether	1.177	1.219	-3.6	101	0.00
12	1,3-Dichlorobenzene	1.465	1.459	0.4	101	0.00
13 C	1,4-Dichlorobenzene	1.485	1.490	-0.3	101	0.00
14	1,2-Dichlorobenzene	1.427	1.435	-0.6	102	0.00
15	Benzyl Alcohol	1.205	1.274	-5.7	101	0.00
16	2,2'-oxybis(1-Chloropropane	1.805	1.897	-5.1	104	0.00
17	2-Methylphenol	1.057	1.100	-4.1	101	0.00
18	Hexachloroethane	0.540	0.551	-2.0	102	0.00
19 P	n-Nitroso-di-n-propylamine	1.074	1.139	-6.1	99	0.00
20	3+4-Methylphenols	1.415	1.495	-5.7	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.501	0.509	-1.6	101	0.00
23 S	Nitrobenzene-d5	0.412	0.429	-4.1	101	0.00
24	Nitrobenzene	0.430	0.443	-3.0	101	0.00
25	Isophorone	0.708	0.739	-4.4	99	0.00
26 C	2-Nitrophenol	0.143	0.155	-8.4	97	0.00
27	2,4-Dimethylphenol	0.275	0.277	-0.7	99	0.00
28	bis(2-Chloroethoxy)methane	0.374	0.385	-2.9	100	0.00
29 C	2,4-Dichlorophenol	0.299	0.307	-2.7	98	0.00
30	1,2,4-Trichlorobenzene	0.339	0.333	1.8	99	0.00
31	Naphthalene	1.054	1.052	0.2	100	0.00
32	Benzoic acid	0.160	0.170	-6.3	95	0.00
33	4-Chloroaniline	0.417	0.430	-3.1	99	0.00
34 C	Hexachlorobutadiene	0.197	0.190	3.6	98	0.00
35	Caprolactam	0.091	0.097	-6.6	100	0.00
36 C	4-Chloro-3-methylphenol	0.319	0.335	-5.0	99	0.00
37	2-Methylnaphthalene	0.737	0.743	-0.8	99	0.00
38	1-Methylnaphthalene	0.701	0.708	-1.0	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	0.560	0.541	3.4	96	0.00
41 P	Hexachlorocyclopentadiene	0.308	0.296	3.9	94	0.00
42 S	2,4,6-Tribromophenol	0.185	0.189	-2.2	93	0.00
43 C	2,4,6-Trichlorophenol	0.365	0.375	-2.7	97	0.00
44	2,4,5-Trichlorophenol	0.406	0.417	-2.7	98	0.00
45 S	2-Fluorobiphenyl	1.410	1.399	0.8	98	0.00
46	1,1'-Biphenyl	1.520	1.520	0.0	98	0.00
47	2-Chloronaphthalene	1.197	1.191	0.5	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032621\
 Data File : BM029239.D
 Acq On : 26 Mar 2021 18:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 26 19:11:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 12:33:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.351	0.392	-11.7	101	0.00
49	Acenaphthylene	1.821	1.862	-2.3	98	0.00
50	Dimethylphthalate	1.398	1.414	-1.1	97	0.00
51	2,6-Dinitrotoluene	0.302	0.324	-7.3	99	0.00
52 C	Acenaphthene	1.157	1.162	-0.4	97	0.00
53	3-Nitroaniline	0.306	0.333	-8.8	99	0.00
54 P	2,4-Dinitrophenol	0.151	0.150	0.7	92	0.00
55	Dibenzofuran	1.821	1.832	-0.6	98	0.00
56 P	4-Nitrophenol	0.280	0.295	-5.4	99	-0.01
57	2,4-Dinitrotoluene	0.398	0.431	-8.3	97	0.00
58	Fluorene	1.476	1.500	-1.6	97	0.00
59	2,3,4,6-Tetrachlorophenol	0.332	0.337	-1.5	96	0.00
60	Diethylphthalate	1.377	1.417	-2.9	97	0.00
61	4-Chlorophenyl-phenylether	0.681	0.685	-0.6	97	0.00
62	4-Nitroaniline	0.322	0.358	-11.2	98	0.00
63	Azobenzene	1.574	1.694	-7.6	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	97	-0.01
65	4,6-Dinitro-2-methylphenol	0.111	0.114	-2.7	94	0.00
66 c	n-Nitrosodiphenylamine	0.640	0.651	-1.7	97	0.00
67	4-Bromophenyl-phenylether	0.183	0.174	4.9	91	0.00
68	Hexachlorobenzene	0.216	0.213	1.4	96	0.00
69	Atrazine	0.202	0.211	-4.5	94	-0.01
70 C	Pentachlorophenol	0.133	0.134	-0.8	93	0.00
71	Phenanthrene	1.160	1.174	-1.2	97	0.00
72	Anthracene	1.127	1.163	-3.2	98	0.00
73	Carbazole	1.109	1.157	-4.3	98	0.00
74	Di-n-butylphthalate	1.069	1.150	-7.6	96	0.00
75 C	Fluoranthene	1.281	1.329	-3.7	97	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	97	-0.01
77	Benzydine	0.578	0.545	5.7	90	0.00
78	Pyrene	1.354	1.351	0.2	97	0.00
79 S	Terphenyl-d14	0.996	0.996	0.0	96	-0.01
80	Butylbenzylphthalate	0.457	0.502	-9.8	97	0.00
81	Benzo(a)anthracene	1.305	1.333	-2.1	99	0.00
82	3,3'-Dichlorobenzidine	0.390	0.411	-5.4	97	0.00
83	Chrysene	1.306	1.300	0.5	97	-0.01
84	Bis(2-ethylhexyl)phthalate	0.658	0.757	-15.0	100	0.00
85 c	Di-n-octyl phthalate	1.066	1.232	-15.6	99	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	99	-0.01
87	Indeno(1,2,3-cd)pyrene	1.463	1.522	-4.0	100	-0.02
88	Benzo(b)fluoranthene	1.301	1.345	-3.4	99	-0.01
89	Benzo(k)fluoranthene	1.264	1.283	-1.5	100	-0.01
90 C	Benzo(a)pyrene	1.179	1.222	-3.6	100	-0.01
91	Dibenzo(a,h)anthracene	1.252	1.302	-4.0	101	-0.02
92	Benzo(g,h,i)perylene	1.227	1.260	-2.7	101	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032621\
Data File : BM029239.D
Acq On : 26 Mar 2021 18:13
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 26 19:11:41 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 12:33:48 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0