

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032621\
 Data File : BM029216.D
 Acq On : 25 Mar 2021 11:01
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

mohammad
 3/26/2021 1:31:16 PM

Quant Time: Mar 25 13:55:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 13:53:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.757	152	83950	20.00	ng	0.00
21) Naphthalene-d8	10.539	136	324259	20.00	ng	0.00
39) Acenaphthene-d10	14.386	164	193285	20.00	ng	0.00
64) Phenanthrene-d10	17.121	188	385562	20.00	ng	0.00
76) Chrysene-d12	21.292	240	371836	20.00	ng	0.00
86) Perylene-d12	23.533	264	383982	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.351	112	45512	9.48	ng	0.00
7) Phenol-d6	6.922	99	64386	9.29	ng	0.00
23) Nitrobenzene-d5	8.904	82	58176	8.80	ng	0.00
42) 2,4,6-Tribromophenol	15.868	330	13425	7.05	ng	0.00
45) 2-Fluorobiphenyl	13.010	172	129209	9.42	ng	0.00
79) Terphenyl-d14	19.745	244	169236	9.04	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.305	88	11234	5.19	ng	# 91
3) Pyridine	3.716	79	22635	4.15	ng	# 88
4) n-Nitrosodimethylamine	3.634	42	8496	3.43	ng	# 82
6) Aniline	7.087	93	35631	4.69	ng	96
8) 2-Chlorophenol	7.322	128	25764	4.71	ng	95
9) Benzaldehyde	6.904	77	23087	5.22	ng	97
10) Phenol	6.945	94	32273	4.70	ng	99
11) bis(2-Chloroethyl)ether	7.187	93	22560	4.48	ng	97
12) 1,3-Dichlorobenzene	7.651	146	30898	5.02	ng	95
13) 1,4-Dichlorobenzene	7.792	146	31083	5.02	ng	95
14) 1,2-Dichlorobenzene	8.110	146	29663	4.91	ng	94
15) Benzyl Alcohol	7.998	79	22459	4.31	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.287	45	38205	4.92	ng	99
17) 2-Methylphenol	8.192	107	20149	4.52	ng	96
18) Hexachloroethane	8.834	117	10913	4.79	ng	98
19) n-Nitroso-di-n-propyla...	8.557	70	20403	4.39	ng	98
20) 3+4-Methylphenols	8.516	107	25506	4.21	ng	99
22) Acetophenone	8.575	105	38591	4.73	ng	# 97
24) Nitrobenzene	8.951	77	31815	4.58	ng	96
25) Isophorone	9.475	82	48536	4.16	ng	99
26) 2-Nitrophenol	9.657	139	8947	4.12	ng	# 87
27) 2,4-Dimethylphenol	9.716	122	19745	4.42	ng	99
28) bis(2-Chloroethoxy)met...	9.957	93	28342	4.59	ng	95
29) 2,4-Dichlorophenol	10.186	162	20574	4.27	ng	93
30) 1,2,4-Trichlorobenzene	10.404	180	25939	4.71	ng	90
31) Naphthalene	10.592	128	81610	4.76	ng	97
33) 4-Chloroaniline	10.698	127	29332	4.28	ng	98
34) Hexachlorobutadiene	10.881	225	15318	4.72	ng	99
36) 4-Chloro-3-methylphenol	11.816	107	21248	4.05	ng	94
37) 2-Methylnaphthalene	12.204	142	55192	4.50	ng	97
38) 1-Methylnaphthalene	12.422	142	52699	4.54	ng	99
40) 1,2,4,5-Tetrachloroben...	12.575	216	26115	4.85	ng	98
41) Hexachlorocyclopentadiene	12.557	237	12998	4.34	ng	97
43) 2,4,6-Trichlorophenol	12.810	196	14493	4.17	ng	95
44) 2,4,5-Trichlorophenol	12.880	196	16372	4.20	ng	94
46) 1,1'-Biphenyl	13.222	154	69653	4.70	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032621\
 Data File : BM029216.D
 Acq On : 25 Mar 2021 11:01
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Manual Integrations
 APPROVED

mohammad
 3/26/2021 1:31:16 PM

Quant Time: Mar 25 13:55:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 13:53:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 2-Chloronaphthalene	13.257	162	54997	4.78	ng	98
48) 2-Nitroaniline	13.463	65	12006	3.58	ng	97
49) Acenaphthylene	14.104	152	77448	4.39	ng	96
50) Dimethylphthalate	13.845	163	62220	4.52	ng	99
51) 2,6-Dinitrotoluene	13.957	165	11041	3.73	ng	92
52) Acenaphthene	14.451	154	52961m	4.69	ng	
53) 3-Nitroaniline	14.286	138	10866	3.75	ng	# 86
55) Dibenzofuran	14.786	168	83029	4.63	ng	96
57) 2,4-Dinitrotoluene	14.745	165	12892	3.47	ng	91
58) Fluorene	15.433	166	64954	4.46	ng	96
59) 2,3,4,6-Tetrachlorophenol	15.004	232	13430	4.18	ng	# 97
60) Diethylphthalate	15.210	149	58467	4.27	ng	98
61) 4-Chlorophenyl-phenyle...	15.427	204	30012	4.43	ng	97
62) 4-Nitroaniline	15.439	138	10636	3.43	ng	94
63) Azobenzene	15.721	77	64301	4.15	ng	98
66) n-Nitrosodiphenylamine	15.639	169	55111	4.49	ng	98
67) 4-Bromophenyl-phenylether	16.321	248	15889	4.29	ng	90
68) Hexachlorobenzene	16.433	284	19388	4.63	ng	98
69) Atrazine	16.592	200	15622	3.96	ng	99
71) Phenanthrene	17.162	178	103716	4.63	ng	99
72) Anthracene	17.257	178	97165	4.49	ng	98
73) Carbazole	17.521	167	92681	4.34	ng	99
74) Di-n-butylphthalate	18.098	149	80477	3.87	ng	100
75) Fluoranthene	19.180	202	105374	4.24	ng	98
78) Pyrene	19.539	202	115459	4.67	ng	97
80) Butylbenzylphthalate	20.445	149	30223	3.77	ng	100
81) Benzo(a)anthracene	21.280	228	108549	4.48	ng	98
82) 3,3'-Dichlorobenzidine	21.215	252	28363	3.95	ng	98
83) Chrysene	21.327	228	113286	4.72	ng	98
84) Bis(2-ethylhexyl)phtha...	21.221	149	40771	3.40	ng	97
85) Di-n-octyl phthalate	22.097	149	68432	3.38	ng	# 98
87) Indeno(1,2,3-cd)pyrene	25.791	276	120556	4.27	ng	100
88) Benzo(b)fluoranthene	22.856	252	109877	4.38	ng	99
89) Benzo(k)fluoranthene	22.903	252	108677	4.47	ng	99
90) Benzo(a)pyrene	23.433	252	96965	4.30	ng	97
91) Dibenzo(a,h)anthracene	25.803	278	103814	4.26	ng	97
92) Benzo(g,h,i)perylene	26.485	276	104633	4.47	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.757 min Scan# 828
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion: 152 Resp: 83950
Ion Ratio Lower Upper
152 100
150 159.1 124.3 186.5
115 60.3 46.4 69.6

Instrument :
BNA_M
ClientSampled :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#2
1,4-Dioxane
Concen: 5.19 ng
RT: 3.305 min Scan# 71
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion: 88 Resp: 11234
Ion Ratio Lower Upper
88 100
58 69.4 59.9 89.9
43 40.4 26.2 39.4#





#3
Pyridine
Concen: 4.15 ng
RT: 3.716 min Scan# 141
Delta R.T. 0.024 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion: 79 Resp: 22635
Ion Ratio Lower Upper
79 100
52 62.1 53.2 79.8
51 51.2 29.8 44.8#

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#4
n-Nitrosodimethylamine
Concen: 3.43 ng
RT: 3.634 min Scan# 127
Delta R.T. 0.024 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion: 42 Resp: 8496
Ion Ratio Lower Upper
42 100
74 87.5 80.6 121.0
44 35.6 5.4 8.0#









#5
2-Fluorophenol
Concen: 9.48 ng
RT: 5.351 min Scan# 419
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Tgt Ion:	112	Resp:	45512
Ion Ratio	Lower	Upper	
112	100		
64	67.0	51.6	77.4
63	35.3	29.0	43.4




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#6
Aniline
Concen: 4.69 ng
RT: 7.087 min Scan# 714
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	93	Resp:	35631
Ion Ratio	Lower	Upper	
93	100		
66	39.4	33.2	49.8
65	24.1	17.8	26.6









#7
Phenol-d6
Concen: 9.29 ng
RT: 6.922 min Scan# 686
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	99	Resp:	64386
Ion	Ratio	Lower	Upper
99	100		
42	24.8	18.5	27.7
71	39.1	31.8	47.6

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#8
2-Chlorophenol
Concen: 4.71 ng
RT: 7.322 min Scan# 754
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion:	128	Resp:	25764
Ion	Ratio	Lower	Upper
128	100		
130	30.1	13.1	53.1
64	53.9	30.9	70.9





#9
Benzaldehyde
Concen: 5.22 ng
RT: 6.904 min Scan# 683
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:	77	Resp:	23087
Ion	Ratio	Lower	Upper
77	100		
105	88.7	71.4	111.4
106	84.0	66.5	106.5

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#10
Phenol
Concen: 4.70 ng
RT: 6.945 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion:	94	Resp:	32273
Ion	Ratio	Lower	Upper
94	100		
65	33.4	13.0	53.0
66	44.6	24.1	64.1









#11
bis(2-Chloroethyl)ether
Concen: 4.48 ng
RT: 7.187 min Scan# 731
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	93	Resp:	22560
Ion	Ratio	Lower	Upper
93	100		
63	80.9	63.9	103.9
95	31.4	12.8	52.8

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#12
1,3-Dichlorobenzene
Concen: 5.02 ng
RT: 7.651 min Scan# 810
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	146	Resp:	30898
Ion	Ratio	Lower	Upper
146	100		
148	59.0	50.5	75.7
75	35.1	29.7	44.5









#13
1,4-Dichlorobenzene
Concen: 5.02 ng
RT: 7.792 min Scan# 834
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Tgt Ion	Ratio	Lower	Upper
146	100		
148	59.5	51.8	77.6
111	42.6	32.8	49.2




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#14
1,2-Dichlorobenzene
Concen: 4.91 ng
RT: 8.110 min Scan# 888
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion	Ratio	Lower	Upper
146	100		
148	58.6	50.2	75.4
111	40.0	35.8	53.8





#15
Benzyl Alcohol
Concen: 4.31 ng
RT: 7.998 min Scan# 869
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion	Ratio	Lower	Upper
79	100		
108	74.8	56.6	84.8
77	73.7	54.9	82.3

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 4.92 ng
RT: 8.287 min Scan# 918
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion	Ratio	Lower	Upper
45	100		
77	13.8	0.0	34.5
79	10.8	0.0	30.9









#17
2-Methylphenol
Concen: 4.52 ng
RT: 8.192 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Tgt Ion:	107	Resp:	20149
Ion	Ratio	Lower	Upper
107	100		
108	105.9	89.8	134.6
77	52.4	42.8	64.2
79	51.1	40.4	60.6




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM




#18
Hexachloroethane
Concen: 4.79 ng
RT: 8.834 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	117	Resp:	10913
Ion	Ratio	Lower	Upper
117	100		
119	95.4	75.0	112.6
201	91.7	74.6	111.8









#19
n-Nitroso-di-n-propylamine
Concen: 4.39 ng
RT: 8.557 min Scan# 964
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01





Tgt Ion:	70	Resp:	20403
Ion	Ratio	Lower	Upper
70	100		
42	67.3	53.3	79.9
101	9.6	6.7	10.1
130	18.5	16.1	24.1



#20
3+4-Methylphenols
Concen: 4.21 ng
RT: 8.516 min Scan# 957
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:	107	Resp:	25506
Ion	Ratio	Lower	Upper
107	100		
108	89.0	68.5	108.5
77	37.5	17.0	57.0
79	31.2	12.0	52.0



Instrument :
BNA_M
ClientSampleId :
SSTDIC005


Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.539 min Scan# 1301
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:	136	Resp:	324259
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.3	13.9
54	8.9	6.9	10.3
68	6.2	5.4	8.0

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#22
Acetophenone
Concen: 4.73 ng
RT: 8.575 min Scan# 967
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion:	105	Resp:	38591
Ion Ratio	Lower	Upper	
105	100		
71	2.7	2.8	4.2#
51	31.7	24.7	37.1
120	19.2	17.4	26.2





#23

Nitrobenzene-d5

Concen: 8.80 ng

RT: 8.904 min Scan# 1023

Delta R.T. -0.006 min

Lab File: BM029216.D

Acq: 25 Mar 2021 11:01

Tgt Ion: 82	Resp: 58176
Ion Ratio	Lower Upper
82 100	
128 33.9	28.6 42.8
54 51.2	40.7 61.1

Instrument :





BNA_M

ClientSampled :

SSTDIC005

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM



#24

Nitrobenzene

Concen: 4.58 ng




RT: 8.951 min Scan# 1031


Delta R.T. 0.000 min

Lab File: BM029216.D

Acq: 25 Mar 2021 11:01

Tgt Ion: 77	Resp: 31815
Ion Ratio	Lower Upper
77 100	
123 35.8	31.3 46.9
65 13.0	10.8 16.2









#25
Isophorone
Concen: 4.16 ng
RT: 9.475 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
82	100		
95	7.7	5.6	8.4
138	16.3	13.3	19.9

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM




#26
2-Nitrophenol
Concen: 4.12 ng
RT: 9.657 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
139	100		
109	35.1	26.6	39.8
65	70.6	45.6	68.4









#27
2,4-Dimethylphenol
Concen: 4.42 ng
RT: 9.716 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01





Tgt Ion:122 Resp: 19745
Ion Ratio Lower Upper
122 100
107 116.6 95.0 142.6
121 57.2 45.8 68.6



#28
bis(2-Chloroethoxy)methane
Concen: 4.59 ng
RT: 9.957 min Scan# 1202
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion: 93 Resp: 28342
Ion Ratio Lower Upper
93 100
95 30.0 26.6 39.8
123 11.5 8.6 13.0



Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM







#29
2,4-Dichlorophenol
Concen: 4.27 ng
RT: 10.186 min Scan# 1241
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
162	100		
164	57.3	45.7	85.7
98	35.3	16.3	56.3

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#30
1,2,4-Trichlorobenzene
Concen: 4.71 ng
RT: 10.404 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion	Ratio	Lower	Upper
180	100		
182	107.2	77.0	115.4
145	35.7	25.4	38.2





#31
Naphthalene
Concen: 4.76 ng
RT: 10.592 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion	Ratio	Lower	Upper
128	100		
129	9.7	9.0	13.6
127	14.0	10.7	16.1

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#33
4-Chloroaniline
Concen: 4.28 ng
RT: 10.698 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion	Ratio	Lower	Upper
127	100		
129	30.7	25.4	38.0
65	36.3	28.6	42.8
92	17.1	15.8	23.6









#34
Hexachlorobutadiene
Concen: 4.72 ng
RT: 10.881 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01





Tgt Ion: 225 Resp: 15318
Ion Ratio Lower Upper
225 100
223 63.4 51.8 77.8
227 65.8 52.2 78.4



#36
4-Chloro-3-methylphenol
Concen: 4.05 ng
RT: 11.816 min Scan# 1518
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion: 107 Resp: 21248
Ion Ratio Lower Upper
107 100
144 26.6 22.7 34.1
142 77.0 67.0 100.4



Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM







#37
2-Methylnaphthalene
Concen: 4.50 ng
RT: 12.204 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	142	Resp:	55192
Ion Ratio	Lower	Upper	
142	100		
141	92.9	72.0	108.0
115	38.6	29.8	44.6

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM




#38
1-Methylnaphthalene
Concen: 4.54 ng
RT: 12.422 min Scan# 1621
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	142	Resp:	52699
Ion Ratio	Lower	Upper	
142	100		
141	93.0	73.5	110.3
116	4.2	3.2	4.8





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.386 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01






Tgt Ion:164 Resp: 193285
Ion Ratio Lower Upper
164 100
162 101.5 84.4 126.6
160 45.8 36.6 54.8


Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.85 ng
RT: 12.575 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:216 Resp: 26115
Ion Ratio Lower Upper
216 100
214 76.3 62.3 93.5
179 23.5 18.6 27.8
108 21.2 18.2 27.4







#41
Hexachlorocyclopentadiene
Concen: 4.34 ng
RT: 12.557 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion: 237 Resp: 12998
Ion Ratio Lower Upper
237 100
235 63.8 41.8 81.8
272 13.7 0.0 31.1




Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#42
2,4,6-Tribromophenol
Concen: 7.05 ng
RT: 15.868 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion: 330 Resp: 13425
Ion Ratio Lower Upper
330 100
332 95.0 77.1 115.7
141 46.6 39.2 58.8









#43
2,4,6-Trichlorophenol
Concen: 4.17 ng
RT: 12.810 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:196	Resp:	14493
Ion Ratio	Lower	Upper
196	100	
198	89.7	75.3 112.9
200	28.4	25.2 37.8

Instrument :
BNA_M
ClientSampled :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#44
2,4,5-Trichlorophenol
Concen: 4.20 ng
RT: 12.880 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion:196	Resp:	16372
Ion Ratio	Lower	Upper
196	100	
198	85.8	74.6 112.0
97	56.9	46.4 69.6
132	23.5	21.5 32.3





#45
2-Fluorobiphenyl
Concen: 9.42 ng
RT: 13.010 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:172 Resp: 129209
Ion Ratio Lower Upper
172 100
171 36.3 28.7 43.1
170 23.2 18.8 28.2

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#46
1,1'-Biphenyl
Concen: 4.70 ng
RT: 13.222 min Scan# 1757
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:154 Resp: 69653
Ion Ratio Lower Upper
154 100
153 43.4 21.5 61.5
76 13.8 0.0 33.6





#47
2-Chloronaphthalene
Concen: 4.78 ng
RT: 13.257 min Scan# 1763
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion	Ratio	Lower	Upper
162	100		
127	40.1	30.4	45.6
164	32.9	26.3	39.5

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#48
2-Nitroaniline
Concen: 3.58 ng
RT: 13.463 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion	Ratio	Lower	Upper
65	100		
92	64.9	52.4	78.6
138	89.5	75.0	112.6







#49
Acenaphthylene
Concen: 4.39 ng
RT: 14.104 min Scan# 1907
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:152 Resp: 77448
Ion Ratio Lower Upper
152 100
151 18.5 16.2 24.4
153 13.9 10.2 15.2




Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#50
Dimethylphthalate
Concen: 4.52 ng
RT: 13.845 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion:163 Resp: 62220
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.1 7.9 11.9









#51
2,6-Dinitrotoluene
Concen: 3.73 ng
RT: 13.957 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Tgt Ion:165 Resp: 11041
Ion Ratio Lower Upper
165 100
63 70.5 51.4 77.2
89 70.6 51.4 77.2




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#52
Acenaphthene
Concen: 4.69 ng m
RT: 14.451 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion:154 Resp: 52961
Ion Ratio Lower Upper
154 100
153 118.9 90.5 135.7
152 59.7 43.4 65.2





#53
3-Nitroaniline
Concen: 3.75 ng
RT: 14.286 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:138 Resp: 10866
Ion Ratio Lower Upper
138 100
108 14.7 9.0 13.4#
92 143.3 101.4 152.0

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#55
Dibenzofuran
Concen: 4.63 ng
RT: 14.786 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:168 Resp: 83029
Ion Ratio Lower Upper
168 100
139 41.3 30.6 46.0
169 11.9 10.2 15.4





#57
2,4-Dinitrotoluene
Concen: 3.47 ng
RT: 14.745 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:165 Resp: 12892
Ion Ratio Lower Upper
165 100
63 58.7 42.2 63.4
89 90.7 65.6 98.4
182 3.2 2.2 3.4

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#58
Fluorene
Concen: 4.46 ng
RT: 15.433 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:166 Resp: 64954
Ion Ratio Lower Upper
166 100
165 95.1 79.6 119.4
167 13.4 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 4.18 ng
RT: 15.004 min Scan# 2060
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01





Tgt Ion:232 Resp: 13430
Ion Ratio Lower Upper
232 100
131 47.9 41.1 61.7
130 1.5 2.6 3.8#
166 35.0 27.5 41.3



#60
Diethylphthalate
Concen: 4.27 ng
RT: 15.210 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:149 Resp: 58467
Ion Ratio Lower Upper
149 100
177 22.6 17.3 25.9
150 11.8 9.8 14.6



Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM







#61
4-Chlorophenyl-phenylether
Concen: 4.43 ng
RT: 15.427 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
204	100		
206	30.7	26.7	40.1
141	62.4	48.4	72.6

Instrument :
BNA_M
ClientSampled :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#62
4-Nitroaniline
Concen: 3.43 ng
RT: 15.439 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
138	100		
92	61.2	38.2	78.2
108	84.9	71.7	111.7









#63
Azobenzene
Concen: 4.15 ng
RT: 15.721 min Scan# 2182
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
77	100		
182	20.6	2.3	42.3
105	17.4	0.0	36.4
51	33.1	12.9	52.9

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.121 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.1	12.1
80	11.5	9.2	13.8









#66
n-Nitrosodiphenylamine
Concen: 4.49 ng
RT: 15.639 min Scan# 2168
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01





Tgt Ion:169 Resp: 55111
Ion Ratio Lower Upper
169 100
168 70.0 54.6 81.8
167 37.1 29.4 44.0



#67
4-Bromophenyl-phenylether
Concen: 4.29 ng
RT: 16.321 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:248 Resp: 15889
Ion Ratio Lower Upper
248 100
250 89.8 78.4 117.6
141 73.0 67.1 100.7



Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM







#68
Hexachlorobenzene
Concen: 4.63 ng
RT: 16.433 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:284 Resp: 19388
Ion Ratio Lower Upper
284 100
142 43.7 33.4 50.2
249 32.7 25.9 38.9

Instrument :
BNA_M
ClientSampled :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#69
Atrazine
Concen: 3.96 ng
RT: 16.592 min Scan# 2330
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:200 Resp: 15622
Ion Ratio Lower Upper
200 100
173 26.8 6.1 46.1
215 44.3 23.8 63.8









#71
Phenanthrene
Concen: 4.63 ng
RT: 17.162 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:178 Resp: 103716
Ion Ratio Lower Upper
178 100
176 20.0 15.7 23.5
179 15.0 11.9 17.9

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM




#72
Anthracene
Concen: 4.49 ng
RT: 17.257 min Scan# 2443
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:178 Resp: 97165
Ion Ratio Lower Upper
178 100
176 18.0 14.9 22.3
179 14.1 12.1 18.1









#73
 Carbazole
 Concen: 4.34 ng
 RT: 17.521 min Scan# 2488
 Delta R.T. -0.006 min
 Lab File: BM029216.D
 Acq: 25 Mar 2021 11:01





Tgt Ion:167 Resp: 92681
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.5 26.3
 139 12.7 10.6 15.8



#74
 Di-n-butylphthalate
 Concen: 3.87 ng
 RT: 18.098 min Scan# 2586
 Delta R.T. 0.000 min
 Lab File: BM029216.D
 Acq: 25 Mar 2021 11:01




Tgt Ion:149 Resp: 80477
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 5.8 4.6 6.8




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED




mohammad
 3/26/2021 1:31:16 PM




#75
Fluoranthene
Concen: 4.24 ng
RT: 19.180 min Scan# 2770
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01





Tgt Ion:202 Resp: 105374
Ion Ratio Lower Upper
202 100
101 10.3 0.0 31.5
203 17.3 0.0 37.7



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.292 min Scan# 3129
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:240 Resp: 371836
Ion Ratio Lower Upper
240 100
120 11.9 9.0 13.6
236 26.6 20.3 30.5



Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM







#78
Pyrene
Concen: 4.67 ng
RT: 19.539 min Scan# 2831
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	202	Resp:	115459
Ion Ratio	Lower	Upper	
202	100		
200	20.0	17.0	25.6
203	16.4	14.1	21.1

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#79
Terphenyl-d14
Concen: 9.04 ng
RT: 19.745 min Scan# 2866
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:	244	Resp:	169236
Ion Ratio	Lower	Upper	
244	100		
212	7.9	6.5	9.7
122	12.2	9.2	13.8









#80
Butylbenzylphthalate
Concen: 3.77 ng
RT: 20.445 min Scan# 2985
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
149	100		
91	75.7	60.5	90.7
206	20.2	15.7	23.5

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#81
Benzo(a)anthracene
Concen: 4.48 ng
RT: 21.280 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.9	32.9
229	18.6	15.7	23.5









#82
 3,3'-Dichlorobenzidine
 Concen: 3.95 ng
 RT: 21.215 min Scan# 3116
 Delta R.T. 0.000 min
 Lab File: BM029216.D
 Acq: 25 Mar 2021 11:01

Tgt Ion	Ratio	Lower	Upper
252	100		
254	67.4	52.6	78.8
126	12.5	10.8	16.2

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005




Manual Integrations
 APPROVED


mohammad
 3/26/2021 1:31:16 PM



#83
 Chrysene
 Concen: 4.72 ng
 RT: 21.327 min Scan# 3135
 Delta R.T. -0.006 min
 Lab File: BM029216.D
 Acq: 25 Mar 2021 11:01


Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.6	24.7	37.1
229	20.0	15.8	23.6








#84
Bis(2-ethylhexyl)phthalate
Concen: 3.40 ng
RT: 21.221 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Instrument :
BNA_M
ClientSampleId :
SSTDIC005





Tgt Ion:149 Resp: 40771
Ion Ratio Lower Upper
149 100
167 24.2 20.6 31.0
279 3.4 3.3 4.9




#85
Di-n-octyl phthalate
Concen: 3.38 ng
RT: 22.097 min Scan# 3266
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:149 Resp: 68432
Ion Ratio Lower Upper
149 100
167 2.0 1.3 1.9#
43 13.1 9.8 14.6





#86
Perylene-d12
Concen: 20.00 ng
RT: 23.533 min Scan# 3510
Delta R.T. 0.000 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:264 Resp: 383982
Ion Ratio Lower Upper
264 100
260 23.1 19.6 29.4
265 21.6 17.8 26.8

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#87
Indeno(1,2,3-cd)pyrene
Concen: 4.27 ng
RT: 25.791 min Scan# 3894
Delta R.T. -0.012 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01



Tgt Ion:276 Resp: 120556
Ion Ratio Lower Upper
276 100
138 28.0 22.6 34.0
277 25.5 20.2 30.4









#88
Benzo(b)fluoranthene
Concen: 4.38 ng
RT: 22.856 min Scan# 3395
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Tgt Ion:252 Resp: 109877
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 11.8 9.0 13.4

Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Manual Integrations
APPROVED


mohammad
3/26/2021 1:31:16 PM



#89
Benzo(k)fluoranthene
Concen: 4.47 ng
RT: 22.903 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01


Tgt Ion:252 Resp: 108677
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 11.1 8.7 13.1





#90
Benzo(a)pyrene
Concen: 4.30 ng
RT: 23.433 min Scan# 3493
Delta R.T. -0.006 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC005




Tgt Ion:252 Resp: 96965
Ion Ratio Lower Upper
252 100
253 22.5 17.0 25.4
125 13.2 9.5 14.3

Manual Integrations
APPROVED



mohammad
3/26/2021 1:31:16 PM




#91
Dibenzo(a,h)anthracene
Concen: 4.26 ng
RT: 25.803 min Scan# 3896
Delta R.T. -0.012 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01




Tgt Ion:278 Resp: 103814
Ion Ratio Lower Upper
278 100
139 16.5 14.8 22.2
279 23.0 19.3 28.9





#92
Benzo(g,h,i)perylene
Concen: 4.47 ng
RT: 26.485 min Scan# 4012
Delta R.T. -0.012 min
Lab File: BM029216.D
Acq: 25 Mar 2021 11:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC005



Tgt Ion: 276 Resp: 104633
Ion Ratio Lower Upper
276 100
277 25.1 18.7 28.1
138 22.7 19.0 28.4

Manual Integrations
APPROVED

mohammad
3/26/2021 1:31:16 PM

