

Data Path : Z:\HPCHEM1\BNA_M\Data\BM032717\
 Data File : BM009408.D
 Acq On : 27 Mar 2017 11:34
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.465

Quant Time: Mar 27 13:26:59 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 13:27:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
3	Naphthalene	1.153	1.202	-4.2	103	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.606	8.7	91	0.00
5	2-Methylnaphthalene	0.829	0.807	2.7	98	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
7	Acenaphthylene	2.082	2.143	-2.9	96	0.00
8 C	Acenaphthene	1.598	1.601	-0.2	94	0.00
9	Fluorene	1.869	1.798	3.8	85	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
11	Pentachlorophenol	0.194	0.193	0.5	77	0.00
12	Phenanthrene	1.256	1.249	0.6	83	0.00
13	Anthracene	1.241	1.262	-1.7	83	0.00
14 SURR	Fluoranthene-d10	1.251	1.146	8.4	75	0.00
15 C	Fluoranthene	1.638	1.517	7.4	75	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	65	0.00
17	Pyrene	1.376	1.540	-11.9	77	0.00
18	Benzo(a)anthracene	1.329	1.336	-0.5	65	0.00
19	Chrysene	1.261	1.353	-7.3	71	0.00
20 I	Perylene-d12	1.000	1.000	0.0	68	0.00
21	Benzo(b)fluoranthene	1.571	1.425	9.3	64	0.00
22	Benzo(k)fluoranthene	1.389	1.427	-2.7	74	0.00
23 C	Benzo(a)pyrene	1.414	1.394	1.4	68	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.635	9.1	60	0.00
25	Dibenzo(a,h)anthracene	1.510	1.383	8.4	60	0.00
26	Benzo(g,h,i)perylene	1.534	1.430	6.8	61	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0