

Data Path : Z:\HPCHEM1\BNA_M\Data\BM032717\
 Data File : BM009414.D
 Acq On : 27 Mar 2017 15:20
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.466

Quant Time: Mar 27 15:51:47 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 14:15:51 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.02
2 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
3	Naphthalene	1.153	1.128	2.2	72	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.593	10.7	66	0.00
5	2-Methylnaphthalene	0.829	0.786	5.2	71	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	67	-0.02
7	Acenaphthylene	2.082	2.103	-1.0	70	0.00
8 C	Acenaphthene	1.598	1.532	4.1	67	0.00
9	Fluorene	1.869	1.802	3.6	64	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
11	Pentachlorophenol	0.194	0.147	24.2#	43	0.00
12	Phenanthrene	1.256	1.211	3.6	59	0.00
13	Anthracene	1.241	1.256	-1.2	61	0.00
14 SURR	Fluoranthene-d10	1.251	1.184	5.4	57	0.00
15 C	Fluoranthene	1.638	1.564	4.5	57	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	50	0.00
17	Pyrene	1.376	1.549	-12.6	59	0.00
18	Benzo(a)anthracene	1.329	1.365	-2.7	50	0.00
19	Chrysene	1.261	1.347	-6.8	54	0.00
20 I	Perylene-d12	1.000	1.000	0.0	54	0.00
21	Benzo(b)fluoranthene	1.571	1.543	1.8	55	0.00
22	Benzo(k)fluoranthene	1.389	1.398	-0.6	57	0.00
23 C	Benzo(a)pyrene	1.414	1.391	1.6	54	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.753	2.5	51	0.00
25	Dibenzo(a,h)anthracene	1.510	1.421	5.9	48	0.00
26	Benzo(g,h,i)perylene	1.534	1.503	2.0	51	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0