

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM032816\
 Data File : BM004801.D
 Acq On : 28 Mar 2016 11:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.445

Quant Time: Mar 29 02:37:23 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 26 05:39:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2 SURR	1,4-Dioxane-d8	2.253	2.308	-2.4	116	0.00
3	1,4-Dioxane	2.745	2.892	-5.4	119	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
5	Naphthalene	12.077	12.022	0.5	113	0.00
6 SURR	2-Methylnaphthalene-d10	6.972	6.502	6.7	106	0.00
7	2-Methylnaphthalene	8.652	8.597	0.6	111	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
9	Acenaphthylene	23.000	23.077	-0.3	103	0.00
10 C	Acenaphthene	16.271	17.361	-6.7	109	0.00
11	Fluorene	19.428	19.941	-2.6	105	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
13	Pentachlorophenol	1.253	1.031	17.7	87	0.00
14	Phenanthrene	14.049	14.743	-4.9	110	0.00
15	Anthracene	13.267	13.105	1.2	104	0.00
16 SURR	Fluoranthene-d10	11.819	12.234	-3.5	108	0.00
17 C	Fluoranthene	15.993	17.337	-8.4	114	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
19	Pyrene	17.122	16.527	3.5	114	0.00
20	Benzo(a)anthracene	15.370	15.439	-0.4	117	0.00
21	Chrysene	14.847	15.601	-5.1	121	0.00
22 I	Perylene-d12	1.000	1.000	0.0	118	0.00
23	Benzo(b)fluoranthene	16.844	18.117	-7.6	130	0.00
24	Benzo(k)fluoranthene	15.612	15.476	0.9	114	0.00
25 C	Benzo(a)pyrene	15.401	15.788	-2.5	120	0.00
26	Indeno(1,2,3-cd)pyrene	14.255	14.949	-4.9	124	0.00
27	Dibenzo(a,h)anthracene	11.221	11.767	-4.9	124	0.00
28	Benzo(g,h,i)perylene	11.798	12.552	-6.4	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0