

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032922\
 Data File : BM034404.D
 Acq On : 29 Mar 2022 13:16
 Operator : CG/JU
 Sample : N2079-19
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EW607

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/01/2022
 Supervised By :Yogesh Patel 04/02/2022

Quant Time: Mar 30 02:29:46 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM032822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 28 18:27:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.933	152	2422	0.400	ng/ul	0.00
4) Naphthalene-d8	10.741	136	6552	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.574	164	3807	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.312	188	6922m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.498	240	5262	0.400	ng/ul	# 0.00
23) Perylene-d12	23.895	264	23208m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	16375	4.545	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.335	152	3629	0.411	ng/ul	0.00
18) Fluoranthene-d10	19.338	212	7780	0.474	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.343	88	1602	0.403	ng/ul#	77
5) Naphthalene	10.790	128	44882	2.366	ng/ul#	88
7) 2-Methylnaphthalene	12.412	142	4397	0.379	ng/ul	99
8) 1-Methylnaphthalene	12.627	142	3443	0.297	ng/ul	99
14) Pentachlorophenol	16.983	266	269	0.096	ng/ul	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032922\
Data File : BM034404.D
Acq On : 29 Mar 2022 13:16
Operator : CG/JU
Sample : N2079-19
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EW607

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/01/2022
Supervised By :Yogesh Patel 04/02/2022

Quant Time: Mar 30 02:29:46 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM032822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 28 18:27:40 2022
Response via : Initial Calibration

