

Data Path : Z:\HPCHEM1\BNA M\DATA\BM033017\
 Data File : BM009458.D
 Acq On : 30 Mar 2017 19:33
 Operator : SJ/MA
 Sample : I2446-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 032817-PRRT-F-U-M

Quant Time: Mar 31 08:51:00 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 16:43:02 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	106891	20.00	ng	-0.01
21) Naphthalene-d8	10.65	136	390528	20.00	ng	-0.02
38) Acenaphthene-d10	14.49	164	217706	20.00	ng	-0.01
63) Phenanthrene-d10	17.24	188	534759	20.00	ng	-0.01
75) Chrysene-d12	21.41	240	749836	20.00	ng	-0.01
86) Perylene-d12	23.74	264	755943	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	278965	43.50	ng	-0.01
7) Phenol-d6	7.00	99	232618	26.60	ng	-0.02
23) Nitrobenzene-d5	9.02	82	1017752	97.71	ng	-0.01
41) 2,4,6-Tribromophenol	15.99	330	296943	109.79	ng	-0.01
44) 2-Fluorobiphenyl	13.11	172	1890272	104.62	ng	-0.01
78) Terphenyl-d14	19.86	244	2874018	80.09	ng	0.00
Target Compounds						
49) Dimethylphthalate	13.95	163	37226	2.23	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM033017\
Data File : BM009458.D
Acq On : 30 Mar 2017 19:33
Operator : SJ/MA
Sample : I2446-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
032817-PRRT-F-U-M

Quant Time: Mar 31 08:51:00 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 22 16:43:02 2017
Response via : Initial Calibration

