

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033021\
 Data File : BM029284.D
 Acq On : 30 Mar 2021 23:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 31 00:31:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 11:14:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	137	0.00
2	1,4-Dioxane	40.000	35.220	12.0	123	-0.01
3	Pyridine	40.000	39.375	1.6	125	-0.01
4	n-Nitrosodimethylamine	40.000	41.758	-4.4	130	-0.01
5 S	2-Fluorophenol	80.000	79.897	0.1	136	0.00
6	Aniline	40.000	40.583	-1.5	135	0.00
7 S	Phenol-d6	80.000	81.266	-1.6	135	0.00
8	2-Chlorophenol	40.000	40.963	-2.4	140	0.00
9	Benzaldehyde	40.000	39.724	0.7	136	0.00
10 C	Phenol	40.000	40.409	-1.0	135	0.00
11	bis(2-Chloroethyl)ether	40.000	42.214	-5.5	139	-0.01
12	1,3-Dichlorobenzene	40.000	39.395	1.5	136	0.00
13 C	1,4-Dichlorobenzene	40.000	39.859	0.4	137	0.00
14	1,2-Dichlorobenzene	40.000	39.955	0.1	137	0.00
15	Benzyl Alcohol	40.000	43.151	-7.9	140	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.559	-6.4	143	0.00
17	2-Methylphenol	40.000	42.288	-5.7	139	0.00
18	Hexachloroethane	40.000	43.857	-9.6	148	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.779	-9.4	139	0.00
20	3+4-Methylphenols	40.000	42.920	-7.3	140	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	137	0.00
22	Acetophenone	40.000	40.173	-0.4	137	0.00
23 S	Nitrobenzene-d5	80.000	83.086	-3.9	139	0.00
24	Nitrobenzene	40.000	40.461	-1.2	136	0.00
25	Isophorone	40.000	43.861	-9.7	143	0.00
26 C	2-Nitrophenol	40.000	50.402	-26.0#	155	0.00
27	2,4-Dimethylphenol	40.000	41.208	-3.0	138	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.057	-7.6	144	0.00
29 C	2,4-Dichlorophenol	40.000	40.480	-1.2	133	0.00
30	1,2,4-Trichlorobenzene	40.000	39.989	0.0	139	0.00
31	Naphthalene	40.000	39.992	0.0	138	0.00
32	Benzoic acid	40.000	48.630	-21.6	166	0.04
33	4-Chloroaniline	40.000	39.767	0.6	131	0.00
34 C	Hexachlorobutadiene	40.000	41.580	-3.9	145	0.00
35	Caprolactam	40.000	41.229	-3.1	140	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.821	-7.1	138	0.01
37	2-Methylnaphthalene	40.000	41.025	-2.6	137	0.00
38	1-Methylnaphthalene	40.000	40.937	-2.3	138	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	133	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.469	-1.2	138	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.549	1.1	132	0.00
42 S	2,4,6-Tribromophenol	80.000	83.452	-4.3	130	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.254	-10.6	142	0.00
44	2,4,5-Trichlorophenol	40.000	38.257	4.4	125	0.00
45 S	2-Fluorobiphenyl	80.000	81.928	-2.4	139	0.00
46	1,1'-Biphenyl	40.000	40.628	-1.6	137	0.00
47	2-Chloronaphthalene	40.000	40.497	-1.2	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033021\
 Data File : BM029284.D
 Acq On : 30 Mar 2021 23:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 31 00:31:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 11:14:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.834	-4.6	142	0.00
49	Acenaphthylene	40.000	41.824	-4.6	137	0.00
50	Dimethylphthalate	40.000	41.453	-3.6	136	0.00
51	2,6-Dinitrotoluene	40.000	43.346	-8.4	138	0.00
52 C	Acenaphthene	40.000	40.035	-0.1	133	0.00
53	3-Nitroaniline	40.000	42.388	-6.0	133	0.00
54 P	2,4-Dinitrophenol	40.000	36.174	9.6	120	0.00
55	Dibenzofuran	40.000	39.631	0.9	132	0.00
56 P	4-Nitrophenol	40.000	39.042	2.4	125	0.02
57	2,4-Dinitrotoluene	40.000	40.203	-0.5	136	0.00
58	Fluorene	40.000	40.758	-1.9	134	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.134	9.7	118	0.00
60	Diethylphthalate	40.000	44.114	-10.3	143	0.00
61	4-Chlorophenyl-phenylether	40.000	40.936	-2.3	136	0.00
62	4-Nitroaniline	40.000	38.901	2.7	130	0.00
63	Azobenzene	40.000	44.230	-10.6	141	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	129	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.824	2.9	125	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.986	-5.0	133	0.00
67	4-Bromophenyl-phenylether	40.000	44.203	-10.5	141	0.00
68	Hexachlorobenzene	40.000	40.849	-2.1	131	0.00
69	Atrazine	40.000	44.910	-12.3	134	0.00
70 C	Pentachlorophenol	40.000	23.895	40.3#	74	0.00
71	Phenanthrene	40.000	39.625	0.9	126	0.00
72	Anthracene	40.000	41.102	-2.8	129	0.00
73	Carbazole	40.000	40.061	-0.2	125	0.00
74	Di-n-butylphthalate	40.000	51.339	-28.3#	152	0.00
75 C	Fluoranthene	40.000	39.895	0.3	124	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	113	0.00
77	Benzydine	40.000	37.263	6.8	104	0.00
78	Pyrene	40.000	42.121	-5.3	119	0.00
79 S	Terphenyl-d14	80.000	88.803	-11.0	124	0.00
80	Butylbenzylphthalate	40.000	52.589	-31.5#	153	0.00
81	Benzo(a)anthracene	40.000	40.992	-2.5	115	0.00
82	3,3'-Dichlorobenzidine	40.000	45.173	-12.9	121	0.00
83	Chrysene	40.000	39.577	1.1	112	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	55.182	-38.0#	159	0.00
85 c	Di-n-octyl phthalate	40.000	56.493	-41.2#	164	0.00
86 I	Perylene-d12	20.000	20.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.131	-0.3	111	0.00
88	Benzo(b)fluoranthene	40.000	40.916	-2.3	112	0.00
89	Benzo(k)fluoranthene	40.000	39.666	0.8	112	0.00
90 C	Benzo(a)pyrene	40.000	40.806	-2.0	113	0.00
91	Dibenzo(a,h)anthracene	40.000	40.145	-0.4	112	0.00
92	Benzo(g,h,i)perylene	40.000	38.967	2.6	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033021\
Data File : BM029284.D
Acq On : 30 Mar 2021 23:41
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 31 00:31:28 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 29 11:14:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 3