

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033023\
 Data File : BM039227.D
 Acq On : 30 Mar 2023 15:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 16:38:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 30 05:15:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	42.425	-6.1	105	0.00
3	Pyridine	40.000	42.834	-7.1	103	0.00
4	n-Nitrosodimethylamine	40.000	43.514	-8.8	108	0.00
5 S	2-Fluorophenol	80.000	81.960	-2.4	101	0.00
6	Aniline	40.000	40.774	-1.9	100	0.00
7 S	Phenol-d6	80.000	81.252	-1.6	100	0.00
8	2-Chlorophenol	40.000	40.327	-0.8	99	0.00
9	Benzaldehyde	40.000	40.591	-1.5	100	0.00
10 C	Phenol	40.000	40.881	-2.2	101	0.00
11	bis(2-Chloroethyl)ether	40.000	40.704	-1.8	101	0.00
12	1,3-Dichlorobenzene	40.000	39.662	0.8	98	0.00
13 C	1,4-Dichlorobenzene	40.000	39.925	0.2	99	0.00
14	1,2-Dichlorobenzene	40.000	40.082	-0.2	99	0.00
15	Benzyl Alcohol	40.000	40.319	-0.8	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.209	-3.0	103	0.00
17	2-Methylphenol	40.000	40.821	-2.1	101	0.00
18	Hexachloroethane	40.000	39.746	0.6	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.794	-2.0	101	0.00
20	3+4-Methylphenols	40.000	40.807	-2.0	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	41.476	-3.7	101	0.00
23 S	Nitrobenzene-d5	80.000	83.027	-3.8	101	0.00
24	Nitrobenzene	40.000	41.513	-3.8	101	0.00
25	Isophorone	40.000	40.990	-2.5	100	0.00
26 C	2-Nitrophenol	40.000	40.422	-1.1	97	0.00
27	2,4-Dimethylphenol	40.000	40.377	-0.9	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.584	-1.5	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.528	-1.3	97	0.00
30	1,2,4-Trichlorobenzene	40.000	39.815	0.5	98	0.00
31	Naphthalene	40.000	40.316	-0.8	99	0.00
32	Benzoic acid	40.000	39.311	1.7	92	-0.01
33	4-Chloroaniline	40.000	40.345	-0.9	98	0.00
34 C	Hexachlorobutadiene	40.000	39.932	0.2	98	0.00
35	Caprolactam	40.000	39.379	1.6	93	-0.01
36 C	4-Chloro-3-methylphenol	40.000	40.159	-0.4	96	0.00
37	2-Methylnaphthalene	40.000	39.413	1.5	96	0.00
38	1-Methylnaphthalene	40.000	39.559	1.1	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.576	1.1	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.580	-3.9	98	0.00
42 S	2,4,6-Tribromophenol	80.000	82.647	-3.3	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.940	0.2	96	0.00
44	2,4,5-Trichlorophenol	40.000	39.989	0.0	96	0.00
45 S	2-Fluorobiphenyl	80.000	79.756	0.3	99	0.00
46	1,1'-Biphenyl	40.000	40.057	-0.1	99	0.00
47	2-Chloronaphthalene	40.000	39.950	0.1	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033023\
 Data File : BM039227.D
 Acq On : 30 Mar 2023 15:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 16:38:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 30 05:15:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.673	-1.7	99	0.00
49	Acenaphthylene	40.000	39.676	0.8	98	0.00
50	Dimethylphthalate	40.000	39.919	0.2	99	0.00
51	2,6-Dinitrotoluene	40.000	40.407	-1.0	98	0.00
52 C	Acenaphthene	40.000	40.204	-0.5	100	0.00
53	3-Nitroaniline	40.000	40.120	-0.3	96	0.00
54 P	2,4-Dinitrophenol	40.000	40.260	-0.6	96	0.00
55	Dibenzofuran	40.000	39.957	0.1	100	0.00
56 P	4-Nitrophenol	40.000	40.757	-1.9	93	0.00
57	2,4-Dinitrotoluene	40.000	40.500	-1.3	96	0.00
58	Fluorene	40.000	40.514	-1.3	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.126	-0.3	98	0.00
60	Diethylphthalate	40.000	40.304	-0.8	100	0.00
61	4-Chlorophenyl-phenylether	40.000	40.746	-1.9	102	0.00
62	4-Nitroaniline	40.000	41.145	-2.9	97	0.00
63	Azobenzene	40.000	42.552	-6.4	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.922	-2.3	99	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.772	-1.9	101	0.00
67	4-Bromophenyl-phenylether	40.000	40.844	-2.1	101	0.00
68	Hexachlorobenzene	40.000	40.765	-1.9	102	0.00
69	Atrazine	40.000	40.673	-1.7	101	0.00
70 C	Pentachlorophenol	40.000	42.520	-6.3	100	0.00
71	Phenanthrene	40.000	40.590	-1.5	101	0.00
72	Anthracene	40.000	40.887	-2.2	100	0.00
73	Carbazole	40.000	40.460	-1.2	99	0.00
74	Di-n-butylphthalate	40.000	40.918	-2.3	100	0.00
75 C	Fluoranthene	40.000	40.154	-0.4	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzidine	40.000	40.369	-0.9	96	0.00
78	Pyrene	40.000	39.347	1.6	100	0.00
79 S	Terphenyl-d14	80.000	80.206	-0.3	102	0.00
80	Butylbenzylphthalate	40.000	38.705	3.2	97	0.00
81	Benzo(a)anthracene	40.000	40.332	-0.8	103	0.00
82	3,3'-Dichlorobenzidine	40.000	39.855	0.4	100	0.00
83	Chrysene	40.000	40.026	-0.1	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.025	-0.1	102	0.00
85 c	Di-n-octyl phthalate	40.000	38.756	3.1	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.871	-2.2	101	-0.02
88	Benzo(b)fluoranthene	40.000	40.588	-1.5	103	0.00
89	Benzo(k)fluoranthene	40.000	40.140	-0.4	101	-0.01
90 C	Benzo(a)pyrene	40.000	40.390	-1.0	101	0.00
91	Dibenzo(a,h)anthracene	40.000	41.074	-2.7	102	-0.02
92	Benzo(g,h,i)perylene	40.000	40.512	-1.3	100	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033023\
Data File : BM039227.D
Acq On : 30 Mar 2023 15:22
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 30 16:38:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 30 05:15:45 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0