

Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM033115\  
 Data File : BM000763.D  
 Acq On : 31 Mar 2015 21:15  
 Operator : TP/IZ  
 Sample : SSTDICCC001  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICCC001

Quant Time: Apr 01 03:06:06 2015  
 Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\SIMPAH-BM033115.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Apr 01 03:53:31 2015  
 Response via : Initial Calibration


| Internal Standards          | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.62  | 152  | 32402    | 5.00 | ng    | 0.00     |
| 4) Naphthalene-d8           | 10.40 | 136  | 113043   | 5.00 | ng    | 0.00     |
| 8) Acenaphthene-d10         | 14.28 | 164  | 54905    | 5.00 | ng    | 0.00     |
| 13) Phenanthrene-d10        | 17.01 | 188  | 112136   | 5.00 | ng    | 0.00     |
| 19) Chrysene-d12            | 21.21 | 240  | 95475    | 5.00 | ng    | 0.00     |
| 25) Perylene-d12            | 23.40 | 264  | 85149    | 5.00 | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |      |       |          |
| 5) Nitrobenzene-d5          | 8.77  | 82   | 5500     | 0.96 | ng    | 0.00     |
| 9) 2-Fluorobiphenyl         | 12.89 | 172  | 17579    | 1.33 | ng    | 0.00     |
| 21) Terphenyl-d14           | 19.66 | 244  | 11837    | 0.87 | ng    | 0.00     |
| Target Compounds            |       |      |          |      |       | Qvalue   |
| 3) n-Nitrosodimethylamine   | 3.45  | 42   | 3110     | 0.68 | ng    | 100      |
| 6) Naphthalene              | 10.45 | 128  | 26784    | 1.18 | ng    | 100      |
| 7) 2-Methylnaphthalene      | 12.07 | 142  | 16850    | 1.18 | ng    | 100      |
| 10) Acenaphthylene          | 13.98 | 152  | 24243    | 0.35 | ng    | 100      |
| 11) Acenaphthene            | 14.33 | 154  | 16650    | 1.37 | ng    | 100      |
| 12) Fluorene                | 15.32 | 166  | 19063    | 1.27 | ng    | 100      |
| 14) Hexachlorobenzene       | 16.33 | 284  | 6749     | 1.79 | ng    | # 100    |
| 15) Pentachlorophenol       | 16.68 | 266  | 1038     | 0.03 | ng    | 100      |
| 16) Phenanthrene            | 17.05 | 178  | 31288    | 1.28 | ng    | 100      |
| 17) Anthracene              | 17.15 | 178  | 26026    | 1.20 | ng    | 100      |
| 18) Fluoranthene            | 19.08 | 202  | 30363    | 1.30 | ng    | 100      |
| 20) Pyrene                  | 19.44 | 202  | 31542    | 1.29 | ng    | 100      |
| 22) Benzo(a)anthracene      | 21.20 | 228  | 26444    | 1.40 | ng    | 100      |
| 23) Chrysene                | 21.24 | 228  | 29325    | 1.47 | ng    | 100      |
| 24) Indeno(1,2,3-cd)pyrene  | 25.59 | 276  | 29453    | 2.07 | ng    | 100      |
| 26) Benzo(b)fluoranthene    | 22.74 | 252  | 27195    | 1.67 | ng    | 100      |
| 27) Benzo(k)fluoranthene    | 22.79 | 252  | 26455    | 1.38 | ng    | 100      |
| 28) Benzo(a)pyrene          | 23.31 | 252  | 23038    | 1.59 | ng    | 100      |
| 29) Dibenzo(a,h)anthracene  | 25.60 | 278  | 23176    | 2.13 | ng    | 100      |
| 30) Benzo(g,h,i)perylene    | 26.26 | 276  | 25038    | 1.80 | ng    | 100      |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM033115\  
Data File : BM000763.D  
Acq On : 31 Mar 2015 21:15  
Operator : TP/IZ  
Sample : SSTDICCC001  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDICCC001

Quant Time: Apr 01 03:06:06 2015  
Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\SIMPAH-BM033115.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Apr 01 03:53:31 2015  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.62 min Scan# 920

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:152 Resp: 32402

Ion Ratio Lower Upper

152 100

150 151.0 120.8 181.2




115 53.4 42.7 64.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine

Concen: 0.68 ng

RT: 3.45 min Scan# 143

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion: 42 Resp: 3110

Ion Ratio Lower Upper

42 100

74 117.0 93.6 140.4



44 7.9 6.3 9.5




Abundance Ion 42.00 (41.70 to 42.70): BM0

Ion 74.00 (73.70 to 74.70): BM0

Ion 44.00 (43.70 to 44.70): BM0





#4

Naphthalene-d8

Concen: 5.00 ng

RT: 10.40 min Scan# 1359

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion: 136 Resp: 113043

Ion Ratio Lower Upper

136 100

137 10.7 8.6 12.8




54 8.8 7.0 10.6



Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BM



#5

Nitrobenzene-d5

Concen: 0.96 ng

RT: 8.77 min Scan# 1135

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion: 82 Resp: 5500

Ion Ratio Lower Upper

82 100

128 40.2 32.2 48.2



54 62.8 50.2 75.4




Abundance Ion 82.00 (81.70 to 82.70): BM

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BM





#6

Naphthalene

Concen: 1.18 ng

RT: 10.45 min Scan# 1363

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:128 Resp: 26784

Ion Ratio Lower Upper

128 100

129 11.1 8.9 13.3




127 12.5 10.0 15.0



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#7

2-Methylnaphthalene

Concen: 1.18 ng

RT: 12.07 min Scan# 1509

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:142 Resp: 16850

Ion Ratio Lower Upper

142 100

141 90.4 72.3 108.5



115 36.9 29.5 44.3




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E





#8

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.28 min Scan# 1692

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:164 Resp: 54905

Ion Ratio Lower Upper

164 100

162 94.3 75.4 113.2




160 38.5 30.8 46.2



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#9

2-Fluorobiphenyl

Concen: 1.33 ng

RT: 12.89 min Scan# 1588

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:172 Resp: 17579

Ion Ratio Lower Upper

172 100

171 33.2 26.6 39.8



170 21.4 17.1 25.7




Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#10

Acenaphthylene

Concen: 0.35 ng

RT: 13.98 min Scan# 1675

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:152 Resp: 24243

Ion Ratio Lower Upper

152 100

151 18.9 15.1 22.7




153 12.8 10.2 15.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#11

Acenaphthene

Concen: 1.37 ng

RT: 14.33 min Scan# 1695

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:154 Resp: 16650

Ion Ratio Lower Upper

154 100

153 109.1 87.3 130.9



152 53.8 43.0 64.6




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#12

Fluorene

Concen: 1.27 ng

RT: 15.32 min Scan# 1753

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:166 Resp: 19063

Ion Ratio Lower Upper

166 100

165 97.7 78.2 117.2


167 13.4 10.7 16.1





Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



Time-->



#13

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.01 min Scan# 1841

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:188 Resp: 112136

Ion Ratio Lower Upper

188 100

94 15.0 12.0 18.0


80 16.0 12.8 19.2




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0


Ion 80.00 (79.70 to 80.70): BM0



Time-->











#14  
Hexachlorobenzene  
Concen: 1.79 ng  
RT: 16.33 min Scan# 1808  
Delta R.T. 0.00 min  
Lab File: BM000763.D  
Acq: 31 Mar 2015 21:15

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDICCC001

Tgt Ion: 284 Resp: 6749  
Ion Ratio Lower Upper  
284 100  
142 0.0 0.0 0.0  
249 31.5 25.2 37.8




Abundance Ion 283.80 (283.50 to 284.50): E  
Ion 142.00 (141.70 to 142.70): E  
Ion 249.00 (248.70 to 249.70): E






#15  
Pentachlorophenol  
Concen: 0.03 ng  
RT: 16.68 min Scan# 1825  
Delta R.T. 0.00 min  
Lab File: BM000763.D  
Acq: 31 Mar 2015 21:15

Tgt Ion: 266 Resp: 1038  
Ion Ratio Lower Upper  
266 100  
268 68.8 55.0 82.6  
264 63.7 51.0 76.4



Abundance Ion 265.70 (265.40 to 266.40): E  
Ion 268.00 (267.70 to 268.70): E  
Ion 264.00 (263.70 to 264.70): E





#16

Phenanthrene

Concen: 1.28 ng

RT: 17.05 min Scan# 1843

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:178 Resp: 31288

Ion Ratio Lower Upper

178 100

176 19.1 15.3 22.9




179 15.2 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#17

Anthracene

Concen: 1.20 ng

RT: 17.15 min Scan# 1848

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:178 Resp: 26026

Ion Ratio Lower Upper

178 100

176 18.3 14.6 22.0



179 15.3 12.2 18.4




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#18  
 Fluoranthene  
 Concen: 1.30 ng  
 RT: 19.08 min Scan# 1973  
 Delta R.T. 0.00 min  
 Lab File: BM000763.D  
 Acq: 31 Mar 2015 21:15




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC001


Tgt Ion: 202 Resp: 30363  
 Ion Ratio Lower Upper  
 202 100  
 101 13.1 10.5 15.7  
 203 17.3 13.8 20.8



#19  
 Chrysene-d12  
 Concen: 5.00 ng  
 RT: 21.21 min Scan# 2158  
 Delta R.T. 0.00 min  
 Lab File: BM000763.D  
 Acq: 31 Mar 2015 21:15

Tgt Ion: 240 Resp: 95475  
 Ion Ratio Lower Upper  
 240 100  
 120 18.6 14.9 22.3  
 236 30.3 24.2 36.4









#20  
Pyrene  
Concen: 1.29 ng  
RT: 19.44 min Scan# 2003  
Delta R.T. 0.00 min  
Lab File: BM000763.D  
Acq: 31 Mar 2015 21:15




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDICCC001


Tgt Ion:202 Resp: 31542  
Ion Ratio Lower Upper  
202 100  
200 20.2 16.2 24.2  
203 17.4 13.9 20.9



#21  
Terphenyl-d14  
Concen: 0.87 ng  
RT: 19.66 min Scan# 2021  
Delta R.T. 0.00 min  
Lab File: BM000763.D  
Acq: 31 Mar 2015 21:15

Tgt Ion:244 Resp: 11837  
Ion Ratio Lower Upper  
244 100  
212 9.7 7.8 11.6  
122 14.8 11.8 17.8





#22

Benzo(a)anthracene

Concen: 1.40 ng

RT: 21.20 min Scan# 2157

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:228 Resp: 26444

Ion Ratio Lower Upper

228 100

226 23.2 18.6 27.8




229 21.5 17.2 25.8



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#23

Chrysene

Concen: 1.47 ng

RT: 21.24 min Scan# 2161

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:228 Resp: 29325

Ion Ratio Lower Upper

228 100

226 32.3 25.8 38.8



229 17.2 13.8 20.6




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#24

Indeno(1,2,3-cd)pyrene

Concen: 2.07 ng

RT: 25.59 min Scan# 2578

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:276 Resp: 29453

Ion Ratio Lower Upper




276 100

138 34.3 27.4 41.2

277 24.6 19.7 29.5



Abundance Ion 276.00 (275.70 to 276.70): E  
Ion 138.00 (137.70 to 138.70): E  
Ion 277.00 (276.70 to 277.70): E



#25

Perylene-d12

Concen: 5.00 ng

RT: 23.40 min Scan# 2368

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:264 Resp: 85149

Ion Ratio Lower Upper



264 100


260 25.1 20.1 30.1

265 21.7 17.4 26.0



Abundance Ion 264.00 (263.70 to 264.70): E  
Ion 260.00 (259.70 to 260.70): E  
Ion 265.00 (264.70 to 265.70): E









#26  
Benzo(b)fluoranthene  
Concen: 1.67 ng  
RT: 22.74 min Scan# 2305  
Delta R.T. 0.00 min  
Lab File: BM000763.D  
Acq: 31 Mar 2015 21:15




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDICCC001


Tgt Ion:252 Resp: 27195  
Ion Ratio Lower Upper  
252 100  
253 23.1 18.5 27.7  
125 13.3 10.6 16.0



#27  
Benzo(k)fluoranthene  
Concen: 1.38 ng  
RT: 22.79 min Scan# 2309  
Delta R.T. 0.00 min  
Lab File: BM000763.D  
Acq: 31 Mar 2015 21:15

Tgt Ion:252 Resp: 26455  
Ion Ratio Lower Upper  
252 100  
253 22.4 17.9 26.9  
125 13.5 10.8 16.2





#28

Benzo(a)pyrene

Concen: 1.59 ng

RT: 23.31 min Scan# 2359

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001


Tgt Ion:252 Resp: 23038

Ion Ratio Lower Upper




252 100

253 24.3 19.4 29.2

125 14.4 11.5 17.3



Abundance Ion 252.00 (251.70 to 252.70): E  
Ion 253.00 (252.70 to 253.70): E  
Ion 125.00 (124.70 to 125.70): E



#29

Dibenzo(a,h)anthracene

Concen: 2.13 ng

RT: 25.60 min Scan# 2579

Delta R.T. 0.00 min

Lab File: BM000763.D

Acq: 31 Mar 2015 21:15


Tgt Ion:278 Resp: 23176

Ion Ratio Lower Upper



278 100

139 22.2 17.8 26.6


279 24.1 19.3 28.9



Abundance Ion 278.00 (277.70 to 278.70): E  
Ion 139.00 (138.70 to 139.70): E  
Ion 279.00 (278.70 to 279.70): E







#30

Benzo(g,h,i)perylene

Concen: 1.80 ng

RT: 26.26 min Scan# 2642

Delta R.T. 0.00 min

Lab File: BM000763.D


Acq: 31 Mar 2015 21:15

Instrument :

BNA\_M

ClientSampleId :

SSTDICCC001




Tgt Ion: 276 Resp: 25038

Ion Ratio Lower Upper

276 100

277 23.9 19.1 28.7

138 28.9 23.1 34.7



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

