

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM033115\
 Data File : BM000768.D
 Acq On : 01 Apr 2015 00:13
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc : MDL-SOIL-BLK
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM033115

Quant Time: Apr 01 15:08:09 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	36692	5.00	ng	0.00
4) Naphthalene-d8	10.40	136	126557	5.00	ng	0.00
8) Acenaphthene-d10	14.28	164	62160	5.00	ng	0.00
13) Phenanthrene-d10	17.01	188	132608	5.00	ng	0.00
19) Chrysene-d12	21.21	240	112995	5.00	ng	0.00
25) Perylene-d12	23.40	264	101840	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.77	82	6597	1.02	ng	0.00
9) 2-Fluorobiphenyl	12.89	172	20179	1.05	ng	0.00
21) Terphenyl-d14	19.66	244	14277	1.05	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.44	42	3558	0.99	ng	# 97
6) Naphthalene	10.45	128	30725	1.04	ng	100
7) 2-Methylnaphthalene	12.07	142	19342	1.04	ng	98
10) Acenaphthylene	13.98	152	28578	1.04	ng	100
11) Acenaphthene	14.33	154	19147	1.00	ng	98
12) Fluorene	15.32	166	22308	1.05	ng	99
14) Hexachlorobenzene	16.33	284	8094	1.04	ng	# 98
15) Pentachlorophenol	16.68	266	1377	1.02	ng	98
16) Phenanthrene	17.05	178	37201	1.02	ng	100
17) Anthracene	17.15	178	30449	1.00	ng	99
18) Fluoranthene	19.08	202	36174	1.01	ng	100
20) Pyrene	19.44	202	38409	1.04	ng	99
22) Benzo(a)anthracene	21.19	228	32732	1.01	ng	# 89
23) Chrysene	21.24	228	35164	1.04	ng	99
24) Indeno(1,2,3-cd)pyrene	25.58	276	36419	1.04	ng	100
26) Benzo(b)fluoranthene	22.74	252	33710	1.05	ng	100
27) Benzo(k)fluoranthene	22.79	252	31931	1.04	ng	99
28) Benzo(a)pyrene	23.30	252	28610	1.04	ng	96
29) Dibenzo(a,h)anthracene	25.60	278	28602	1.05	ng	99
30) Benzo(g,h,i)perylene	26.25	276	30645	1.04	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM033115\
Data File : BM000768.D
Acq On : 01 Apr 2015 00:13
Operator : TP/IZ
Sample : SSTDICV001
Misc : MDL-SOIL-BLK
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM033115

Quant Time: Apr 01 15:08:09 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 01 04:17:30 2015
Response via : Initial Calibration

