

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM033115\
 Data File : BM000768.D
 Acq On : 01 Apr 2015 00:13
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc : MDL-SOIL-BLK
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM033115

Quant Time: Apr 01 15:08:09 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.485	-4.8	114	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
5 S	Nitrobenzene-d5	0.247	0.261	-5.7	120	0.00
6	Naphthalene	1.162	1.214	-4.5	115	0.00
7	2-Methylnaphthalene	0.734	0.764	-4.1	115	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
9 S	2-Fluorobiphenyl	1.551	1.623	-4.6	115	0.00
10	Acenaphthylene	2.213	2.299	-3.9	118	0.00
11	Acenaphthene	1.533	1.540	-0.5	115	0.00
12	Fluorene	1.716	1.794	-4.5	117	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
14	Hexachlorobenzene	0.293	0.305	-4.1	120	0.00
15	Pentachlorophenol	0.069	0.052	24.6	133	0.00
16	Phenanthrene	1.370	1.403	-2.4	119	0.00
17	Anthracene	1.148	1.148	0.0	117	0.00
18	Fluoranthene	1.353	1.364	-0.8	119	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
20	Pyrene	1.631	1.700	-4.2	122	0.00
21 S	Terphenyl-d14	0.602	0.632	-5.0	121	0.00
22	Benzo(a)anthracene	1.428	1.448	-1.4	124	-0.01
23	Chrysene	1.497	1.556	-3.9	120	0.00
24	Indeno(1,2,3-cd)pyrene	1.549	1.612	-4.1	124	-0.01
25 I	Perylene-d12	1.000	1.000	0.0	120	0.00
26	Benzo(b)fluoranthene	1.575	1.655	-5.1	124	0.00
27	Benzo(k)fluoranthene	1.511	1.568	-3.8	121	0.00
28 C	Benzo(a)pyrene	1.352	1.405	-3.9	124	-0.01
29	Dibenzo(a,h)anthracene	1.335	1.404	-5.2	123	0.00
30	Benzo(g,h,i)perylene	1.441	1.505	-4.4	122	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0