

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM033115\
 Data File : BM000768.D
 Acq On : 01 Apr 2015 00:13
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc : MDL-SOIL-BLK
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM033115

Quant Time: Apr 01 15:08:09 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	113	0.00
2	1,4-Dioxane	1.000	0.000	100.0#	0	-2.47#
3	n-Nitrosodimethylamine	1.000	0.993	0.7	114	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	112	0.00
5 S	Nitrobenzene-d5	1.000	1.023	-2.3	120	0.00
6	Naphthalene	1.000	1.045	-4.5	115	0.00
7	2-Methylnaphthalene	1.000	1.041	-4.1	115	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	113	0.00
9 S	2-Fluorobiphenyl	1.000	1.046	-4.6	115	0.00
10	Acenaphthylene	1.000	1.039	-3.9	118	0.00
11	Acenaphthene	1.000	1.005	-0.5	115	0.00
12	Fluorene	1.000	1.045	-4.5	117	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	118	0.00
14	Hexachlorobenzene	1.000	1.042	-4.2	120	0.00
15	Pentachlorophenol	1.000	1.023	-2.3	133	0.00
16	Phenanthrene	1.000	1.024	-2.4	119	0.00
17	Anthracene	1.000	1.000	0.0	117	0.00
18	Fluoranthene	1.000	1.008	-0.8	119	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	118	0.00
20	Pyrene	1.000	1.042	-4.2	122	0.00
21 S	Terphenyl-d14	1.000	1.050	-5.0	121	0.00
22	Benzo(a)anthracene	1.000	1.014	-1.4	124	-0.01
23	Chrysene	1.000	1.039	-3.9	120	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.040	-4.0	124	-0.01
25 I	Perylene-d12	5.000	5.000	0.0	120	0.00
26	Benzo(b)fluoranthene	1.000	1.051	-5.1	124	0.00
27	Benzo(k)fluoranthene	1.000	1.038	-3.8	121	0.00
28 C	Benzo(a)pyrene	1.000	1.039	-3.9	124	-0.01
29	Dibenzo(a,h)anthracene	1.000	1.052	-5.2	123	0.00
30	Benzo(g,h,i)perylene	1.000	1.044	-4.4	122	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0