

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM033115\
 Data File : BM000775.D
 Acq On : 01 Apr 2015 04:22
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Quant Time: Apr 01 15:09:42 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.472	-1.9	92	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
5 S	Nitrobenzene-d5	0.247	0.240	2.8	91	0.00
6	Naphthalene	1.162	1.194	-2.8	93	0.00
7	2-Methylnaphthalene	0.734	0.748	-1.9	93	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9 S	2-Fluorobiphenyl	1.551	1.601	-3.2	93	0.00
10	Acenaphthylene	2.213	2.182	1.4	92	0.00
11	Acenaphthene	1.533	1.519	0.9	93	0.00
12	Fluorene	1.716	1.760	-2.6	94	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
14	Hexachlorobenzene	0.293	0.295	-0.7	95	0.00
15	Pentachlorophenol	0.069	0.031	55.1#	64	0.00
16	Phenanthrene	1.370	1.375	-0.4	96	0.00
17	Anthracene	1.148	1.097	4.4	92	0.00
18	Fluoranthene	1.353	1.313	3.0	94	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
20	Pyrene	1.631	1.641	-0.6	95	0.00
21 S	Terphenyl-d14	0.602	0.616	-2.3	95	0.00
22	Benzo(a)anthracene	1.428	1.387	2.9	96	-0.01
23	Chrysene	1.497	1.545	-3.2	97	0.00
24	Indeno(1,2,3-cd)pyrene	1.549	1.591	-2.7	99	-0.01
25 I	Perylene-d12	1.000	1.000	0.0	98	0.00
26	Benzo(b)fluoranthene	1.575	1.594	-1.2	98	0.00
27	Benzo(k)fluoranthene	1.511	1.538	-1.8	97	0.00
28 C	Benzo(a)pyrene	1.352	1.340	0.9	97	-0.01
29	Dibenzo(a,h)anthracene	1.335	1.371	-2.7	99	0.00
30	Benzo(g,h,i)perylene	1.441	1.486	-3.1	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0