

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM033115\
 Data File : BM000777.D
 Acq On : 01 Apr 2015 06:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 01 15:09:56 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.461	0.4	90	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
5 S	Nitrobenzene-d5	0.247	0.249	-0.8	94	0.00
6	Naphthalene	1.162	1.198	-3.1	93	0.00
7	2-Methylnaphthalene	0.734	0.752	-2.5	93	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9 S	2-Fluorobiphenyl	1.551	1.609	-3.7	93	0.00
10	Acenaphthylene	2.213	2.224	-0.5	93	0.00
11	Acenaphthene	1.533	1.551	-1.2	95	0.00
12	Fluorene	1.716	1.782	-3.8	95	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
14	Hexachlorobenzene	0.293	0.305	-4.1	98	0.00
15	Pentachlorophenol	0.069	0.038	44.9#	80	0.00
16	Phenanthrene	1.370	1.404	-2.5	97	0.00
17	Anthracene	1.148	1.096	4.5	91	0.00
18	Fluoranthene	1.353	1.353	0.0	97	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	97	0.01
20	Pyrene	1.631	1.653	-1.3	97	0.00
21 S	Terphenyl-d14	0.602	0.618	-2.7	96	0.00
22	Benzo(a)anthracene	1.428	1.421	0.5	99	0.00
23	Chrysene	1.497	1.571	-4.9	99	0.01
24	Indeno(1,2,3-cd)pyrene	1.549	1.638	-5.7	103	0.00
25 I	Perylene-d12	1.000	1.000	0.0	102	0.00
26	Benzo(b)fluoranthene	1.575	1.553	1.4	99	0.00
27	Benzo(k)fluoranthene	1.511	1.561	-3.3	103	0.00
28 C	Benzo(a)pyrene	1.352	1.370	-1.3	104	0.00
29	Dibenzo(a,h)anthracene	1.335	1.360	-1.9	102	0.00
30	Benzo(g,h,i)perylene	1.441	1.470	-2.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0