

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033121\
 Data File : BM029289.D
 Acq On : 31 Mar 2021 12:10
 Operator : CG/JU
 Sample : M1802-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SB-1S-(0.5-1)

Quant Time: Mar 31 12:41:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 11:14:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.745	152	113758	20.00	ng	0.00
21) Naphthalene-d8	10.528	136	446976	20.00	ng	0.00
39) Acenaphthene-d10	14.374	164	260614	20.00	ng	0.00
64) Phenanthrene-d10	17.115	188	499658	20.00	ng	0.00
76) Chrysene-d12	21.286	240	487696	20.00	ng	0.00
86) Perylene-d12	23.527	264	539960	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	671932	101.82	ng	0.00
7) Phenol-d6	6.922	99	888493	93.82	ng	0.00
23) Nitrobenzene-d5	8.898	82	616301	66.94	ng	0.00
42) 2,4,6-Tribromophenol	15.863	330	255848	106.01	ng	0.00
45) 2-Fluorobiphenyl	12.998	172	1158634	63.06	ng	0.00
79) Terphenyl-d14	19.739	244	1506945	62.07	ng	0.00
Target Compounds						Qvalue
50) Dimethylphthalate	13.833	163	36705	2.02	ng	98
75) Fluoranthene	19.168	202	93453	2.92	ng	97
78) Pyrene	19.533	202	92001	2.79	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033121\
Data File : BM029289.D
Acq On : 31 Mar 2021 12:10
Operator : CG/JU
Sample : M1802-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SB-1S-(0.5-1)

Quant Time: Mar 31 12:41:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 29 11:14:33 2021
Response via : Initial Calibration

