

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033123\
 Data File : BM039260.D
 Acq On : 31 Mar 2023 19:23
 Operator : CG/JU
 Sample : 02140-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MW-02

Quant Time: Apr 01 05:21:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 17:03:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	203633	20.000	ng	0.00
21) Naphthalene-d8	10.728	136	807614	20.000	ng	0.00
39) Acenaphthene-d10	14.563	164	460903	20.000	ng	0.00
64) Phenanthrene-d10	17.310	188	912643	20.000	ng	0.00
76) Chrysene-d12	21.492	240	773250	20.000	ng	0.00
86) Perylene-d12	23.904	264	891674	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	855479	63.294	ng	0.00
7) Phenol-d6	7.081	99	636525	36.410	ng	0.00
23) Nitrobenzene-d5	9.087	82	1547252	93.114	ng	0.00
42) 2,4,6-Tribromophenol	16.057	330	820159	140.181	ng	0.00
45) 2-Fluorobiphenyl	13.187	172	3208474	94.319	ng	0.00
79) Terphenyl-d14	19.933	244	3883336	92.786	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033123\
Data File : BM039260.D
Acq On : 31 Mar 2023 19:23
Operator : CG/JU
Sample : 02140-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MW-02

Quant Time: Apr 01 05:21:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 17:03:41 2023
Response via : Initial Calibration

