

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033123\
 Data File : BM039246.D
 Acq On : 31 Mar 2023 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 31 17:03:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 17:03:41 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	40.872	-2.2	99	0.00
3	Pyridine	40.000	42.046	-5.1	99	0.00
4	n-Nitrosodimethylamine	40.000	42.354	-5.9	102	0.00
5 S	2-Fluorophenol	80.000	81.325	-1.7	98	0.00
6	Aniline	40.000	41.436	-3.6	99	0.00
7 S	Phenol-d6	80.000	81.720	-2.1	98	0.00
8	2-Chlorophenol	40.000	40.387	-1.0	97	0.00
9	Benzaldehyde	40.000	39.906	0.2	96	0.00
10 C	Phenol	40.000	40.873	-2.2	98	0.00
11	bis(2-Chloroethyl)ether	40.000	41.263	-3.2	99	0.00
12	1,3-Dichlorobenzene	40.000	39.798	0.5	96	0.00
13 C	1,4-Dichlorobenzene	40.000	39.610	1.0	95	0.00
14	1,2-Dichlorobenzene	40.000	39.813	0.5	96	0.00
15	Benzyl Alcohol	40.000	40.918	-2.3	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.256	-3.1	100	0.00
17	2-Methylphenol	40.000	40.964	-2.4	98	0.00
18	Hexachloroethane	40.000	39.275	1.8	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.550	-3.9	100	0.00
20	3+4-Methylphenols	40.000	41.487	-3.7	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	40.647	-1.6	98	0.00
23 S	Nitrobenzene-d5	80.000	81.581	-2.0	98	0.00
24	Nitrobenzene	40.000	40.875	-2.2	99	0.00
25	Isophorone	40.000	41.362	-3.4	100	0.00
26 C	2-Nitrophenol	40.000	40.905	-2.3	97	0.00
27	2,4-Dimethylphenol	40.000	40.726	-1.8	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.915	-2.3	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.774	-1.9	97	0.00
30	1,2,4-Trichlorobenzene	40.000	39.390	1.5	96	0.00
31	Naphthalene	40.000	39.953	0.1	97	0.00
32	Benzoic acid	40.000	41.276	-3.2	96	0.00
33	4-Chloroaniline	40.000	40.950	-2.4	98	0.00
34 C	Hexachlorobutadiene	40.000	39.287	1.8	96	0.00
35	Caprolactam	40.000	41.633	-4.1	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.898	-2.2	97	0.00
37	2-Methylnaphthalene	40.000	40.189	-0.5	98	0.00
38	1-Methylnaphthalene	40.000	39.896	0.3	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.475	1.3	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.138	2.2	93	0.00
42 S	2,4,6-Tribromophenol	80.000	84.320	-5.4	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.454	-1.1	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.341	-0.9	98	0.00
45 S	2-Fluorobiphenyl	80.000	79.266	0.9	99	0.00
46	1,1'-Biphenyl	40.000	39.760	0.6	99	0.00
47	2-Chloronaphthalene	40.000	39.690	0.8	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033123\
 Data File : BM039246.D
 Acq On : 31 Mar 2023 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 31 17:03:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 17:03:41 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.799	-2.0	100	0.00
49	Acenaphthylene	40.000	39.764	0.6	99	0.00
50	Dimethylphthalate	40.000	40.301	-0.8	101	0.00
51	2,6-Dinitrotoluene	40.000	40.777	-1.9	100	0.00
52 C	Acenaphthene	40.000	40.068	-0.2	100	0.00
53	3-Nitroaniline	40.000	40.920	-2.3	99	0.00
54 P	2,4-Dinitrophenol	40.000	41.215	-3.0	99	0.00
55	Dibenzofuran	40.000	39.967	0.1	100	0.00
56 P	4-Nitrophenol	40.000	41.882	-4.7	97	0.00
57	2,4-Dinitrotoluene	40.000	41.630	-4.1	100	0.00
58	Fluorene	40.000	40.424	-1.1	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.511	-1.3	100	0.00
60	Diethylphthalate	40.000	40.783	-2.0	102	0.00
61	4-Chlorophenyl-phenylether	40.000	40.448	-1.1	102	0.00
62	4-Nitroaniline	40.000	41.466	-3.7	99	0.00
63	Azobenzene	40.000	41.917	-4.8	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.943	-2.4	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.904	0.2	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.256	-0.6	103	0.00
68	Hexachlorobenzene	40.000	40.261	-0.7	104	0.00
69	Atrazine	40.000	40.644	-1.6	104	0.00
70 C	Pentachlorophenol	40.000	42.234	-5.6	103	0.00
71	Phenanthrene	40.000	39.932	0.2	103	0.00
72	Anthracene	40.000	40.388	-1.0	103	0.00
73	Carbazole	40.000	40.500	-1.3	103	0.00
74	Di-n-butylphthalate	40.000	42.012	-5.0	106	0.00
75 C	Fluoranthene	40.000	40.275	-0.7	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	39.701	0.7	96	0.00
78	Pyrene	40.000	39.995	0.0	103	0.00
79 S	Terphenyl-d14	80.000	81.934	-2.4	105	0.00
80	Butylbenzylphthalate	40.000	41.117	-2.8	104	0.00
81	Benzo(a)anthracene	40.000	40.683	-1.7	105	0.00
82	3,3'-Dichlorobenzidine	40.000	41.041	-2.6	104	0.00
83	Chrysene	40.000	39.866	0.3	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.549	-6.4	109	0.00
85 c	Di-n-octyl phthalate	40.000	40.029	-0.1	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.750	0.6	94	0.00
88	Benzo(b)fluoranthene	40.000	42.223	-5.6	101	0.00
89	Benzo(k)fluoranthene	40.000	40.460	-1.2	96	0.00
90 C	Benzo(a)pyrene	40.000	40.559	-1.4	96	0.00
91	Dibenzo(a,h)anthracene	40.000	39.710	0.7	93	0.00
92	Benzo(g,h,i)perylene	40.000	38.989	2.5	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM033123\
Data File : BM039246.D
Acq On : 31 Mar 2023 10:31
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 31 17:03:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 17:03:41 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0