

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040115\
 Data File : BM000801.D
 Acq On : 02 Apr 2015 01:18
 Operator : TP/IZ
 Sample : PB82421BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB82421BS

Manual Integrations
 APPROVED

apatel
 4/2/2015 5:38:55 PM

Quant Time: Apr 02 03:00:31 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	28593	5.00	ng	0.00
4) Naphthalene-d8	10.40	136	99949	5.00	ng	0.00
8) Acenaphthene-d10	14.28	164	51074	5.00	ng	0.00
13) Phenanthrene-d10	17.01	188	107969	5.00	ng	0.00
19) Chrysene-d12	21.21	240	92794	5.00	ng	0.00
25) Perylene-d12	23.39	264	82852	5.00	ng	-0.01
System Monitoring Compounds						
5) Nitrobenzene-d5	8.77	82	44721	7.31	ng	0.00
9) 2-Fluorobiphenyl	12.89	172	135850	8.57	ng	0.00
21) Terphenyl-d14	19.66	244	107131	9.59	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.43	42	19834	6.15	ng	# 93
6) Naphthalene	10.45	128	159264	6.86	ng	99
7) 2-Methylnaphthalene	12.07	142	105843	7.21	ng	92
10) Acenaphthylene	13.98	152	186483	8.25	ng	100
11) Acenaphthene	14.33	154	119551	7.64	ng	94
12) Fluorene	15.32	166	142988	8.16	ng	99
14) Hexachlorobenzene	16.33	284	51231	8.10	ng	# 96
15) Pentachlorophenol	16.68	266	31425	10.61	ng	90
16) Phenanthrene	17.05	178	231073	7.81	ng	99
17) Anthracene	17.13	178	202682m	8.18	ng	
18) Fluoranthene	19.08	202	237492	8.13	ng	100
20) Pyrene	19.44	202	255382	8.44	ng	100
22) Benzo(a)anthracene	21.19	228	227415	8.58	ng	94
23) Chrysene	21.24	228	218013	7.85	ng	95
24) Indeno(1,2,3-cd)pyrene	25.57	276	229327	7.98	ng	100
26) Benzo(b)fluoranthene	22.73	252	211773	8.11	ng	96
27) Benzo(k)fluoranthene	22.78	252	215907	8.62	ng	97
28) Benzo(a)pyrene	23.30	252	202322	9.03	ng	97
29) Dibenzo(a,h)anthracene	25.59	278	185287	8.38	ng	99
30) Benzo(g,h,i)perylene	26.25	276	195217	8.18	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040115\
Data File : BM000801.D
Acq On : 02 Apr 2015 01:18
Operator : TP/IZ
Sample : PB82421BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB82421BS

Manual Integrations
APPROVED

apatel
4/2/2015 5:38:55 PM

Quant Time: Apr 02 03:00:31 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 01 04:17:30 2015
Response via : Initial Calibration

