

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040115\
 Data File : BM000802.D
 Acq On : 02 Apr 2015 01:53
 Operator : TP/IZ
 Sample : PB82421BSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB82421BSD

Manual Integrations
 APPROVED

apatel
 4/2/2015 5:38:58 PM

Quant Time: Apr 02 03:02:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	33599	5.00	ng	0.00
4) Naphthalene-d8	10.40	136	117003	5.00	ng	0.00
8) Acenaphthene-d10	14.28	164	58789	5.00	ng	0.00
13) Phenanthrene-d10	17.01	188	123539	5.00	ng	0.00
19) Chrysene-d12	21.21	240	105964	5.00	ng	0.00
25) Perylene-d12	23.39	264	94350	5.00	ng	-0.01
System Monitoring Compounds						
5) Nitrobenzene-d5	8.77	82	47150	6.60	ng	0.00
9) 2-Fluorobiphenyl	12.89	172	146230	8.02	ng	0.00
21) Terphenyl-d14	19.66	244	116128	9.11	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.43	42	20721	5.49	ng	97
6) Naphthalene	10.45	128	172406	6.34	ng	99
7) 2-Methylnaphthalene	12.07	142	117925	6.86	ng	93
10) Acenaphthylene	13.98	152	217804	8.37	ng	100
11) Acenaphthene	14.33	154	137148	7.61	ng	93
12) Fluorene	15.32	166	164667	8.16	ng	100
14) Hexachlorobenzene	16.33	284	58239	8.05	ng	# 97
15) Pentachlorophenol	16.68	266	40267	11.42	ng	91
16) Phenanthrene	17.05	178	268155	7.92	ng	99
17) Anthracene	17.13	178	242524m	8.55	ng	
18) Fluoranthene	19.08	202	276922	8.28	ng	100
20) Pyrene	19.44	202	297158	8.60	ng	100
22) Benzo(a)anthracene	21.19	228	259053	8.56	ng	93
23) Chrysene	21.24	228	250408	7.89	ng	97
24) Indeno(1,2,3-cd)pyrene	25.58	276	259637	7.91	ng	99
26) Benzo(b)fluoranthene	22.73	252	238065	8.01	ng	96
27) Benzo(k)fluoranthene	22.78	252	249895	8.76	ng	96
28) Benzo(a)pyrene	23.30	252	230746	9.04	ng	97
29) Dibenzo(a,h)anthracene	25.59	278	210815	8.37	ng	98
30) Benzo(g,h,i)perylene	26.25	276	220744	8.12	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040115\
Data File : BM000802.D
Acq On : 02 Apr 2015 01:53
Operator : TP/IZ
Sample : PB82421BSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB82421BSD

Manual Integrations
APPROVED

apatel
4/2/2015 5:38:58 PM

Quant Time: Apr 02 03:02:16 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 01 04:17:30 2015
Response via : Initial Calibration

