

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040115\
 Data File : BM000790.D
 Acq On : 01 Apr 2015 17:54
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 02 01:57:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.466	-0.6	91	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
5 S	Nitrobenzene-d5	0.247	0.251	-1.6	95	0.00
6	Naphthalene	1.162	1.192	-2.6	93	0.00
7	2-Methylnaphthalene	0.734	0.752	-2.5	93	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9 S	2-Fluorobiphenyl	1.551	1.602	-3.3	93	0.00
10	Acenaphthylene	2.213	2.260	-2.1	95	0.00
11	Acenaphthene	1.533	1.562	-1.9	96	0.00
12	Fluorene	1.716	1.809	-5.4	97	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
14	Hexachlorobenzene	0.293	0.287	2.0	96	0.00
15	Pentachlorophenol	0.069	0.033	52.2#	72	0.00
16	Phenanthrene	1.370	1.376	-0.4	100	0.00
17	Anthracene	1.148	1.117	2.7	97	0.00
18	Fluoranthene	1.353	1.354	-0.1	101	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
20	Pyrene	1.631	1.632	-0.1	102	0.00
21 S	Terphenyl-d14	0.602	0.613	-1.8	102	0.00
22	Benzo(a)anthracene	1.428	1.424	0.3	106	-0.01
23	Chrysene	1.497	1.517	-1.3	102	0.00
24	Indeno(1,2,3-cd)pyrene	1.549	1.579	-1.9	106	-0.01
25 I	Perylene-d12	1.000	1.000	0.0	107	0.00
26	Benzo(b)fluoranthene	1.575	1.680	-6.7	112	-0.01
27	Benzo(k)fluoranthene	1.511	1.425	5.7	98	0.00
28 C	Benzo(a)pyrene	1.352	1.373	-1.6	108	-0.01
29	Dibenzo(a,h)anthracene	1.335	1.337	-0.1	105	-0.01
30	Benzo(g,h,i)perylene	1.441	1.444	-0.2	105	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0